

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080318\
 Data File : BN002217.D
 Acq On : 03 Aug 2018 11:01
 Operator : JU/SJ
 Sample : J4178-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AB8

Quant Time: Aug 03 15:37:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 03 15:22:46 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	270245	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	1260531	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	744496	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	1571788	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	1269400	20.00	ng/ul	0.00
85) Perylene-d12	23.49	264	1094309	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	29856	4.98	ng/uL	0.00
5) Phenol-d5	6.86	99	607768	23.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	422788	27.27	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	496831	24.72	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	507834	24.77	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	268693	25.73	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	275324	23.32	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	492916	23.06	ng/ul	0.00
29) 4-Chloroaniline-d4	10.60	131	720034	29.30	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	1656129	26.24	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	2086578	26.58	ng/ul	0.00
51) 4-Nitrophenol-d4	14.53	143	168172	13.95	ng/ul	0.00
57) Fluorene-d10	15.32	176	1430007	26.35	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.44	200	160145	15.64	ng/ul	0.00
70) Anthracene-d10	17.16	188	2129912	27.79	ng/ul	0.00
78) Pyrene-d10	19.46	212	2165636	32.82	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	1665585	28.19	ng/ul	0.00
Target Compounds						
6) Phenol	6.89	94	349646	13.681	ng/ul	93
44) Dimethylphthalate	13.77	163	365863	5.919	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080318\
Data File : BN002217.D
Acq On : 03 Aug 2018 11:01
Operator : JU/SJ
Sample : J4178-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AB8

Quant Time: Aug 03 15:37:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 03 15:22:46 2018
Response via : Initial Calibration






#6

Phenol




Concen: 13.681 ng/ul
 RT: 6.89 min Scan# 664
 Delta R.T. 0.00 min
 Lab File: BN002217.D
 Acq: 03 Aug 2018 11:01

Instrument :
 BNA_N
 ClientSampleId :
 C0AB8

Tot Ion: 94 Resp: 349646
 Ion Ratio Lower Upper
 94 100
 65 29.9 21.0 31.6
 66 40.6 29.3 43.9



Abundance Ion 94.00 (93.70 to 94.70): BN0
 Ion 65.00 (64.70 to 65.70): BN0
 Ion 66.00 (65.70 to 66.70): BN0




#44

Dimethylphthalate

Concen: 5.919 ng/ul
 RT: 13.77 min Scan# 1834
 Delta R.T. -0.01 min
 Lab File: BN002217.D
 Acq: 03 Aug 2018 11:01

Tgt Ion:163 Resp: 365863
 Ion Ratio Lower Upper
 163 100
 194 4.0 3.2 4.8
 164 10.0 7.8 11.8



Abundance Ion 163.00 (162.70 to 163.70): E
 Ion 194.00 (193.70 to 194.70): E
 Ion 164.00 (163.70 to 164.70): E

