

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080318\
 Data File : BN002227.D
 Acq On : 03 Aug 2018 18:06
 Operator : JU/SJ
 Sample : J4265-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AB2

Quant Time: Aug 03 18:43:44 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 03 15:22:46 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	241728	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	1116070	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	673169	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	1515923	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	1297545	20.00	ng/ul	0.00
85) Perylene-d12	23.49	264	1084039	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	35665	6.66	ng/uL	0.00
5) Phenol-d5	6.87	99	141674	6.22	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	349749	25.22	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	363747	20.23	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	257586	14.05	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	231739	25.06	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	247456	23.67	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	412480	21.79	ng/ul	0.00
29) 4-Chloroaniline-d4	10.62	131	2775	0.13	ng/ul	0.02
43) Dimethylphthalate-d6	13.73	166	1527556	26.77	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	1843939	25.98	ng/ul	0.00
51) 4-Nitrophenol-d4	14.54	143	46799	4.29	ng/ul	0.00
57) Fluorene-d10	15.32	176	1294574	26.38	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.44	200	221722	22.46	ng/ul	0.00
70) Anthracene-d10	17.16	188	2089171	28.26	ng/ul	0.00
78) Pyrene-d10	19.46	212	2203321	32.66	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.35	264	1618732	27.65	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080318\
Data File : BN002227.D
Acq On : 03 Aug 2018 18:06
Operator : JU/SJ
Sample : J4265-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AB2

Quant Time: Aug 03 18:43:44 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 03 15:22:46 2018
Response via : Initial Calibration

