

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
 Data File : BN026806.D
 Acq On : 03 Aug 2023 15:14
 Operator : MA/JU
 Sample : SSTD3.224
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2242

Quant Time: Aug 04 03:39:05 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 03 15:46:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.118	152	6960	0.400	ng/ul	0.00
4) Naphthalene-d8	10.949	136	21637	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.745	164	12402	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.489	188	27123	0.400	ng/ul	0.00
17) Chrysene-d12	21.667	240	24009	0.400	ng/ul	0.00
23) Perylene-d12	24.222	264	25855	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.444	96	27006	3.268	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.482	88	27931	3.311	ng/ul#	90
14) Pentachlorophenol	17.113	266	28307	3.689	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
Data File : BN026806.D
Acq On : 03 Aug 2023 15:14
Operator : MA/JU
Sample : SSTD3.224
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2242

Quant Time: Aug 04 03:39:05 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 03 15:46:03 2023
Response via : Initial Calibration

