

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
 Data File : BN026807.D
 Acq On : 03 Aug 2023 15:51
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV243

Quant Time: Aug 04 03:47:56 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 03:46:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.118	152	8994	0.400	ng/ul	0.00
4) Naphthalene-d8	10.949	136	25615	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.745	164	13416	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.489	188	32951	0.400	ng/ul	0.00
17) Chrysene-d12	21.665	240	29767	0.400	ng/ul	0.00
23) Perylene-d12	24.220	264	34278	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.448	96	4298	0.403	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.522	152	15623	0.415	ng/ul	0.00
18) Fluoranthene-d10	19.506	212	37707	0.401	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.486	88	4287	0.393	ng/ul	92
5) Naphthalene	10.999	128	28825	0.398	ng/ul	100
7) 2-Methylnaphthalene	12.594	142	18202	0.415	ng/ul	100
8) 1-Methylnaphthalene	12.808	142	19069	0.416	ng/ul	100
10) Acenaphthylene	14.472	152	27976	0.372	ng/ul	99
11) Acenaphthene	14.809	153	20143	0.387	ng/ul	99
12) Fluorene	15.790	166	21621	0.367	ng/ul	98
14) Pentachlorophenol	17.114	266	3199	0.343	ng/ul	100
15) Phenanthrene	17.532	178	41763	0.386	ng/ul	100
16) Anthracene	17.624	178	37156	0.380	ng/ul	99
19) Fluoranthene	19.538	202	49482	0.400	ng/ul	100
20) Pyrene	19.901	202	51666	0.392	ng/ul	98
21) Benzo(a)anthracene	21.647	228	46538	0.393	ng/ul	99
22) Chrysene	21.703	228	46305	0.396	ng/ul	100
24) Benzo(b)fluoranthene	23.424	252	52142	0.402	ng/ul	99
25) Benzo(k)fluoranthene	23.480	252	50775	0.397	ng/ul	99
26) Benzo(a)pyrene	24.103	252	47866	0.409	ng/ul#	96
27) Indeno(1,2,3-cd)pyrene	26.918	276	62082	0.406	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.958	278	49065	0.402	ng/ul	99
29) Benzo(g,h,i)perylene	27.763	276	50313	0.395	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
Data File : BN026807.D
Acq On : 03 Aug 2023 15:51
Operator : MA/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SICV243

Quant Time: Aug 04 03:47:56 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 03:46:36 2023
Response via : Initial Calibration

