

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
 Data File : BN026813.D
 Acq On : 03 Aug 2023 19:32
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4380

Quant Time: Aug 04 04:25:40 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 03:46:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.118	152	8042	0.400	ng/ul	0.00
4) Naphthalene-d8	10.944	136	23490	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.744	164	12334	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.489	188	29617	0.400	ng/ul	0.00
17) Chrysene-d12	21.662	240	25811	0.400	ng/ul	0.00
23) Perylene-d12	24.214	264	29307	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.448	96	3854	0.404	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.516	152	13863	0.401	ng/ul	0.00
18) Fluoranthene-d10	19.506	212	33040	0.405	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.481	88	3978	0.408	ng/ul	97
5) Naphthalene	10.999	128	26662	0.402	ng/ul	100
7) 2-Methylnaphthalene	12.593	142	16318	0.406	ng/ul	99
8) 1-Methylnaphthalene	12.808	142	17105	0.406	ng/ul	99
10) Acenaphthylene	14.471	152	25511	0.369	ng/ul	99
11) Acenaphthene	14.809	153	18545	0.388	ng/ul	99
12) Fluorene	15.790	166	20416	0.377	ng/ul	98
14) Pentachlorophenol	17.113	266	2604	0.311	ng/ul	99
15) Phenanthrene	17.531	178	37545	0.386	ng/ul	100
16) Anthracene	17.624	178	32966	0.375	ng/ul	100
19) Fluoranthene	19.534	202	43563	0.406	ng/ul	98
20) Pyrene	19.901	202	45395	0.397	ng/ul	97
21) Benzo(a)anthracene	21.644	228	39715	0.387	ng/ul	99
22) Chrysene	21.700	228	40265	0.397	ng/ul	99
24) Benzo(b)fluoranthene	23.421	252	46108	0.416	ng/ul	98
25) Benzo(k)fluoranthene	23.474	252	43430	0.397	ng/ul	100
26) Benzo(a)pyrene	24.100	252	41595	0.416	ng/ul#	96
27) Indeno(1,2,3-cd)pyrene	26.914	276	53987	0.413	ng/ul#	98
28) Dibenzo(a,h)anthracene	26.951	278	42587	0.408	ng/ul	99
29) Benzo(g,h,i)perylene	27.753	276	43384	0.398	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080323\
Data File : BN026813.D
Acq On : 03 Aug 2023 19:32
Operator : MA/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4380

Quant Time: Aug 04 04:25:40 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 03:46:36 2023
Response via : Initial Calibration

