

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080423\
 Data File : BN026817.D
 Acq On : 04 Aug 2023 14:15
 Operator : MA/JU
 Sample : 03805-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X3H69

Quant Time: Aug 04 22:30:42 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 03:46:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.131	152	9695	0.400	ng/ul	0.01
4) Naphthalene-d8	10.960	136	29266	0.400	ng/ul	0.02
9) Acenaphthene-d10	14.754	164	16972	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.498	188	38368	0.400	ng/ul	0.00
17) Chrysene-d12	21.673	240	32724	0.400	ng/ul	0.00
23) Perylene-d12	24.234	264	36801	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.460	96	4105	0.357	ng/ul	0.02
6) 2-Methylnaphthalene-d10	12.528	152	12978	0.301	ng/ul	0.01
18) Fluoranthene-d10	19.515	212	39677	0.384	ng/ul	0.00
Target Compounds					Qvalue	
5) Naphthalene	11.010	128	1202377	14.539	ng/ul	99
10) Acenaphthylene	14.485	152	1376566	14.477	ng/ul	98
11) Acenaphthene	14.814	153	776949	11.807	ng/ul	98
12) Fluorene	15.799	166	505006	6.784	ng/ul	98
15) Phenanthrene	17.540	178	846547	6.716	ng/ul	100
16) Anthracene	17.633	178	640265	5.621	ng/ul	99
19) Fluoranthene	19.543	202	1131598	8.325	ng/ul	99
20) Pyrene	19.906	202	1356853	9.355	ng/ul	99
21) Benzo(a)anthracene	21.656	228	384882	2.957	ng/ul	100
22) Chrysene	21.714	228	667464	5.196	ng/ul	99
24) Benzo(b)fluoranthene	23.439	252	819970	5.891	ng/ul	97
25) Benzo(k)fluoranthene	23.495	252	690096	5.026	ng/ul	99
26) Benzo(a)pyrene	24.117	252	935802	7.450	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	26.945	276	1034125	6.294	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.975	278	664038	5.070	ng/ul	96
29) Benzo(g,h,i)perylene	27.783	276	701979	5.131	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080423\
Data File : BN026817.D
Acq On : 04 Aug 2023 14:15
Operator : MA/JU
Sample : 03805-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X3H69

Quant Time: Aug 04 22:30:42 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 03:46:36 2023
Response via : Initial Calibration

