

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
 Data File : BN033249.D
 Acq On : 06 Aug 2024 00:12
 Operator : MA/JU
 Sample : PB162464BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB162464BL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 04:31:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.676	152	1029m	0.400	ng	0.17
7) Naphthalene-d8	10.522	136	3111m	0.400	ng	0.20
13) Acenaphthene-d10	14.222	164	1921m	0.400	ng	0.10
19) Phenanthrene-d10	17.019	188	4706m	0.400	ng	0.10
29) Chrysene-d12	21.160	240	5871	0.400	ng	# 0.04
35) Perylene-d12	23.317	264	6677	0.400	ng	# 0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.336	112	827m	0.324	ng	0.04
5) Phenol-d6	7.300	99	1121m	0.320	ng	0.31
8) Nitrobenzene-d5	9.272	82	888m	0.355	ng	0.35
11) 2-Methylnaphthalene-d10	12.135	152	1535m	0.333	ng	0.21
14) 2,4,6-Tribromophenol	15.840	330	210m	0.293	ng	0.12
15) 2-Fluorobiphenyl	12.961	172	2185m	0.333	ng	0.16
27) Fluoranthene-d10	19.001	212	5283	0.401	ng	0.05
31) Terphenyl-d14	19.577	244	4047	0.399	ng	0.03

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
Data File : BN033249.D
Acq On : 06 Aug 2024 00:12
Operator : MA/JU
Sample : PB162464BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB162464BL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 04:31:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 05 23:57:48 2024
Response via : Initial Calibration

