

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
 Data File : BN033252.D
 Acq On : 06 Aug 2024 02:04
 Operator : MA/JU
 Sample : PB162464BL
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB162464BL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
 Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 04:32:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 05 23:57:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.625	152	1306m	0.400	ng	0.12
7) Naphthalene-d8	10.511	136	4124m	0.400	ng	0.19
13) Acenaphthene-d10	14.211	164	2380m	0.400	ng	0.09
19) Phenanthrene-d10	17.007	188	5709m	0.400	ng	0.09
29) Chrysene-d12	21.160	240	6775	0.400	ng	# 0.04
35) Perylene-d12	23.317	264	7732	0.400	ng	# 0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.293	112	1354m	0.419	ng	0.00
5) Phenol-d6	7.243	99	1473m	0.331	ng	0.25
8) Nitrobenzene-d5	9.187	82	1299m	0.392	ng	0.27
11) 2-Methylnaphthalene-d10	12.109	152	2054m	0.336	ng	0.18
14) 2,4,6-Tribromophenol	15.828	330	179m	0.202	ng	0.11
15) 2-Fluorobiphenyl	12.950	172	2828m	0.347	ng	0.15
27) Fluoranthene-d10	19.001	212	6381	0.399	ng	0.05
31) Terphenyl-d14	19.577	244	4918	0.420	ng	0.03

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080524\
Data File : BN033252.D
Acq On : 06 Aug 2024 02:04
Operator : MA/JU
Sample : PB162464BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB162464BL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/06/2024
Supervised By :mohammad ahmed 08/07/2024

Quant Time: Aug 06 04:32:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 05 23:57:48 2024
Response via : Initial Calibration

