

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080719\
 Data File : BN006996.D
 Acq On : 07 Aug 2019 20:21
 Operator : HP/JU
 Sample : K3893-17DL 10X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BEP21DL

Manual Integrations
 APPROVED

mohammad
 8/9/2019 6:15:07 AM

Quant Time: Aug 08 05:20:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN080719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 17:30:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	7023	0.40	ng/ul	0.00
2) Naphthalene-d8	10.23	136	28601	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.12	164	18290	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.86	188	46178	0.40	ng/ul	0.00
16) Chrysene-d12	21.08	240	41002	0.40	ng/ul	0.00
20) Perylene-d12	23.21	264	36494	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.83	152	1220	0.02	ng/ul	0.00
14) Fluoranthene-d10	18.90	212	3423	0.02	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.28	128	1647	0.020	ng/ul#	77
7) Acenaphthylene	13.83	152	4987	0.047	ng/ul#	95
9) Fluorene	15.17	166	1810	0.021	ng/ul#	96
12) Phenanthrene	16.90	178	44342	0.301	ng/ul	100
13) Anthracene	16.99	178	7490	0.050	ng/ul	96
15) Fluoranthene	18.93	202	133043	0.727	ng/ul	99
17) Pyrene	19.29	202	114438	0.625	ng/ul	100
18) Benzo(a)anthracene	21.06	228	59702	0.331	ng/ul	97
19) Chrysene	21.11	228	71054	0.421	ng/ul	97
21) Benzo(b)fluoranthene	22.58	252	96335m	0.651	ng/ul	
22) Benzo(k)fluoranthene	22.62	252	28762m	0.201	ng/ul	
23) Benzo(a)pyrene	23.12	252	54788	0.393	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	25.30	276	36499	0.243	ng/ul#	93
25) Dibenzo(a,h)anthracene	25.32	278	8949	0.075	ng/ul#	79
26) Benzo(g,h,i)perylene	25.94	276	31568	0.251	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080719\
Data File : BN006996.D
Acq On : 07 Aug 2019 20:21
Operator : HP/JU
Sample : K3893-17DL 10X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BEP21DL

Manual Integrations
APPROVED

mohammad
8/9/2019 6:15:07 AM

Quant Time: Aug 08 05:20:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN080719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 17:30:08 2019
Response via : Initial Calibration

