

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080719\
 Data File : BN006987.D
 Acq On : 07 Aug 2019 14:58
 Operator : HP/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.427

Quant Time: Aug 07 17:41:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN080719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 17:30:08 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
3	Naphthalene	1.151	2.144	-86.3#	216#	0.00
4 SURR	2-Methylnaphthalene-d10	0.711	1.354	-90.4#	220#	0.00
5	2-Methylnaphthalene	0.864	1.605	-85.8#	216#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
7	Acenaphthylene	2.318	4.278	-84.6#	212#	0.00
8 C	Acenaphthene	1.586	2.939	-85.3#	212#	0.00
9	Fluorene	1.922	3.554	-84.9#	210#	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
11	Pentachlorophenol	0.110	0.178	-61.8#	165#	0.00
12	Phenanthrene	1.275	2.383	-86.9#	208#	0.00
13	Anthracene	1.304	2.395	-83.7#	204#	0.00
14 SURR	Fluoranthene-d10	1.289	2.425	-88.1#	207#	0.00
15 C	Fluoranthene	1.585	2.928	-84.7#	204#	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
17	Pyrene	1.786	3.363	-88.3#	203#	0.00
18	Benzo(a)anthracene	1.757	3.257	-85.4#	199#	0.00
19	Chrysene	1.645	3.055	-85.7#	198#	0.00
20 I	Perylene-d12	1.000	1.000	0.0	107	0.00
21	Benzo(b)fluoranthene	1.622	2.972	-83.2#	190#	0.00
22	Benzo(k)fluoranthene	1.572	2.964	-88.5#	200#	0.00
23 C	Benzo(a)pyrene	1.527	2.804	-83.6#	194#	0.00
24	Indeno(1,2,3-cd)pyrene	1.646	2.980	-81.0#	196#	0.00
25	Dibenzo(a,h)anthracene	1.315	2.391	-81.8#	198#	0.00
26	Benzo(g,h,i)perylene	1.381	2.506	-81.5#	197#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 3