



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080723\
Data File : BN026821.D
Acq On : 07 Aug 2023 08:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Aug 02 00:35:47 2023






AutoFind: Scans 2547, 2548, 2549; Background Corrected with Scan 2539

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.1	195104	PASS
68	69	0.00	2	1.7	3885	PASS
69	198	0.00	100	32.2	231341	PASS
70	69	0.00	2	0.3	709	PASS
127	198	10	80	44.3	318954	PASS
197	198	0.00	2	0.5	3517	PASS
198	198	100	100	100.0	719317	PASS
199	198	5	9	6.8	48984	PASS
275	198	10	60	27.5	198101	PASS
365	198	1	100	3.1	22195	PASS
441	198	0.01	100	13.0	93512	PASS
442	442	50	100	100.0	602133	PASS
443	442	15	24	18.8	113323	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080723\
 Data File : BN026821.D
 Acq On : 07 Aug 2023 08:24
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Aug 07 09:25:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 09:25:41 2023
 Response via : Initial Calibration



TIC: BN026821.D\data.ms

(70) Pentachlorophenol (C)

17.110min (0.000) 26558.46 ng




response 880175

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.82
264.00	61.60	57.52
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080723\
Data File : BN026821.D
Acq On : 07 Aug 2023 08:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 07 09:25:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 09:25:41 2023
Response via : Initial Calibration



TIC: BN026821.D\data.ms

(77) Benzidine

19.745min (0.000) 1335663.38 ng

response 5270057

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.24
183.00	13.20	12.34
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/7/2023	BNA_N	<u>BN026821.D</u>
Compound Name	Response	Retention Time
DDT	2313106	20.945
DDD	33384	20.551
DDE	972	19.992
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
34356	2347462	1.46

Instrument :
BNA_N
ClientSampleId :
DFTPP