

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080724\
 Data File : BN033296.D
 Acq On : 07 Aug 2024 23:37
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN080724

Quant Time: Aug 08 04:40:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 08 04:39:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	117	0.00
2	1,4-Dioxane	0.400	0.393	1.8	112	0.00
3	n-Nitrosodimethylamine	0.400	0.382	4.5	113	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	108	0.00
5 S	Phenol-d6	0.400	0.357	10.8	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	117	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.407	-1.7	122	0.00
9	Naphthalene	0.400	0.393	1.8	112	0.00
10	Hexachlorobutadiene	0.400	0.389	2.8	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	112	0.00
12	2-Methylnaphthalene	0.400	0.399	0.3	113	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.400	0.367	8.3	101	0.00
15 S	2-Fluorobiphenyl	0.400	0.432	-8.0	116	0.00
16	Acenaphthylene	0.400	0.431	-7.7	115	0.00
17	Acenaphthene	0.400	0.416	-4.0	112	0.00
18	Fluorene	0.400	0.407	-1.7	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.387	3.3	121	0.00
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	107	0.00
22	Hexachlorobenzene	0.400	0.378	5.5	109	-0.01
23	Atrazine	0.400	0.387	3.3	112	0.00
24	Pentachlorophenol	0.400	0.355	11.3	107	0.00
25	Phenanthrene	0.400	0.395	1.3	112	0.00
26	Anthracene	0.400	0.402	-0.5	116	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	109	0.00
28	Fluoranthene	0.400	0.381	4.8	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
30	Pyrene	0.400	0.390	2.5	111	0.00
31 S	Terphenyl-d14	0.400	0.468	-17.0	133	0.00
32	Benzo(a)anthracene	0.400	0.401	-0.3	116	0.00
33	Chrysene	0.400	0.402	-0.5	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.341	14.8	102	0.00
35 I	Perylene-d12	0.400	0.400	0.0	111	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	115	0.00
37	Benzo(b)fluoranthene	0.400	0.382	4.5	108	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	116	0.00
39 C	Benzo(a)pyrene	0.400	0.413	-3.2	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	114	0.00
41	Benzo(g,h,i)perylene	0.400	0.361	9.8	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN080724