


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080818\
Data File : BN002241.D
Acq On : 08 Aug 2018 08:18
Operator : JU/SJ
Sample : DFTPP97
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP97

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 08 02:41:21 2018



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	187029	PASS
68	69	0.00	2	1.7	3363	PASS
69	198	0.00	100	39.8	195178	PASS
70	69	0.00	2	0.5	975	PASS
127	198	10	80	47.4	232789	PASS
197	198	0.00	2	0.4	1863	PASS
198	198	100	100	100.0	490837	PASS
199	198	5	9	6.9	34074	PASS
275	198	10	60	26.8	131640	PASS
365	198	1	100	3.1	15364	PASS
441	198	0.01	100	13.8	67954	PASS
442	198	50	100	91.5	449024	PASS
443	442	15	24	19.9	89338	PASS

m/z	Abundance
37.00	831.0
38.00	2152.0
39.00	14043.0
40.00	1747.0
41.00	821.0
43.00	501.0
44.00	10852.0
45.00	648.0
49.10	1475.0
50.00	46680.0
51.00	184384.0
52.00	10015.0
53.00	445.0
55.00	1270.0
56.00	6178.0
57.00	13181.0
57.90	594.0
59.00	325.0
61.00	2514.0
62.00	2765.0
63.00	7773.0
63.90	1266.0
65.00	3989.0
66.00	504.0
66.90	489.0
68.00	3260.0
69.00	189440.0
70.00	1164.0
73.00	1470.0
74.00	18216.0
75.00	29416.0
76.00	8944.0
77.00	208640.0
78.00	14197.0
79.00	14037.0
80.00	10556.0
81.00	14866.0
82.00	3812.0
83.00	3485.0
83.90	491.0
85.00	2232.0
86.00	3937.0
87.00	1596.0
87.90	748.0
88.90	386.0
91.00	4042.0
92.00	3834.0
93.00	22528.0
93.90	1350.0
95.00	876.0
96.00	1340.0
97.10	742.0
98.00	15582.0
99.00	14485.0
99.90	1361.0
101.00	8669.0
101.90	431.0
103.00	2408.0
104.00	5107.0
105.00	4624.0
105.90	1509.0
107.00	60992.0
108.00	9860.0
109.00	1906.0
110.00	116936.0
111.00	17720.0
111.90	1983.0
113.00	821.0
116.00	3189.0
117.00	41640.0
118.00	3233.0
119.00	704.0
119.90	941.0
121.00	336.0
122.00	4032.0
123.00	5984.0
123.90	2587.0
125.00	3251.0
127.00	211904.0
128.00	16272.0
129.00	82840.0
130.00	7610.0
131.00	1587.0
132.00	505.0
132.80	353.0
134.00	2430.0
135.00	6815.0
136.00	2814.0
137.00	3604.0
137.80	741.0
139.00	488.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

139.90	1023.0
141.00	10342.0
142.00	3067.0
143.00	2705.0
143.90	900.0
144.90	929.0
145.90	1797.0
147.00	5000.0
148.00	11301.0
148.90	2677.0
149.80	714.0
151.10	1192.0
151.80	1147.0
153.00	3553.0
154.00	2688.0
155.00	5956.0
156.00	8622.0
157.10	1635.0
157.90	2248.0
158.90	1603.0
160.00	3429.0
161.00	5132.0
161.90	1567.0
162.90	418.0
163.90	706.0
165.00	3901.0
166.00	3665.0
167.00	19784.0
168.00	10632.0
168.90	1426.0
170.00	853.0
170.80	965.0
172.00	2050.0
172.90	2676.0
173.90	4085.0
175.00	9085.0
175.90	2159.0
177.00	3627.0
178.00	1172.0
179.00	16023.0
180.00	11554.0
181.00	5619.0
181.90	895.0
183.00	523.0
183.90	1021.0
185.00	7386.0
186.00	57296.0
187.00	16440.0
188.00	2299.0
188.90	3755.0
190.00	571.0
191.00	1854.0
192.00	4271.0
193.00	4864.0
194.00	1202.0
195.10	476.0
196.00	12108.0
197.90	434048.0
198.90	30912.0
200.00	2154.0
201.30	2695.0
203.00	3614.0
204.00	14020.0
205.00	26592.0
206.00	109592.0
207.00	15429.0
208.00	4071.0
208.90	1096.0
211.00	3727.0
211.80	1027.0
214.90	1317.0
216.00	1992.0
216.90	27984.0
218.00	3851.0
219.00	539.0
221.00	15546.0
223.00	6877.0
224.00	58928.0
225.00	15528.0
226.00	1769.0
227.00	22880.0
227.90	3805.0
228.90	4279.0
229.90	859.0
231.00	2292.0
231.80	312.0
233.20	458.0
233.90	1253.0
234.90	1721.0
235.90	1119.0
236.90	2023.0
237.90	444.0
239.00	1184.0
240.10	641.0
241.00	1552.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

242.00	2940.0
243.10	3457.0
244.00	45008.0
245.00	6692.0
246.00	8547.0
247.00	1645.0
247.90	367.0
248.90	1426.0
249.90	584.0
251.00	460.0
253.00	759.0
254.00	1233.0
255.00	238912.0
256.00	37256.0
257.00	3248.0
258.00	12287.0
258.90	2376.0
259.80	378.0
260.90	512.0
264.10	665.0
265.00	4959.0
266.00	707.0
270.90	654.0
272.00	618.0
273.00	6902.0
274.00	18464.0
275.00	105000.0
276.00	14134.0
276.90	7783.0
278.00	1666.0
278.90	483.0
283.10	1322.0
283.80	386.0
284.20	585.0
285.00	1577.0
288.90	455.0
292.10	466.0
293.00	1797.0
294.00	435.0
295.00	745.0
296.00	26056.0
296.90	3340.0
298.00	425.0
300.80	425.0
302.00	423.0
303.00	3020.0
304.00	875.0
307.80	482.0
309.90	381.0
313.00	301.0
313.90	1327.0
315.00	3486.0
315.90	1902.0
317.10	315.0
321.00	630.0
322.10	549.0
323.00	9202.0
324.00	1847.0
326.90	1856.0
328.00	879.0
331.90	606.0
332.90	1129.0
334.00	5353.0
335.00	1539.0
341.00	1189.0
341.80	357.0
346.00	1693.0
346.80	303.0
352.00	2146.0
353.00	1984.0
353.90	2758.0
355.00	637.0
365.00	11706.0
365.90	1578.0
370.10	344.0
371.10	553.0
372.00	4393.0
373.00	985.0
383.00	1060.0
390.00	789.0
390.90	474.0
402.00	1387.0
403.00	2995.0
403.90	1012.0
420.90	2419.0
422.00	2462.0
423.00	14886.0
424.00	4770.0
424.90	592.0
441.00	44896.0
442.00	293760.0
443.00	60872.0
444.00	6129.0
444.80	361.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

m/z	Abundance
37.00	963.0
38.00	2629.0
39.00	16321.0
40.00	1599.0
41.10	1043.0
44.00	10656.0
44.90	515.0
49.10	1662.0
50.00	51912.0
51.00	219136.0
52.00	10016.0
53.00	723.0
55.00	1835.0
56.00	6039.0
57.00	15913.0
58.00	713.0
58.80	333.0
61.00	2913.0
62.00	2955.0
63.00	8753.0
64.10	1243.0
65.00	4854.0
67.00	450.0
68.10	4157.0
69.00	225984.0
70.00	1033.0
73.00	2053.0
74.00	21192.0
75.00	35392.0
76.00	11385.0
77.00	250112.0
78.00	17328.0
79.00	15978.0
80.00	12669.0
81.00	17144.0
82.00	4914.0
83.00	3756.0
84.00	533.0
85.00	2983.0
86.00	4915.0
87.00	2344.0
87.90	698.0
88.90	413.0
90.90	4176.0
92.00	3882.0
93.00	26496.0
94.00	1889.0
95.00	886.0
96.00	1468.0
96.90	488.0
98.00	20664.0
99.00	15501.0
100.00	1460.0
101.00	9672.0
101.90	405.0
102.90	3273.0
104.00	6095.0
105.00	5628.0
106.00	1988.0
107.00	76984.0
108.00	13053.0
109.00	2472.0
110.00	144512.0
111.00	21584.0
112.00	2524.0
113.00	523.0
114.70	301.0
116.00	3976.0
117.00	51912.0
118.00	4214.0
119.00	893.0
120.00	1193.0
121.10	554.0
122.00	5343.0
123.00	7307.0
124.00	3290.0
124.90	2977.0
127.00	273344.0
128.00	19560.0
129.00	104528.0
130.00	8307.0
131.00	1763.0
132.00	1015.0
133.00	556.0
133.90	2900.0
135.00	8258.0
135.90	3683.0
137.00	4273.0
137.90	1184.0
139.00	658.0
139.90	992.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

141.00	13356.0
142.00	4480.0
142.90	2914.0
143.90	858.0
145.00	718.0
146.00	2116.0
146.90	6778.0
148.00	14884.0
149.00	2874.0
150.00	700.0
151.00	1751.0
151.90	1390.0
153.00	4404.0
154.00	3212.0
155.00	6978.0
156.00	10257.0
157.00	2343.0
157.90	2594.0
159.00	2026.0
160.00	4165.0
161.00	6195.0
161.90	1868.0
162.90	878.0
164.10	544.0
164.90	4683.0
166.00	4273.0
167.00	23352.0
168.00	13655.0
169.00	2523.0
170.20	642.0
171.00	951.0
171.80	2228.0
173.00	3036.0
174.00	5796.0
175.00	10665.0
176.00	4011.0
177.00	5111.0
177.90	1476.0
178.90	19144.0
180.00	13067.0
180.90	5696.0
182.00	1231.0
182.80	723.0
184.00	1777.0
185.00	9314.0
186.00	75480.0
187.00	20912.0
188.00	2295.0
189.00	4650.0
189.90	1058.0
191.00	1988.0
192.00	6604.0
193.00	6901.0
194.00	1690.0
194.90	851.0
196.00	16255.0
197.00	3063.0
197.90	568448.0
198.90	39824.0
200.00	3035.0
201.40	2268.0
202.90	4045.0
204.00	20448.0
205.00	32632.0
206.00	148352.0
207.00	20168.0
207.90	4504.0
208.90	1667.0
210.10	2259.0
211.00	5248.0
211.70	1408.0
213.00	386.0
214.90	1604.0
216.00	3051.0
216.90	37592.0
217.90	4524.0
218.90	608.0
221.00	22280.0
221.80	5951.0
223.00	8514.0
224.00	80336.0
225.00	20144.0
226.00	1682.0
227.00	32040.0
228.00	4133.0
228.90	7055.0
230.00	899.0
231.00	2916.0
232.00	380.0
233.00	700.0
234.00	2221.0
234.90	2302.0
235.90	1159.0
236.90	2505.0
237.90	436.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

238.90	1090.0
240.00	839.0
241.00	2163.0
242.00	3560.0
243.00	4148.0
244.00	59544.0
245.00	8640.0
245.90	11802.0
246.90	2540.0
247.80	648.0
248.90	2212.0
250.00	627.0
251.00	336.0
251.90	594.0
252.90	1657.0
255.00	334720.0
256.00	49384.0
257.00	4116.0
257.90	18032.0
259.00	2740.0
259.90	445.0
260.90	479.0
264.10	649.0
265.00	7357.0
265.90	1269.0
269.90	328.0
270.90	698.0
272.10	818.0
273.00	9990.0
274.00	25040.0
275.00	151744.0
276.00	19504.0
277.00	11432.0
278.00	2119.0
278.90	301.0
282.00	351.0
282.90	1402.0
284.00	631.0
284.90	1918.0
286.00	476.0
288.90	410.0
289.90	587.0
291.00	377.0
291.90	841.0
293.00	2400.0
293.90	917.0
294.90	736.0
296.00	38840.0
297.00	5907.0
298.10	361.0
301.00	526.0
302.00	624.0
303.00	5128.0
304.00	1293.0
307.90	839.0
309.00	483.0
309.90	623.0
312.90	353.0
314.00	2122.0
314.90	4843.0
316.00	2563.0
316.90	612.0
321.00	1243.0
322.00	836.0
323.00	12614.0
324.00	2611.0
324.80	329.0
326.90	2293.0
328.00	1346.0
332.00	1010.0
333.00	1421.0
334.00	8499.0
335.00	2310.0
339.00	435.0
340.00	322.0
341.00	1498.0
342.00	472.0
345.90	2720.0
347.00	714.0
352.00	4622.0
353.00	3055.0
354.00	4402.0
354.90	888.0
358.90	424.0
364.90	18120.0
365.90	2462.0
370.90	1069.0
372.00	7121.0
373.00	1873.0
373.90	491.0
382.90	2298.0
384.10	541.0
390.00	974.0
391.00	876.0
392.00	680.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

400.90	592.0
401.90	3406.0
403.00	3930.0
404.00	1724.0
414.80	383.0
421.00	4155.0
422.00	3336.0
423.00	26768.0
424.00	6347.0
425.10	535.0
441.00	80752.0
442.00	511296.0
443.00	100664.0
444.00	10305.0
444.90	613.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

m/z	Abundance
36.90	448.0
38.00	1705.0
39.00	12068.0
40.00	1592.0
41.00	980.0
42.90	404.0
44.00	10992.0
45.00	437.0
49.20	886.0
50.00	38928.0
51.00	157568.0
52.00	7939.0
53.10	527.0
55.00	1193.0
56.00	4949.0
57.00	11932.0
57.90	568.0
61.00	2083.0
62.00	1993.0
63.00	6678.0
64.00	874.0
65.00	3564.0
66.90	483.0
68.00	2674.0
69.00	170112.0
70.10	728.0
72.10	321.0
73.00	1376.0
74.00	16284.0
75.00	26296.0
76.00	7635.0
77.00	188160.0
78.00	12311.0
79.00	11776.0
79.90	9640.0
81.00	13915.0
82.00	3405.0
83.00	3522.0
84.00	472.0
84.90	2252.0
86.00	3716.0
87.00	2015.0
88.00	556.0
89.00	533.0
91.00	3698.0
92.00	3318.0
93.00	19904.0
94.00	1469.0
95.00	611.0
96.00	1060.0
97.00	640.0
98.00	15751.0
99.00	11547.0
99.90	1123.0
101.00	8217.0
102.10	362.0
102.90	2117.0
103.90	4653.0
104.90	4593.0
106.10	1760.0
107.00	59432.0
108.00	9416.0
109.00	2016.0
110.00	106896.0
111.00	15850.0
112.00	1823.0
112.90	495.0
114.80	557.0
116.00	3410.0
117.00	42432.0
117.90	2998.0
119.00	581.0
120.00	649.0
121.00	415.0
121.90	4298.0
123.00	6042.0
124.00	2648.0
125.00	2613.0
127.00	213120.0
128.00	15933.0
129.00	82824.0
129.90	6469.0
130.90	1457.0
132.10	819.0
133.10	394.0
134.00	2492.0
135.00	7176.0
135.90	2299.0
137.00	3272.0
137.80	708.0
138.80	664.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

139.90	1150.0
140.90	9837.0
141.90	3187.0
142.90	1876.0
143.90	620.0
145.00	868.0
145.90	1680.0
147.00	5475.0
148.00	12289.0
149.00	2321.0
149.90	758.0
151.10	1315.0
151.80	722.0
153.00	3544.0
154.00	2545.0
155.00	5772.0
156.00	8573.0
157.00	1613.0
157.90	2063.0
158.90	1582.0
160.00	3158.0
161.00	5040.0
161.90	1771.0
163.00	585.0
164.00	693.0
165.00	4212.0
166.00	3295.0
167.00	20880.0
168.00	10955.0
168.90	1694.0
169.90	655.0
170.90	973.0
171.90	1873.0
173.00	2748.0
174.00	4477.0
175.00	8588.0
176.00	2037.0
177.00	3860.0
178.00	1507.0
179.00	14929.0
180.00	11402.0
181.00	5654.0
182.00	746.0
183.00	414.0
183.90	1117.0
185.00	8089.0
186.00	63336.0
187.00	19032.0
188.00	2262.0
189.00	4027.0
190.00	640.0
191.00	1918.0
191.90	5017.0
193.00	6476.0
194.00	1268.0
196.00	12623.0
196.80	2527.0
197.90	470016.0
198.90	31488.0
200.00	2723.0
201.30	2075.0
202.90	3247.0
204.00	17656.0
205.00	27896.0
206.00	122936.0
207.00	15927.0
208.00	3946.0
209.00	1439.0
210.20	2316.0
210.90	5531.0
211.80	883.0
212.80	518.0
215.00	1254.0
215.90	2703.0
216.90	31896.0
218.00	4064.0
218.90	546.0
221.00	19784.0
221.70	4384.0
223.00	7493.0
224.00	65816.0
225.00	17864.0
225.90	1817.0
227.00	28312.0
228.00	3066.0
229.00	6415.0
230.00	676.0
231.00	2473.0
232.00	501.0
233.00	637.0
233.80	1791.0
235.00	1895.0
235.90	1307.0
237.00	2310.0
238.00	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

239.00	732.0
239.90	1065.0
240.90	1587.0
242.00	3338.0
243.00	3417.0
244.00	53736.0
245.00	6597.0
246.00	10929.0
246.90	2004.0
247.90	509.0
249.00	2048.0
250.00	306.0
250.90	485.0
252.00	660.0
252.90	1089.0
254.00	2350.0
255.00	288448.0
256.00	42136.0
257.00	3983.0
258.00	15351.0
259.00	2074.0
260.00	363.0
261.00	537.0
263.80	310.0
265.00	6324.0
265.90	858.0
270.10	426.0
270.90	681.0
271.90	895.0
272.90	8339.0
274.00	21560.0
275.00	138176.0
276.00	17968.0
276.90	11734.0
277.90	1473.0
278.90	422.0
283.00	1306.0
284.10	977.0
285.00	2210.0
288.90	486.0
290.00	383.0
290.90	430.0
291.90	624.0
293.00	2447.0
293.80	774.0
295.10	771.0
296.00	35656.0
297.00	4686.0
298.00	418.0
300.90	606.0
301.90	704.0
303.00	4669.0
303.90	1230.0
308.00	487.0
309.90	563.0
312.90	402.0
314.00	1416.0
315.00	4654.0
315.90	2719.0
316.90	466.0
321.00	1299.0
322.00	657.0
323.00	12231.0
324.00	1701.0
325.90	315.0
327.00	2274.0
328.00	842.0
332.10	811.0
333.00	1236.0
334.00	7980.0
335.00	1832.0
339.10	345.0
340.10	348.0
341.00	1500.0
342.00	546.0
345.90	2695.0
347.00	335.0
352.00	3499.0
353.00	2269.0
354.00	4463.0
355.00	732.0
358.90	411.0
364.90	16266.0
366.00	2021.0
370.90	960.0
372.00	7480.0
373.00	1823.0
382.90	1890.0
384.00	684.0
389.90	1208.0
390.80	464.0
392.00	500.0
401.10	354.0
402.00	2990.0
403.00	4601.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

404.00	1482.0
421.00	3200.0
422.00	3475.0
423.00	27096.0
424.00	6088.0
424.90	593.0
441.00	78216.0
442.00	542016.0
443.00	106480.0
444.00	10327.0
445.00	547.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97

Scan 2463 (17.475 min): BN002241.D
DFTPP97

m/z	Abundance
40.00	1114.0
41.00	361.0
44.00	12584.0
72.90	387.0
81.00	345.0
95.00	525.0
105.00	335.0
206.90	1020.0

Instrument :
BNA_N
ClientSampleId :
DFTPP97