

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015826.D
 Acq On : 09 Aug 2021 16:49
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Aug 09 17:27:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8919	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	42061	0.400	ng	0.00
13) Acenaphthene-d10	14.560	164	24278	0.400	ng	0.00
19) Phenanthrene-d10	17.309	188	48129	0.400	ng	0.00
29) Chrysene-d12	21.493	240	51544	0.400	ng	0.00
35) Perylene-d12	23.926	264	49495	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.484	112	18356	0.699	ng	0.00
5) Phenol-d6	7.080	99	24789	0.685	ng	0.00
8) Nitrobenzene-d5	9.076	82	24694	0.780	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	54504	0.844	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	8317	0.656	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	75256	0.831	ng	0.00
27) Fluoranthene-d10	19.336	212	117779	0.873	ng	0.00
31) Terphenyl-d14	19.936	244	118901	0.748	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	2958	0.823	ng	97
3) n-Nitrosodimethylamine	3.769	42	7501	0.887	ng	# 99
6) bis(2-Chloroethyl)ether	7.347	93	19802	0.770	ng	99
9) Naphthalene	10.774	128	87588	0.794	ng	100
10) Hexachlorobutadiene	11.065	225	23004	1.144	ng	# 99
12) 2-Methylnaphthalene	12.389	142	63986	0.855	ng	100
16) Acenaphthylene	14.281	152	82265	0.727	ng	100
17) Acenaphthene	14.624	154	54285	0.763	ng	99
18) Fluorene	15.601	166	68138	0.790	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	3423	0.910	ng	# 88
21) 4-Bromophenyl-phenylether	16.494	248	20752	0.806	ng	99
22) Hexachlorobenzene	16.616	284	26750	0.843	ng	100
23) Atrazine	16.762	200	17591	0.666	ng	99
24) Pentachlorophenol	16.956	266	9678	0.808	ng	99
25) Phenanthrene	17.346	178	107543	0.783	ng	99
26) Anthracene	17.431	178	97389	0.731	ng	100
28) Fluoranthene	19.370	202	131383	0.855	ng	100
30) Pyrene	19.733	202	129353	0.709	ng	100
32) Benzo(a)anthracene	21.482	228	132351	0.752	ng	100
33) Chrysene	21.535	228	135133	0.813	ng	100
34) Bis(2-ethylhexyl)phtha...	21.408	149	53782	0.414	ng	99
36) Indeno(1,2,3-cd)pyrene	26.441	276	138077	0.739	ng	100
37) Benzo(b)fluoranthene	23.183	252	139273	0.803	ng	99
38) Benzo(k)fluoranthene	23.234	252	128086	0.784	ng	99
39) Benzo(a)pyrene	23.816	252	118850	0.775	ng	98
40) Dibenzo(a,h)anthracene	26.465	278	113242	0.741	ng	99
41) Benzo(g,h,i)perylene	27.215	276	120616	0.747	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015826.D
 Acq On : 09 Aug 2021 16:49
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Aug 09 17:27:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.928 min Scan# 526
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	152	Resp:	8919
Ion	Ratio	Lower	Upper
152	100		
150	153.7	123.0	184.4
115	64.8	52.2	78.2



1,4-Dioxane
 Concen: 0.823 ng
 RT: 3.410 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49

Tgt Ion:	88	Resp:	2958
Ion	Ratio	Lower	Upper
88	100		
43	37.2	27.4	41.2
58	82.9	64.3	96.5









n-Nitrosodimethylamine
 Concen: 0.887 ng
 RT: 3.769 min Scan# 75
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	42	Resp:	7501
Ion	Ratio	Lower	Upper
42	100		
74	102.9	83.4	125.2
44	7.5	0.0	0.0#



2-Fluorophenol
 Concen: 0.699 ng
 RT: 5.484 min Scan# 261
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49

Tgt Ion:	112	Resp:	18356
Ion	Ratio	Lower	Upper
112	100		
64	67.2	53.4	80.0
63	36.4	28.6	43.0









Phenol-d6
 Concen: 0.685 ng
 RT: 7.080 min Scan# 434
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	99	Resp:	24789
Ion	Ratio	Lower	Upper
99	100		
42	23.4	19.0	28.4
71	39.5	31.2	46.8




bis(2-Chloroethyl)ether
 Concen: 0.770 ng
 RT: 7.347 min Scan# 463
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49

Tgt Ion:	93	Resp:	19802
Ion	Ratio	Lower	Upper
93	100		
63	80.1	64.3	96.5
95	32.8	25.6	38.4





Abundance Scan 758 (10.724 min): BN015825.D\data.ms (-757) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.723 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	42061
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.8	13.2
54	9.0	7.1	10.7
68	6.0	4.9	7.3



Abundance Scan 758 (10.723 min): BN015826.D\data.ms (-742) (-)





Abundance Scan 628 (9.076 min): BN015825.D\data.ms (-625) (-)




Nitrobenzene-d5
Concen: 0.780 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	82	Resp:	24694
Ion Ratio	Lower	Upper	
82	100		
128	33.2	26.2	39.2
54	56.4	45.4	68.0




Abundance Scan 628 (9.076 min): BN015826.D\data.ms (-612) (-)






Abundance Scan 762 (10.774 min): BN015825.D\data.ms (-756) (-)

Naphthalene
Concen: 0.794 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:	128	Resp:	87588
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.4	10.8	16.2






Abundance Scan 785 (11.066 min): BN015825.D\data.ms (-779) (-)


Hexachlorobutadiene
Concen: 1.144 ng
RT: 11.065 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49



Tgt Ion:	225	Resp:	23004
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.8	77.8






Abundance Scan 1001 (12.314 min): BN015825.D\data.ms (-971) (-)




2-Methylnaphthalene-d10
Concen: 0.844 ng
RT: 12.314 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:152 Resp: 54504
Ion Ratio Lower Upper
152 100
151 20.4 16.4 24.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Abundance Scan 1018 (12.389 min): BN015825.D\data.ms (-1012) (-)




2-Methylnaphthalene
Concen: 0.855 ng
RT: 12.389 min Scan# 1018
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:142 Resp: 63986
Ion Ratio Lower Upper
142 100
141 88.1 70.6 106.0
115 36.6 29.1 43.7






Abundance Scan 1206 (14.561 min): BN015825.D\data.ms (-1206) (-)




Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	164	Resp:	24278
Ion	Ratio	Lower	Upper
164	100		
162	101.1	81.2	121.8
160	43.9	35.0	52.4







Abundance Scan 1323 (16.044 min): BN015825.D\data.ms (-1319) (-)



2,4,6-Tribromophenol
Concen: 0.656 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	330	Resp:	8317
Ion	Ratio	Lower	Upper
330	100		
332	99.8	86.2	129.2
141	30.7	25.0	37.6









Scan 1098 (13.190 min): BN015825.D\data.ms (-1098) (-)
 2-Fluorobiphenyl
 Concen: 0.831 ng
 RT: 13.190 min Scan# 1098
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	172	Resp:	75256
Ion Ratio	Lower	Upper	
172	100		
171	34.1	26.9	40.3
170	21.6	17.3	25.9



Scan 1184 (14.281 min): BN015825.D\data.ms (-1184) (-)
 Acenaphthylene
 Concen: 0.727 ng
 RT: 14.281 min Scan# 1184
 Delta R.T. -0.000 min
 Lab File: BN015826.D
 Acq: 09 Aug 2021 16:49


Tgt Ion:	152	Resp:	82265
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.6	23.4
153	13.0	10.5	15.7






Abundance Scan 1211 (14.624 min): BN015825.D\data.ms (-1205) (-)

Acenaphthene
Concen: 0.763 ng
RT: 14.624 min Scan# 1211
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion	Ratio	Lower	Upper
154	100		
153	111.7	90.1	135.1
152	53.0	42.7	64.1






Abundance Scan 1288 (15.601 min): BN015825.D\data.ms (-1214) (-)


Fluorene
Concen: 0.790 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49



Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.0	78.8	118.2
167	13.5	10.9	16.3





Abundance Scan 1427 (17.310 min): BN015825.D\data.ms (-1419) (-)




Phenanthrene-d10
Concen: 0.400 ng
RT: 17.309 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	188	Resp:	48129
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	5.9	4.7	7.1



Abundance Scan 1427 (17.309 min): BN015826.D\data.ms (-1411) (-)





Abundance Scan 1294 (15.677 min): BN015825.D\data.ms (-1290) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.910 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	198	Resp:	3423
Ion Ratio	Lower	Upper	
198	100		
182	8.7	10.8	16.2#
77	0.0	0.0	0.0




Abundance Scan 1294 (15.677 min): BN015826.D\data.ms (-1286) (-)





Abundance Scan 1360 (16.494 min): BN015825.D\data.ms (-1352) (-)

4-Bromophenyl-phenylether
Concen: 0.806 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion:	248	Resp:	20752
Ion	Ratio	Lower	Upper
248	100		
250	93.5	74.8	112.2
141	74.5	60.5	90.7





Abundance Scan 1360 (16.494 min): BN015826.D\data.ms (-1352) (-)




Abundance Scan 1370 (16.616 min): BN015825.D\data.ms (-1362) (-)

Hexachlorobenzene
Concen: 0.843 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	284	Resp:	26750
Ion	Ratio	Lower	Upper
284	100		
142	28.5	23.0	34.4
249	31.7	25.3	37.9




Abundance Scan 1370 (16.616 min): BN015826.D\data.ms (-1362) (-)





Abundance Scan 1382 (16.762 min): BN015825.D\data.ms (-1328) (-)

Atrazine
Concen: 0.666 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.5	19.4	29.2
215	50.4	39.4	59.2






Abundance Scan 1382 (16.762 min): BN015826.D\data.ms (-1341) (-)




Abundance Scan 1398 (16.957 min): BN015825.D\data.ms (-1324) (-)

Pentachlorophenol
Concen: 0.808 ng
RT: 16.956 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.5	51.0	76.6
268	63.8	51.8	77.6



Abundance Scan 1398 (16.956 min): BN015826.D\data.ms (-1390) (-)





Abundance Scan 1430 (17.346 min): BN015825.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.783 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 107543
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.4 12.7 19.1



Abundance Scan 1430 (17.346 min): BN015826.D\data.ms (-1405) (-)





Abundance Scan 1437 (17.431 min): BN015825.D\data.ms (-1426) (-)




Anthracene
Concen: 0.731 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion:178 Resp: 97389
Ion Ratio Lower Upper
178 100
176 17.9 14.3 21.5
179 15.3 12.3 18.5



Abundance Scan 1437 (17.431 min): BN015826.D\data.ms (-1429) (-)





Abundance Scan 1752 (19.341 min): BN015825.D\data.ms (-1732) (-)




Fluoranthene-d10
Concen: 0.873 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.005 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	212	Resp:	117779
Ion Ratio	Lower	Upper	
212	100		
106	9.6	7.8	11.8
104	5.6	4.6	6.8



Abundance Scan 1751 (19.336 min): BN015826.D\data.ms (-1732) (-)





Abundance Scan 1758 (19.371 min): BN015825.D\data.ms (-1732) (-)




Fluoranthene
Concen: 0.855 ng
RT: 19.370 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	202	Resp:	131383
Ion Ratio	Lower	Upper	
202	100		
101	8.7	7.0	10.6
203	17.4	13.8	20.8




Abundance Scan 1758 (19.370 min): BN015826.D\data.ms (-1718) (-)






Abundance Scan 2026 (21.493 min): BN015825.D\data.ms (-2029) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:	240	Resp:	51544
Ion	Ratio	Lower	Upper
240	100		
120	8.3	7.8	11.8
236	26.0	21.1	31.7






Abundance Scan 1831 (19.733 min): BN015825.D\data.ms (-1830) (-)

Pyrene
Concen: 0.709 ng
RT: 19.733 min Scan# 1831
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Tgt Ion:	202	Resp:	129353
Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.2	24.2
203	17.5	14.1	21.1





Abundance Scan 1872 (19.937 min): BN015825.D\data.ms (-1842) (-)

Terphenyl-d14
Concen: 0.748 ng
RT: 19.936 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.1	9.1
122	8.5	7.0	10.4







Abundance Scan 1872 (19.936 min): BN015826.D\data.ms (-1842) (-)




Abundance Scan 2025 (21.482 min): BN015825.D\data.ms (-2066) (-)

Benzo(a)anthracene
Concen: 0.752 ng
RT: 21.482 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.7	20.6	31.0
229	19.9	15.8	23.6









Chrysene
Concen: 0.813 ng
RT: 21.535 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.1	22.7	34.1
229	19.8	15.8	23.8



Bis(2-ethylhexyl)phthalate
Concen: 0.414 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.011 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.9	22.7	34.1
279	5.2	4.2	6.4






Abundance Scan 2556 (23.926 min): BN015825.D\data.ms (-2535) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.926 min Scan# 2556
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.6	29.4
265	26.1	21.7	32.5






Abundance Scan 3292 (26.445 min): BN015825.D\data.ms (-3236) (-)


Indeno(1,2,3-cd)pyrene
Concen: 0.739 ng
RT: 26.441 min Scan# 3291
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49



Tgt Ion	Ratio	Lower	Upper
276	100		
138	18.1	14.5	21.7
277	25.2	20.2	30.4





Abundance Scan 2340 (23.187 min): BN015825.D\data.ms (-2327) (-)




Benzo(b)fluoranthene
Concen: 0.803 ng
RT: 23.183 min Scan# 2339
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	252	Resp:	139273
Ion Ratio	Lower	Upper	
252	100		
253	22.2	18.2	27.2
125	7.7	6.3	9.5



Abundance Scan 2339 (23.183 min): BN015826.D\data.ms (-2282) (-)





Abundance Scan 2354 (23.235 min): BN015825.D\data.ms (-2348) (-)




Benzo(k)fluoranthene
Concen: 0.784 ng
RT: 23.234 min Scan# 2354
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion:	252	Resp:	128086
Ion Ratio	Lower	Upper	
252	100		
253	22.3	18.1	27.1
125	7.3	6.2	9.4



Abundance Scan 2354 (23.234 min): BN015826.D\data.ms (-2340) (-)





Abundance Scan 2525 (23.820 min): BN015825.D\data.ms (-2509) (-)




Benzo(a)pyrene
Concen: 0.775 ng
RT: 23.816 min Scan# 2524
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	252	Resp:	118850
Ion Ratio	Lower	Upper	
252	100		
253	22.3	18.5	27.7
125	8.4	7.0	10.6



Abundance Scan 2524 (23.816 min): BN015826.D\data.ms (-2467) (-)





Abundance Scan 3297 (26.462 min): BN015825.D\data.ms (-3274) (-)




Dibenzo(a,h)anthracene
Concen: 0.741 ng
RT: 26.465 min Scan# 3298
Delta R.T. 0.003 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion:	278	Resp:	113242
Ion Ratio	Lower	Upper	
278	100		
139	12.7	10.5	15.7
279	24.2	19.5	29.3



Abundance Scan 3298 (26.465 min): BN015826.D\data.ms (-3268) (-)




Abundance Scan 3516 (27.212 min): BN015825.D\data.ms (-3458)



Benzo(g,h,i)perylene
Concen: 0.747 ng
RT: 27.215 min Scan# 3517
Delta R.T. 0.003 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Abundance Scan 3517 (27.215 min): BN015826.D\data.ms



Tgt Ion:	276	Resp:	120616
Ion Ratio		Lower	Upper
276	100		
277	24.0	19.0	28.6
138	15.3	12.2	18.4

Abundance Scan 3517 (27.215 min): BN015826.D\data.ms (-3458)

