

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015830.D
 Acq On : 09 Aug 2021 19:52
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN080921

Quant Time: Aug 10 05:24:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 19:13:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.373	6.8	81	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	100	0.00
4 S	2-Fluorophenol	0.400	0.395	1.3	99	0.00
5 S	Phenol-d6	0.400	0.396	1.0	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.372	7.0	108	0.00
9	Naphthalene	0.400	0.384	4.0	99	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	101	0.00
12	2-Methylnaphthalene	0.400	0.381	4.8	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.322	19.5	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	101	0.00
16	Acenaphthylene	0.400	0.364	9.0	101	0.00
17	Acenaphthene	0.400	0.373	6.8	102	0.00
18	Fluorene	0.400	0.372	7.0	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.412	-3.0	101	0.00
21	4-Bromophenyl-phenylether	0.400	0.349	12.8	95	0.00
22	Hexachlorobenzene	0.400	0.385	3.8	105	0.00
23	Atrazine	0.400	0.353	11.8	106	0.00
24	Pentachlorophenol	0.400	0.330	17.5	109	0.00
25	Phenanthrene	0.400	0.375	6.3	103	0.00
26	Anthracene	0.400	0.368	8.0	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	106	0.00
28	Fluoranthene	0.400	0.380	5.0	103	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	108	0.00
30	Pyrene	0.400	0.370	7.5	104	0.00
31 S	Terphenyl-d14	0.400	0.377	5.8	109	0.00
32	Benzo(a)anthracene	0.400	0.377	5.8	107	-0.01
33	Chrysene	0.400	0.377	5.8	105	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.321	19.8	104	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.373	6.8	113	0.00
37	Benzo(b)fluoranthene	0.400	0.377	5.8	111	0.00
38	Benzo(k)fluoranthene	0.400	0.385	3.8	108	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	110	0.00
40	Dibenzo(a,h)anthracene	0.400	0.378	5.5	115	0.00
41	Benzo(g,h,i)perylene	0.400	0.368	8.0	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN080921