

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015823.D
 Acq On : 09 Aug 2021 14:59
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Aug 09 17:25:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8614	0.400	ng	0.00
7) Naphthalene-d8	10.724	136	41158	0.400	ng	0.00
13) Acenaphthene-d10	14.561	164	22987	0.400	ng	0.00
19) Phenanthrene-d10	17.310	188	45569	0.400	ng	0.00
29) Chrysene-d12	21.493	240	45961	0.400	ng	0.00
35) Perylene-d12	23.926	264	44069	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	2655	0.105	ng	0.00
5) Phenol-d6	7.080	99	3643	0.104	ng	0.00
8) Nitrobenzene-d5	9.076	82	3339	0.108	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	6579	0.104	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	1028	0.086	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	11173	0.130	ng	0.00
27) Fluoranthene-d10	19.336	212	13225	0.104	ng	0.00
31) Terphenyl-d14	19.937	244	15793	0.111	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	332	0.096	ng	89
3) n-Nitrosodimethylamine	3.788	42	775	0.095	ng	# 94
6) bis(2-Chloroethyl)ether	7.347	93	2629	0.106	ng	100
9) Naphthalene	10.774	128	11859	0.110	ng	98
10) Hexachlorobutadiene	11.066	225	2926	0.149	ng	# 97
12) 2-Methylnaphthalene	12.390	142	7922	0.108	ng	99
16) Acenaphthylene	14.282	152	9931	0.093	ng	100
17) Acenaphthene	14.624	154	6917	0.103	ng	98
18) Fluorene	15.601	166	8713	0.107	ng	99
20) 4,6-Dinitro-2-methylph...	15.690	198	222	0.062	ng	# 29
21) 4-Bromophenyl-phenylether	16.494	248	2615	0.107	ng	100
22) Hexachlorobenzene	16.616	284	3461	0.115	ng	# 87
23) Atrazine	16.762	200	2023	0.081	ng	98
24) Pentachlorophenol	16.957	266	1147	0.101	ng	97
25) Phenanthrene	17.346	178	14059	0.108	ng	99
26) Anthracene	17.431	178	12063	0.096	ng	99
28) Fluoranthene	19.371	202	16212	0.111	ng	100
30) Pyrene	19.728	202	16390	0.101	ng	100
32) Benzo(a)anthracene	21.482	228	15865	0.101	ng	99
33) Chrysene	21.536	228	16765	0.113	ng	100
34) Bis(2-ethylhexyl)phtha...	21.408	149	6963	0.060	ng	98
36) Indeno(1,2,3-cd)pyrene	26.442	276	15108	0.091	ng	# 96
37) Benzo(b)fluoranthene	23.183	252	16811	0.109	ng	# 92
38) Benzo(k)fluoranthene	23.231	252	15410	0.106	ng	# 93
39) Benzo(a)pyrene	23.817	252	14192	0.104	ng	# 90
40) Dibenzo(a,h)anthracene	26.466	278	11904	0.087	ng	# 82
41) Benzo(g,h,i)perylene	27.216	276	14120	0.098	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
Data File : BN015823.D
Acq On : 09 Aug 2021 14:59
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Aug 09 17:25:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 16:46:20 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.928 min Scan# 526
 Delta R.T. 0.000 min
 Lab File: BN015823.D
 Acq: 09 Aug 2021 14:59

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion: 152 Resp: 8614




Ion	Ratio	Lower	Upper
152	100		
150	152.8	123.0	184.4
115	64.7	52.2	78.2




#2
 1,4-Dioxane
 Concen: 0.096 ng
 RT: 3.410 min Scan# 36
 Delta R.T. 0.000 min
 Lab File: BN015823.D
 Acq: 09 Aug 2021 14:59

Tgt Ion: 88 Resp: 332

Ion	Ratio	Lower	Upper
88	100		
43	40.4	27.4	41.2
58	90.1	64.3	96.5









#3
n-Nitrosodimethylamine
Concen: 0.095 ng
RT: 3.788 min Scan# 77
Delta R.T. 0.019 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 775
Ion Ratio Lower Upper
42 100
74 98.6 83.4 125.2
44 4.0 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.105 ng
RT: 5.485 min Scan# 261
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion: 112 Resp: 2655
Ion Ratio Lower Upper
112 100
64 69.5 53.4 80.0
63 36.9 28.6 43.0









#5
Phenol-d6
Concen: 0.104 ng
RT: 7.080 min Scan# 434
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion: 99 Resp: 3643
Ion Ratio Lower Upper
99 100
42 24.6 19.0 28.4
71 38.5 31.2 46.8



#6
bis(2-Chloroethyl)ether
Concen: 0.106 ng
RT: 7.347 min Scan# 463
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion: 93 Resp: 2629
Ion Ratio Lower Upper
93 100
63 80.4 64.3 96.5
95 32.9 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.724 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	41158
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.8	13.2
54	8.8	7.1	10.7
68	5.9	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.108 ng
RT: 9.076 min Scan# 628
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	82	Resp:	3339
Ion	Ratio	Lower	Upper
82	100		
128	35.5	26.2	39.2
54	59.0	45.4	68.0









#9
Naphthalene
Concen: 0.110 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	128	Resp:	11859
Ion	Ratio	Lower	Upper
128	100		
129	11.8	8.9	13.3
127	14.1	10.8	16.2



#10
Hexachlorobutadiene
Concen: 0.149 ng
RT: 11.066 min Scan# 785
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	225	Resp:	2926
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.8	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 0.104 ng
RT: 12.314 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 6579
Ion Ratio Lower Upper
152 100
151 20.6 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.108 ng
RT: 12.390 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:142 Resp: 7922
Ion Ratio Lower Upper
142 100
141 87.7 70.6 106.0
115 37.7 29.1 43.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:164 Resp: 22987
Ion Ratio Lower Upper
164 100
162 101.0 81.2 121.8
160 43.6 35.0 52.4



#14
2,4,6-Tribromophenol
Concen: 0.086 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:330 Resp: 1028
Ion Ratio Lower Upper
330 100
332 133.2 86.2 129.2#
141 32.3 25.0 37.6









#15
2-Fluorobiphenyl
Concen: 0.130 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	172	Resp:	11173
Ion	Ratio	Lower	Upper
172	100		
171	35.2	26.9	40.3
170	22.3	17.3	25.9



#16
Acenaphthylene
Concen: 0.093 ng
RT: 14.282 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	152	Resp:	9931
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.6	23.4
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.103 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:154 Resp: 6917
Ion Ratio Lower Upper
154 100
153 114.9 90.1 135.1
152 55.1 42.7 64.1



#18
Fluorene
Concen: 0.107 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:166 Resp: 8713
Ion Ratio Lower Upper
166 100
165 97.8 78.8 118.2
167 14.0 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.310 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	188	Resp:	45569
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	5.9	4.7	7.1



#20
4,6-Dinitro-2-methylphenol
Concen: 0.062 ng
RT: 15.690 min Scan# 1295
Delta R.T. 0.013 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	198	Resp:	222
Ion	Ratio	Lower	Upper
198	100		
182	42.0	10.8	16.2
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.107 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion:	248	Resp:	2615
Ion	Ratio	Lower	Upper
248	100		
250	92.8	74.8	112.2
141	75.4	60.5	90.7



#22
Hexachlorobenzene
Concen: 0.115 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	284	Resp:	3461
Ion	Ratio	Lower	Upper
284	100		
142	43.0	23.0	34.4#
249	31.8	25.3	37.9









#23
Atrazine
Concen: 0.081 ng
RT: 16.762 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	200	Resp:	2023
Ion	Ratio	Lower	Upper
200	100		
173	25.4	19.4	29.2
215	50.3	39.4	59.2



#24
Pentachlorophenol
Concen: 0.101 ng
RT: 16.957 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:	266	Resp:	1147
Ion	Ratio	Lower	Upper
266	100		
264	64.9	51.0	76.6
268	67.9	51.8	77.6









#25
Phenanthrene
Concen: 0.108 ng
RT: 17.346 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 14059
Ion Ratio Lower Upper
178 100
176 18.7 14.6 22.0
179 16.0 12.7 19.1



#26
Anthracene
Concen: 0.096 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:178 Resp: 12063
Ion Ratio Lower Upper
178 100
176 18.3 14.3 21.5
179 15.3 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.104 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.005 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 13225
Ion Ratio Lower Upper
212 100
106 10.0 7.8 11.8
104 6.0 4.6 6.8



#28
Fluoranthene
Concen: 0.111 ng
RT: 19.371 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:202 Resp: 16212
Ion Ratio Lower Upper
202 100
101 9.0 7.0 10.6
203 17.2 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 45961
Ion Ratio Lower Upper
240 100
120 8.8 7.8 11.8
236 26.3 21.1 31.7



#30
Pyrene
Concen: 0.101 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.005 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:202 Resp: 16390
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.2
203 17.9 14.1 21.1









#31
Terphenyl-d14
Concen: 0.111 ng
RT: 19.937 min Scan# 1872
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:244 Resp: 15793
Ion Ratio Lower Upper
244 100
212 8.0 6.1 9.1
122 9.1 7.0 10.4



#32
Benzo(a)anthracene
Concen: 0.101 ng
RT: 21.482 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:228 Resp: 15865
Ion Ratio Lower Upper
228 100
226 25.5 20.6 31.0
229 20.4 15.8 23.6









#33
Chrysene
Concen: 0.113 ng
RT: 21.536 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	228	Resp:	16765
Ion	Ratio	Lower	Upper
228	100		
226	28.4	22.7	34.1
229	20.1	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.060 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.010 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59


Tgt Ion:	149	Resp:	6963
Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.7	34.1
279	5.0	4.2	6.4







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.926 min Scan# 2556
Delta R.T. 0.000 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:264 Resp: 44069

Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.6	29.4
265	27.6	21.7	32.5






#36
Indeno(1,2,3-cd)pyrene
Concen: 0.091 ng
RT: 26.442 min Scan# 3291
Delta R.T. -0.003 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59



Tgt Ion:276 Resp: 15108

Ion	Ratio	Lower	Upper
276	100		
138	22.5	14.5	21.7#
277	25.4	20.2	30.4









#37
Benzo(b)fluoranthene
Concen: 0.109 ng
RT: 23.183 min Scan# 2339
Delta R.T. -0.003 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 16811
Ion Ratio Lower Upper
252 100
253 26.1 18.2 27.2
125 11.3 6.3 9.5#



#38
Benzo(k)fluoranthene
Concen: 0.106 ng
RT: 23.231 min Scan# 2353
Delta R.T. -0.003 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:252 Resp: 15410
Ion Ratio Lower Upper
252 100
253 25.9 18.1 27.1
125 10.7 6.2 9.4#









#39
Benzo(a)pyrene
Concen: 0.104 ng
RT: 23.817 min Scan# 2524
Delta R.T. -0.003 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252	Resp:	14192
Ion Ratio	Lower	Upper
252	100	
253	27.3	18.5 27.7
125	13.4	7.0 10.6#



#40
Dibenzo(a,h)anthracene
Concen: 0.087 ng
RT: 26.466 min Scan# 3298
Delta R.T. 0.004 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Tgt Ion:278	Resp:	11904
Ion Ratio	Lower	Upper
278	100	
139	28.3	10.5 15.7#
279	28.0	19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.098 ng
RT: 27.216 min Scan# 3517
Delta R.T. 0.004 min
Lab File: BN015823.D
Acq: 09 Aug 2021 14:59

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:276 Resp: 14120
Ion Ratio Lower Upper
276 100
277 25.3 19.0 28.6
138 17.8 12.2 18.4

