

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015825.D
 Acq On : 09 Aug 2021 16:12
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Aug 09 17:26:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8604	0.400	ng	0.00
7) Naphthalene-d8	10.724	136	42136	0.400	ng	0.00
13) Acenaphthene-d10	14.561	164	24417	0.400	ng	0.00
19) Phenanthrene-d10	17.310	188	49391	0.400	ng	0.00
29) Chrysene-d12	21.493	240	51770	0.400	ng	0.00
35) Perylene-d12	23.926	264	48615	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	8897	0.351	ng	0.00
5) Phenol-d6	7.080	99	11892	0.341	ng	0.00
8) Nitrobenzene-d5	9.076	82	11535	0.364	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	26477	0.409	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	3761	0.295	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	37030	0.406	ng	0.00
27) Fluoranthene-d10	19.341	212	56286	0.407	ng	0.00
31) Terphenyl-d14	19.937	244	56930	0.357	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.410	88	1650	0.476	ng	92
3) n-Nitrosodimethylamine	3.770	42	3287	0.403	ng	# 100
6) bis(2-Chloroethyl)ether	7.347	93	9666	0.389	ng	100
9) Naphthalene	10.774	128	43197	0.391	ng	100
10) Hexachlorobutadiene	11.066	225	11091	0.550	ng	# 100
12) 2-Methylnaphthalene	12.390	142	29535	0.394	ng	100
16) Acenaphthylene	14.281	152	38872	0.342	ng	100
17) Acenaphthene	14.624	154	26243	0.367	ng	100
18) Fluorene	15.601	166	33312	0.384	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	1292	0.335	ng	100
21) 4-Bromophenyl-phenylether	16.494	248	10122	0.383	ng	100
22) Hexachlorobenzene	16.616	284	13133	0.403	ng	100
23) Atrazine	16.762	200	8110	0.299	ng	100
24) Pentachlorophenol	16.957	266	4127	0.336	ng	100
25) Phenanthrene	17.346	178	52669	0.374	ng	100
26) Anthracene	17.431	178	46562	0.340	ng	100
28) Fluoranthene	19.371	202	63421	0.402	ng	100
30) Pyrene	19.733	202	62975	0.343	ng	100
32) Benzo(a)anthracene	21.482	228	64051	0.363	ng	100
33) Chrysene	21.535	228	65546	0.393	ng	100
34) Bis(2-ethylhexyl)phtha...	21.419	149	25607	0.196	ng	100
36) Indeno(1,2,3-cd)pyrene	26.445	276	60864	0.332	ng	100
37) Benzo(b)fluoranthene	23.187	252	63734	0.374	ng	100
38) Benzo(k)fluoranthene	23.235	252	61538	0.384	ng	100
39) Benzo(a)pyrene	23.820	252	55054	0.365	ng	100
40) Dibenzo(a,h)anthracene	26.462	278	49116	0.327	ng	100
41) Benzo(g,h,i)perylene	27.212	276	54926	0.346	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
Data File : BN015825.D
Acq On : 09 Aug 2021 16:12
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Aug 09 17:26:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 16:46:20 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.928 min Scan# 526
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 8604
Ion Ratio Lower Upper
152 100
150 153.7 123.0 184.4
115 65.2 52.2 78.2



#2
1,4-Dioxane
Concen: 0.476 ng
RT: 3.410 min Scan# 36
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion: 88 Resp: 1650
Ion Ratio Lower Upper
88 100
43 33.3 27.4 41.2
58 70.5 64.3 96.5









#3
n-Nitrosodimethylamine
Concen: 0.403 ng
RT: 3.770 min Scan# 75
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 3287
Ion Ratio Lower Upper
42 100
74 104.0 83.4 125.2
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.351 ng
RT: 5.485 min Scan# 261
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion: 112 Resp: 8897
Ion Ratio Lower Upper
112 100
64 66.7 53.4 80.0
63 35.8 28.6 43.0









#5
Phenol-d6
Concen: 0.341 ng
RT: 7.080 min Scan# 434
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 11892
Ion Ratio Lower Upper
99 100
42 23.7 19.0 28.4
71 39.0 31.2 46.8



#6
bis(2-Chloroethyl)ether
Concen: 0.389 ng
RT: 7.347 min Scan# 463
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion: 93 Resp: 9666
Ion Ratio Lower Upper
93 100
63 80.4 64.3 96.5
95 32.0 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.724 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	42136
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	8.9	7.1	10.7
68	6.1	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.364 ng
RT: 9.076 min Scan# 628
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	82	Resp:	11535
Ion	Ratio	Lower	Upper
82	100		
128	32.7	26.2	39.2
54	56.7	45.4	68.0









#9
Naphthalene
Concen: 0.391 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	128	Resp:	43197
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.5	10.8	16.2



#10
Hexachlorobutadiene
Concen: 0.550 ng
RT: 11.066 min Scan# 785
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	225	Resp:	11091
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.8	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 0.409 ng
RT: 12.314 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 26477
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.394 ng
RT: 12.390 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:142 Resp: 29535
Ion Ratio Lower Upper
142 100
141 88.3 70.6 106.0
115 36.4 29.1 43.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.561 min Scan# 1206
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 24417
Ion Ratio Lower Upper
164 100
162 101.5 81.2 121.8
160 43.7 35.0 52.4



#14
2,4,6-Tribromophenol
Concen: 0.295 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:330 Resp: 3761
Ion Ratio Lower Upper
330 100
332 106.9 86.2 129.2
141 31.4 25.0 37.6









#15
2-Fluorobiphenyl
Concen: 0.406 ng
RT: 13.190 min Scan# 1098
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	37030
Ion	Ratio	Lower	Upper
172	100		
171	33.6	26.9	40.3
170	21.6	17.3	25.9



#16
Acenaphthylene
Concen: 0.342 ng
RT: 14.281 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	152	Resp:	38872
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.6	23.4
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.367 ng
RT: 14.624 min Scan# 1211
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 26243
Ion Ratio Lower Upper
154 100
153 112.6 90.1 135.1
152 53.4 42.7 64.1



#18
Fluorene
Concen: 0.384 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:166 Resp: 33312
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.310 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	188	Resp:	49391
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	5.9	4.7	7.1



#20
4,6-Dinitro-2-methylphenol
Concen: 0.335 ng
RT: 15.677 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	198	Resp:	1292
Ion Ratio	Lower	Upper	
198	100		
182	13.5	10.8	16.2
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.383 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	10122
Ion Ratio	Lower	Upper	
248	100		
250	93.5	74.8	112.2
141	75.6	60.5	90.7



#22
Hexachlorobenzene
Concen: 0.403 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	284	Resp:	13133
Ion Ratio	Lower	Upper	
284	100		
142	28.7	23.0	34.4
249	31.6	25.3	37.9









#23
Atrazine
Concen: 0.299 ng
RT: 16.762 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 8110
Ion Ratio Lower Upper
200 100
173 24.3 19.4 29.2
215 49.3 39.4 59.2



#24
Pentachlorophenol
Concen: 0.336 ng
RT: 16.957 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:266 Resp: 4127
Ion Ratio Lower Upper
266 100
264 63.8 51.0 76.6
268 64.7 51.8 77.6









#25
Phenanthrene
Concen: 0.374 ng
RT: 17.346 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 52669
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.9 12.7 19.1



#26
Anthracene
Concen: 0.340 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:178 Resp: 46562
Ion Ratio Lower Upper
178 100
176 17.9 14.3 21.5
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.407 ng
RT: 19.341 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 56286
Ion Ratio Lower Upper
212 100
106 9.8 7.8 11.8
104 5.7 4.6 6.8



#28
Fluoranthene
Concen: 0.402 ng
RT: 19.371 min Scan# 1758
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:202 Resp: 63421
Ion Ratio Lower Upper
202 100
101 8.8 7.0 10.6
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 51770
Ion Ratio Lower Upper
240 100
120 9.8 7.8 11.8
236 26.4 21.1 31.7



#30
Pyrene
Concen: 0.343 ng
RT: 19.733 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:202 Resp: 62975
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.2
203 17.6 14.1 21.1









#31
Terphenyl-d14
Concen: 0.357 ng
RT: 19.937 min Scan# 1872
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:244 Resp: 56930
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 8.7 7.0 10.4



#32
Benzo(a)anthracene
Concen: 0.363 ng
RT: 21.482 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:228 Resp: 64051
Ion Ratio Lower Upper
228 100
226 25.8 20.6 31.0
229 19.7 15.8 23.6









#33
Chrysene
Concen: 0.393 ng
RT: 21.535 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	228	Resp:	65546
Ion	Ratio	Lower	Upper
228	100		
226	28.4	22.7	34.1
229	19.8	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.196 ng
RT: 21.419 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:	149	Resp:	25607
Ion	Ratio	Lower	Upper
149	100		
167	28.6	22.7	34.1
279	5.3	4.2	6.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.926 min Scan# 2556
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:264	Resp:	48615
Ion	Ratio	Lower Upper
264	100	
260	24.5	19.6 29.4
265	27.1	21.7 32.5



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.332 ng
RT: 26.445 min Scan# 3292
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:276	Resp:	60864
Ion	Ratio	Lower Upper
276	100	
138	18.2	14.5 21.7
277	25.3	20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 0.374 ng
RT: 23.187 min Scan# 2340
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 63734
Ion Ratio Lower Upper
252 100
253 22.7 18.2 27.2
125 7.9 6.3 9.5



#38
Benzo(k)fluoranthene
Concen: 0.384 ng
RT: 23.235 min Scan# 2354
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:252 Resp: 61538
Ion Ratio Lower Upper
252 100
253 22.6 18.1 27.1
125 7.8 6.2 9.4









#39
Benzo(a)pyrene
Concen: 0.365 ng
RT: 23.820 min Scan# 2525
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252	Resp:	55054
Ion Ratio	Lower	Upper
252	100	
253	23.1	18.5 27.7
125	8.8	7.0 10.6



#40
Dibenzo(a,h)anthracene
Concen: 0.327 ng
RT: 26.462 min Scan# 3297
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Tgt Ion:278	Resp:	49116
Ion Ratio	Lower	Upper
278	100	
139	13.1	10.5 15.7
279	24.4	19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.346 ng
RT: 27.212 min Scan# 3516
Delta R.T. 0.000 min
Lab File: BN015825.D
Acq: 09 Aug 2021 16:12

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:276 Resp: 54926
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 15.3 12.2 18.4

