

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015826.D
 Acq On : 09 Aug 2021 16:49
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Aug 09 17:27:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	8919	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	42061	0.400	ng	0.00
13) Acenaphthene-d10	14.560	164	24278	0.400	ng	0.00
19) Phenanthrene-d10	17.309	188	48129	0.400	ng	0.00
29) Chrysene-d12	21.493	240	51544	0.400	ng	0.00
35) Perylene-d12	23.926	264	49495	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.484	112	18356	0.699	ng	0.00
5) Phenol-d6	7.080	99	24789	0.685	ng	0.00
8) Nitrobenzene-d5	9.076	82	24694	0.780	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	54504	0.844	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	8317	0.656	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	75256	0.831	ng	0.00
27) Fluoranthene-d10	19.336	212	117779	0.873	ng	0.00
31) Terphenyl-d14	19.936	244	118901	0.748	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.410	88	2958	0.823	ng	97
3) n-Nitrosodimethylamine	3.769	42	7501	0.887	ng	# 99
6) bis(2-Chloroethyl)ether	7.347	93	19802	0.770	ng	99
9) Naphthalene	10.774	128	87588	0.794	ng	100
10) Hexachlorobutadiene	11.065	225	23004	1.144	ng	# 99
12) 2-Methylnaphthalene	12.389	142	63986	0.855	ng	100
16) Acenaphthylene	14.281	152	82265	0.727	ng	100
17) Acenaphthene	14.624	154	54285	0.763	ng	99
18) Fluorene	15.601	166	68138	0.790	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	3423	0.910	ng	# 88
21) 4-Bromophenyl-phenylether	16.494	248	20752	0.806	ng	99
22) Hexachlorobenzene	16.616	284	26750	0.843	ng	100
23) Atrazine	16.762	200	17591	0.666	ng	99
24) Pentachlorophenol	16.956	266	9678	0.808	ng	99
25) Phenanthrene	17.346	178	107543	0.783	ng	99
26) Anthracene	17.431	178	97389	0.731	ng	100
28) Fluoranthene	19.370	202	131383	0.855	ng	100
30) Pyrene	19.733	202	129353	0.709	ng	100
32) Benzo(a)anthracene	21.482	228	132351	0.752	ng	100
33) Chrysene	21.535	228	135133	0.813	ng	100
34) Bis(2-ethylhexyl)phtha...	21.408	149	53782	0.414	ng	99
36) Indeno(1,2,3-cd)pyrene	26.441	276	138077	0.739	ng	100
37) Benzo(b)fluoranthene	23.183	252	139273	0.803	ng	99
38) Benzo(k)fluoranthene	23.234	252	128086	0.784	ng	99
39) Benzo(a)pyrene	23.816	252	118850	0.775	ng	98
40) Dibenzo(a,h)anthracene	26.465	278	113242	0.741	ng	99
41) Benzo(g,h,i)perylene	27.215	276	120616	0.747	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
Data File : BN015826.D
Acq On : 09 Aug 2021 16:49
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Aug 09 17:27:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 16:46:20 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.928 min Scan# 526
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampled :
SSTDICC0.8


Tgt Ion:152 Resp: 8919
Ion Ratio Lower Upper
152 100
150 153.7 123.0 184.4
115 64.8 52.2 78.2



#2
1,4-Dioxane
Concen: 0.823 ng
RT: 3.410 min Scan# 36
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion: 88 Resp: 2958
Ion Ratio Lower Upper
88 100
43 37.2 27.4 41.2
58 82.9 64.3 96.5









#3
n-Nitrosodimethylamine
Concen: 0.887 ng
RT: 3.769 min Scan# 75
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion: 42 Resp: 7501
Ion Ratio Lower Upper
42 100
74 102.9 83.4 125.2
44 7.5 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.699 ng
RT: 5.484 min Scan# 261
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion: 112 Resp: 18356
Ion Ratio Lower Upper
112 100
64 67.2 53.4 80.0
63 36.4 28.6 43.0









#5
Phenol-d6
Concen: 0.685 ng
RT: 7.080 min Scan# 434
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampled :
SSTDIC0.8


Tgt Ion: 99 Resp: 24789
Ion Ratio Lower Upper
99 100
42 23.4 19.0 28.4
71 39.5 31.2 46.8



#6
bis(2-Chloroethyl)ether
Concen: 0.770 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion: 93 Resp: 19802
Ion Ratio Lower Upper
93 100
63 80.1 64.3 96.5
95 32.8 25.6 38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.723 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	42061
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	9.0	7.1	10.7
68	6.0	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.780 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	82	Resp:	24694
Ion	Ratio	Lower	Upper
82	100		
128	33.2	26.2	39.2
54	56.4	45.4	68.0









#9
Naphthalene
Concen: 0.794 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	128	Resp:	87588
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.9	13.3
127	13.4	10.8	16.2



#10
Hexachlorobutadiene
Concen: 1.144 ng
RT: 11.065 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	225	Resp:	23004
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 0.844 ng
RT: 12.314 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 54504
Ion Ratio Lower Upper
152 100
151 20.4 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.855 ng
RT: 12.389 min Scan# 1018
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:142 Resp: 63986
Ion Ratio Lower Upper
142 100
141 88.1 70.6 106.0
115 36.6 29.1 43.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164	Resp:	24278
Ion	Ratio	Lower Upper
164	100	
162	101.1	81.2 121.8
160	43.9	35.0 52.4



#14
2,4,6-Tribromophenol
Concen: 0.656 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:330	Resp:	8317
Ion	Ratio	Lower Upper
330	100	
332	99.8	86.2 129.2
141	30.7	25.0 37.6









#15
2-Fluorobiphenyl
Concen: 0.831 ng
RT: 13.190 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	172	Resp:	75256
Ion Ratio	Lower	Upper	
172	100		
171	34.1	26.9	40.3
170	21.6	17.3	25.9



#16
Acenaphthylene
Concen: 0.727 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:	152	Resp:	82265
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.6	23.4
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.763 ng
RT: 14.624 min Scan# 1211
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 54285
Ion Ratio Lower Upper
154 100
153 111.7 90.1 135.1
152 53.0 42.7 64.1



#18
Fluorene
Concen: 0.790 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:166 Resp: 68138
Ion Ratio Lower Upper
166 100
165 98.0 78.8 118.2
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.309 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:188 Resp: 48129
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 5.9 4.7 7.1



#20
4,6-Dinitro-2-methylphenol
Concen: 0.910 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:198 Resp: 3423
Ion Ratio Lower Upper
198 100
182 8.7 10.8 16.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.806 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:248 Resp: 20752
Ion Ratio Lower Upper
248 100
250 93.5 74.8 112.2
141 74.5 60.5 90.7



#22
Hexachlorobenzene
Concen: 0.843 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:284 Resp: 26750
Ion Ratio Lower Upper
284 100
142 28.5 23.0 34.4
249 31.7 25.3 37.9









#23
Atrazine
Concen: 0.666 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 17591
Ion Ratio Lower Upper
200 100
173 24.5 19.4 29.2
215 50.4 39.4 59.2



#24
Pentachlorophenol
Concen: 0.808 ng
RT: 16.956 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:266 Resp: 9678
Ion Ratio Lower Upper
266 100
264 62.5 51.0 76.6
268 63.8 51.8 77.6









#25
Phenanthrene
Concen: 0.783 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 107543
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.4 12.7 19.1



#26
Anthracene
Concen: 0.731 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:178 Resp: 97389
Ion Ratio Lower Upper
178 100
176 17.9 14.3 21.5
179 15.3 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.873 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.005 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:212 Resp: 117779
Ion Ratio Lower Upper
212 100
106 9.6 7.8 11.8
104 5.6 4.6 6.8



#28
Fluoranthene
Concen: 0.855 ng
RT: 19.370 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:202 Resp: 131383
Ion Ratio Lower Upper
202 100
101 8.7 7.0 10.6
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 51544
Ion Ratio Lower Upper
240 100
120 8.3 7.8 11.8
236 26.0 21.1 31.7



#30
Pyrene
Concen: 0.709 ng
RT: 19.733 min Scan# 1831
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:202 Resp: 129353
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.2
203 17.5 14.1 21.1









#31
Terphenyl-d14
Concen: 0.748 ng
RT: 19.936 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:244 Resp: 118901




Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.1	9.1
122	8.5	7.0	10.4




#32
Benzo(a)anthracene
Concen: 0.752 ng
RT: 21.482 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:228 Resp: 132351

Ion	Ratio	Lower	Upper
228	100		
226	25.7	20.6	31.0
229	19.9	15.8	23.6









#33
Chrysene
Concen: 0.813 ng
RT: 21.535 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:228 Resp: 135133
Ion Ratio Lower Upper
228 100
226 28.1 22.7 34.1
229 19.8 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.414 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.011 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Tgt Ion:149 Resp: 53782
Ion Ratio Lower Upper
149 100
167 28.9 22.7 34.1
279 5.2 4.2 6.4







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.926 min Scan# 2556
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:264 Resp: 49495
Ion Ratio Lower Upper
264 100
260 24.1 19.6 29.4
265 26.1 21.7 32.5




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.739 ng
RT: 26.441 min Scan# 3291
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49



Tgt Ion:276 Resp: 138077
Ion Ratio Lower Upper
276 100
138 18.1 14.5 21.7
277 25.2 20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 0.803 ng
RT: 23.183 min Scan# 2339
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 139273
Ion Ratio Lower Upper
252 100
253 22.2 18.2 27.2
125 7.7 6.3 9.5



#38
Benzo(k)fluoranthene
Concen: 0.784 ng
RT: 23.234 min Scan# 2354
Delta R.T. -0.000 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:252 Resp: 128086
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 7.3 6.2 9.4









#39
Benzo(a)pyrene
Concen: 0.775 ng
RT: 23.816 min Scan# 2524
Delta R.T. -0.004 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 118850
Ion Ratio Lower Upper
252 100
253 22.3 18.5 27.7
125 8.4 7.0 10.6



#40
Dibenzo(a,h)anthracene
Concen: 0.741 ng
RT: 26.465 min Scan# 3298
Delta R.T. 0.003 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Tgt Ion:278 Resp: 113242
Ion Ratio Lower Upper
278 100
139 12.7 10.5 15.7
279 24.2 19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.747 ng
RT: 27.215 min Scan# 3517
Delta R.T. 0.003 min
Lab File: BN015826.D
Acq: 09 Aug 2021 16:49

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:	276	Resp:	120616
Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.0	28.6
138	15.3	12.2	18.4

