

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
 Data File : BN015827.D
 Acq On : 09 Aug 2021 17:26
 Operator : CG/JU
 Sample : SSTDIC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC1.6

Quant Time: Aug 09 18:03:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 09 16:46:20 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	9395	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	43287	0.400	ng	0.00
13) Acenaphthene-d10	14.560	164	25233	0.400	ng	0.00
19) Phenanthrene-d10	17.309	188	50998	0.400	ng	0.00
29) Chrysene-d12	21.493	240	56842	0.400	ng	0.00
35) Perylene-d12	23.926	264	54674	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	38803	1.403	ng	0.00
5) Phenol-d6	7.080	99	52514	1.379	ng	0.00
8) Nitrobenzene-d5	9.076	82	55903	1.715	ng	0.00
11) 2-Methylnaphthalene-d10	12.314	152	114588	1.724	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	19206	1.458	ng	0.00
15) 2-Fluorobiphenyl	13.190	172	156895	1.666	ng	0.00
27) Fluoranthene-d10	19.336	212	263026	1.840	ng	0.00
31) Terphenyl-d14	19.936	244	262690	1.499	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.410	88	6200	1.638	ng	96
3) n-Nitrosodimethylamine	3.760	42	16858	1.892	ng	# 99
6) bis(2-Chloroethyl)ether	7.347	93	41331	1.525	ng	99
9) Naphthalene	10.774	128	182904	1.610	ng	100
10) Hexachlorobutadiene	11.066	225	48194	2.328	ng	# 99
12) 2-Methylnaphthalene	12.389	142	132129	1.716	ng	99
16) Acenaphthylene	14.281	152	180513	1.536	ng	100
17) Acenaphthene	14.624	154	116473	1.576	ng	98
18) Fluorene	15.601	166	146892	1.639	ng	99
20) 4,6-Dinitro-2-methylph...	15.677	198	9396	2.358	ng	# 83
21) 4-Bromophenyl-phenylether	16.494	248	45168	1.655	ng	99
22) Hexachlorobenzene	16.616	284	57952	1.724	ng	99
23) Atrazine	16.762	200	41865	1.497	ng	99
24) Pentachlorophenol	16.957	266	23842	1.879	ng	99
25) Phenanthrene	17.346	178	231589	1.592	ng	99
26) Anthracene	17.431	178	217170	1.538	ng	100
28) Fluoranthene	19.366	202	286427	1.759	ng	100
30) Pyrene	19.728	202	286031	1.421	ng	100
32) Benzo(a)anthracene	21.482	228	304384	1.569	ng	99
33) Chrysene	21.535	228	297362	1.623	ng	100
34) Bis(2-ethylhexyl)phtha...	21.408	149	143991	1.006	ng	99
36) Indeno(1,2,3-cd)pyrene	26.438	276	317913	1.541	ng	100
37) Benzo(b)fluoranthene	23.183	252	307589	1.606	ng	99
38) Benzo(k)fluoranthene	23.231	252	294243	1.631	ng	99
39) Benzo(a)pyrene	23.816	252	271901	1.604	ng	98
40) Dibenzo(a,h)anthracene	26.462	278	261401	1.548	ng	98
41) Benzo(g,h,i)perylene	27.215	276	273454	1.533	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080921\
Data File : BN015827.D
Acq On : 09 Aug 2021 17:26
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Aug 09 18:03:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN080921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 09 16:46:20 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.928 min Scan# 526
 Delta R.T. 0.000 min
 Lab File: BN015827.D
 Acq: 09 Aug 2021 17:26

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion: 152 Resp: 9395




Ion	Ratio	Lower	Upper
152	100		
150	154.5	123.0	184.4
115	63.9	52.2	78.2




#2
 1,4-Dioxane
 Concen: 1.638 ng
 RT: 3.410 min Scan# 36
 Delta R.T. -0.000 min
 Lab File: BN015827.D
 Acq: 09 Aug 2021 17:26

Tgt Ion: 88 Resp: 6200

Ion	Ratio	Lower	Upper
88	100		
43	37.6	27.4	41.2
58	83.3	64.3	96.5









#3
n-Nitrosodimethylamine
Concen: 1.892 ng
RT: 3.760 min Scan# 74
Delta R.T. -0.009 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 16858
Ion Ratio Lower Upper
42 100
74 105.4 83.4 125.2
44 6.4 0.0 0.0#



#4
2-Fluorophenol
Concen: 1.403 ng
RT: 5.494 min Scan# 262
Delta R.T. 0.009 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion: 112 Resp: 38803
Ion Ratio Lower Upper
112 100
64 67.0 53.4 80.0
63 35.7 28.6 43.0









#5
Phenol-d6
Concen: 1.379 ng
RT: 7.080 min Scan# 434
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Instrument :
BNA_N
ClientSampled :
SSTDICC1.6

Tgt Ion: 99 Resp: 52514




Ion	Ratio	Lower	Upper
99	100		
42	23.1	19.0	28.4
71	39.3	31.2	46.8




#6
bis(2-Chloroethyl)ether
Concen: 1.525 ng
RT: 7.347 min Scan# 463
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion: 93 Resp: 41331

Ion	Ratio	Lower	Upper
93	100		
63	79.9	64.3	96.5
95	32.9	25.6	38.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.723 min Scan# 758
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	43287
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.8	13.2
54	8.9	7.1	10.7
68	6.0	4.9	7.3



#8
Nitrobenzene-d5
Concen: 1.715 ng
RT: 9.076 min Scan# 628
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:	82	Resp:	55903
Ion	Ratio	Lower	Upper
82	100		
128	33.6	26.2	39.2
54	55.3	45.4	68.0









#9
Naphthalene
Concen: 1.610 ng
RT: 10.774 min Scan# 762
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 182904
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.3 10.8 16.2



#10
Hexachlorobutadiene
Concen: 2.328 ng
RT: 11.066 min Scan# 785
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:225 Resp: 48194
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.8 77.8









#11
2-Methylnaphthalene-d10
Concen: 1.724 ng
RT: 12.314 min Scan# 1001
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 114588
Ion Ratio Lower Upper
152 100
151 20.4 16.4 24.6



#12
2-Methylnaphthalene
Concen: 1.716 ng
RT: 12.389 min Scan# 1018
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:142 Resp: 132129
Ion Ratio Lower Upper
142 100
141 88.9 70.6 106.0
115 36.6 29.1 43.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.560 min Scan# 1206
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 25233
Ion Ratio Lower Upper
164 100
162 99.8 81.2 121.8
160 43.5 35.0 52.4



#14
2,4,6-Tribromophenol
Concen: 1.458 ng
RT: 16.044 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:330 Resp: 19206
Ion Ratio Lower Upper
330 100
332 98.0 86.2 129.2
141 30.8 25.0 37.6









#15
2-Fluorobiphenyl
Concen: 1.666 ng
RT: 13.190 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 156895
Ion Ratio Lower Upper
172 100
171 34.2 26.9 40.3
170 21.7 17.3 25.9



#16
Acenaphthylene
Concen: 1.536 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:152 Resp: 180513
Ion Ratio Lower Upper
152 100
151 19.5 15.6 23.4
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 1.576 ng
RT: 14.624 min Scan# 1211
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 116473
Ion Ratio Lower Upper
154 100
153 110.9 90.1 135.1
152 52.3 42.7 64.1



#18
Fluorene
Concen: 1.639 ng
RT: 15.601 min Scan# 1288
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:166 Resp: 146892
Ion Ratio Lower Upper
166 100
165 97.6 78.8 118.2
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.309 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 50998
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.2 4.7 7.1



#20
4,6-Dinitro-2-methylphenol
Concen: 2.358 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:198 Resp: 9396
Ion Ratio Lower Upper
198 100
182 6.7 10.8 16.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.655 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 45168
Ion Ratio Lower Upper
248 100
250 92.6 74.8 112.2
141 75.6 60.5 90.7



#22
Hexachlorobenzene
Concen: 1.724 ng
RT: 16.616 min Scan# 1370
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:284 Resp: 57952
Ion Ratio Lower Upper
284 100
142 28.2 23.0 34.4
249 31.3 25.3 37.9









#23
Atrazine
Concen: 1.497 ng
RT: 16.762 min Scan# 1382
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 41865
Ion Ratio Lower Upper
200 100
173 24.1 19.4 29.2
215 49.8 39.4 59.2



#24
Pentachlorophenol
Concen: 1.879 ng
RT: 16.957 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:266 Resp: 23842
Ion Ratio Lower Upper
266 100
264 62.7 51.0 76.6
268 63.9 51.8 77.6









#25
Phenanthrene
Concen: 1.592 ng
RT: 17.346 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 231589
Ion Ratio Lower Upper
178 100
176 18.5 14.6 22.0
179 15.5 12.7 19.1



#26
Anthracene
Concen: 1.538 ng
RT: 17.431 min Scan# 1437
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:178 Resp: 217170
Ion Ratio Lower Upper
178 100
176 17.9 14.3 21.5
179 15.4 12.3 18.5









#27
Fluoranthene-d10
Concen: 1.840 ng
RT: 19.336 min Scan# 1751
Delta R.T. -0.005 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 263026
Ion Ratio Lower Upper
212 100
106 9.7 7.8 11.8
104 5.6 4.6 6.8



#28
Fluoranthene
Concen: 1.759 ng
RT: 19.366 min Scan# 1757
Delta R.T. -0.005 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:202 Resp: 286427
Ion Ratio Lower Upper
202 100
101 8.8 7.0 10.6
203 17.5 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 56842
Ion Ratio Lower Upper
240 100
120 8.0 7.8 11.8
236 26.0 21.1 31.7



#30
Pyrene
Concen: 1.421 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.005 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:202 Resp: 286031
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.2
203 17.6 14.1 21.1









#31
Terphenyl-d14
Concen: 1.499 ng
RT: 19.936 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 262690
Ion Ratio Lower Upper
244 100
212 7.4 6.1 9.1
122 8.1 7.0 10.4



#32
Benzo(a)anthracene
Concen: 1.569 ng
RT: 21.482 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:228 Resp: 304384
Ion Ratio Lower Upper
228 100
226 25.2 20.6 31.0
229 19.9 15.8 23.6









#33
Chrysene
Concen: 1.623 ng
RT: 21.535 min Scan# 2030
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 297362
Ion Ratio Lower Upper
228 100
226 28.3 22.7 34.1
229 19.8 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 1.006 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.011 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26


Tgt Ion:149 Resp: 143991
Ion Ratio Lower Upper
149 100
167 29.1 22.7 34.1
279 5.3 4.2 6.4







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.926 min Scan# 2556
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:264 Resp: 54674
Ion Ratio Lower Upper
264 100
260 24.0 19.6 29.4
265 25.9 21.7 32.5




#36
Indeno(1,2,3-cd)pyrene
Concen: 1.541 ng
RT: 26.438 min Scan# 3290
Delta R.T. -0.007 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26



Tgt Ion:276 Resp: 317913
Ion Ratio Lower Upper
276 100
138 17.8 14.5 21.7
277 25.3 20.2 30.4









#37
Benzo(b)fluoranthene
Concen: 1.606 ng
RT: 23.183 min Scan# 2339
Delta R.T. -0.004 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 307589
Ion Ratio Lower Upper
252 100
253 22.1 18.2 27.2
125 7.4 6.3 9.5



#38
Benzo(k)fluoranthene
Concen: 1.631 ng
RT: 23.231 min Scan# 2353
Delta R.T. -0.004 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26


Tgt Ion:252 Resp: 294243
Ion Ratio Lower Upper
252 100
253 21.9 18.1 27.1
125 7.2 6.2 9.4








#39
Benzo(a)pyrene
Concen: 1.604 ng
RT: 23.816 min Scan# 2524
Delta R.T. -0.004 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:252 Resp: 271901
Ion Ratio Lower Upper
252 100
253 22.0 18.5 27.7
125 8.2 7.0 10.6



#40
Dibenzo(a,h)anthracene
Concen: 1.548 ng
RT: 26.462 min Scan# 3297
Delta R.T. -0.000 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Tgt Ion:278 Resp: 261401
Ion Ratio Lower Upper
278 100
139 11.8 10.5 15.7
279 23.6 19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 1.533 ng
RT: 27.215 min Scan# 3517
Delta R.T. 0.003 min
Lab File: BN015827.D
Acq: 09 Aug 2021 17:26

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:276 Resp: 273454
Ion Ratio Lower Upper
276 100
277 23.8 19.0 28.6
138 15.1 12.2 18.4

