

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/9/2023	BNA_N	<u>BN026845.D</u>
Compound Name	Response	Retention Time
DDT	2348027	20.939
DDD	24219	20.545
DDE	590	19.986
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
24809	2372836	1.05