

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081019\
 Data File : BN007052.D
 Acq On : 10 Aug 2019 16:22
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Quant Time: Aug 10 21:57:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 10 21:44:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	n-Nitrosodimethylamine	0.400	0.388	3.0	129	0.00
3 S	2-Fluorophenol	0.400	0.377	5.8	103	0.00
4 S	Phenol-d6	0.400	0.361	9.8	101	0.00
5	bis(2-Chloroethyl)ether	0.400	0.387	3.3	106	0.00
6 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
7 S	Nitrobenzene-d5	0.400	0.364	9.0	105	0.00
8	Naphthalene	0.400	0.373	6.8	106	0.00
9	Hexachlorobutadiene	0.400	0.387	3.3	108	-0.01
10 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	104	0.00
11	2-Methylnaphthalene	0.400	0.364	9.0	104	0.00
12 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.400	0.278	30.5#	99	0.00
14 S	2-Fluorobiphenyl	0.400	0.419	-4.7	99	0.00
15	Acenaphthylene	0.400	0.368	8.0	102	0.00
16	Acenaphthene	0.400	0.375	6.3	104	0.00
17	Fluorene	0.400	0.330	17.5	103	0.00
18 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
19	4-Bromophenyl-phenylether	0.400	0.393	1.8	103	0.00
20	Hexachlorobenzene	0.400	0.405	-1.3	106	0.00
21	Pentachlorophenol	0.400	0.316	21.0	94	0.00
22	Phenanthrene	0.400	0.392	2.0	105	0.00
23	Anthracene	0.400	0.369	7.8	102	0.00
24 SURR	Fluoranthene-d10	0.400	0.378	5.5	102	0.00
25	Fluoranthene	0.400	0.382	4.5	102	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	105	-0.01
27	Pyrene	0.400	0.378	5.5	103	0.00
28 S	Terphenyl-d14	0.400	0.383	4.3	105	0.00
29	Benzo(a)anthracene	0.400	0.381	4.8	103	0.00
30	Chrysene	0.400	0.396	1.0	109	-0.01
31	Indeno(1,2,3-cd)pyrene	0.400	0.354	11.5	99	-0.01
32 I	Perylene-d12	0.400	0.400	0.0	99	0.00
33	Benzo(b)fluoranthene	0.400	0.397	0.8	100	0.00
34	Benzo(k)fluoranthene	0.400	0.408	-2.0	101	0.00
35 C	Benzo(a)pyrene	0.400	0.380	5.0	98	-0.01
36	Dibenzo(a,h)anthracene	0.400	0.385	3.8	99	-0.01
37	Benzo(g,h,i)perylene	0.400	0.391	2.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0