

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081019\
 Data File : BN007052.D
 Acq On : 10 Aug 2019 16:22
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Quant Time: Aug 10 21:57:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 10 21:44:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	n-Nitrosodimethylamine	0.263	0.255	3.0	129	0.00
3 S	2-Fluorophenol	0.961	0.905	5.8	103	0.00
4 S	Phenol-d6	1.191	1.074	9.8	101	0.00
5	bis(2-Chloroethyl)ether	1.052	1.018	3.2	106	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
7 S	Nitrobenzene-d5	0.323	0.293	9.3	105	0.00
8	Naphthalene	1.122	1.045	6.9	106	0.00
9	Hexachlorobutadiene	0.239	0.231	3.3	108	-0.01
10 SURR	2-Methylnaphthalene-d10	0.746	0.677	9.2	104	0.00
11	2-Methylnaphthalene	0.795	0.723	9.1	104	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.215	0.150	30.2#	99	0.00
14 S	2-Fluorobiphenyl	0.582	0.610	-4.8	99	0.00
15	Acenaphthylene	2.091	1.925	7.9	102	0.00
16	Acenaphthene	1.339	1.256	6.2	104	0.00
17	Fluorene	2.004	1.653	17.5	103	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
19	4-Bromophenyl-phenylether	0.247	0.243	1.6	103	0.00
20	Hexachlorobenzene	0.238	0.241	-1.3	106	0.00
21	Pentachlorophenol	0.089	0.070	21.3	94	0.00
22	Phenanthrene	1.199	1.175	2.0	105	0.00
23	Anthracene	1.095	1.012	7.6	102	0.00
24 SURR	Fluoranthene-d10	1.419	1.339	5.6	102	0.00
25	Fluoranthene	1.533	1.465	4.4	102	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
27	Pyrene	1.403	1.327	5.4	103	0.00
28 S	Terphenyl-d14	1.090	1.044	4.2	105	0.00
29	Benzo(a)anthracene	1.463	1.395	4.6	103	0.00
30	Chrysene	1.421	1.408	0.9	109	-0.01
31	Indeno(1,2,3-cd)pyrene	1.566	1.385	11.6	99	-0.01
32 I	Perylene-d12	1.000	1.000	0.0	99	0.00
33	Benzo(b)fluoranthene	1.372	1.361	0.8	100	0.00
34	Benzo(k)fluoranthene	1.355	1.383	-2.1	101	0.00
35 C	Benzo(a)pyrene	1.244	1.180	5.1	98	-0.01
36	Dibenzo(a,h)anthracene	1.241	1.193	3.9	99	-0.01
37	Benzo(g,h,i)perylene	1.276	1.247	2.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0