

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081019\
 Data File : BN007071.D
 Acq On : 11 Aug 2019 04:59
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 05:30:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 10 21:44:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	0.00
2	n-Nitrosodimethylamine	0.400	0.358	10.5	144	0.00
3 S	2-Fluorophenol	0.400	0.400	0.0	131	0.00
4 S	Phenol-d6	0.400	0.404	-1.0	137	0.00
5	bis(2-Chloroethyl)ether	0.400	0.383	4.3	126	0.00
6 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
7 S	Nitrobenzene-d5	0.400	0.377	5.8	137	0.00
8	Naphthalene	0.400	0.373	6.8	133	0.00
9	Hexachlorobutadiene	0.400	0.376	6.0	131	-0.01
10 SURR	2-Methylnaphthalene-d10	0.400	0.371	7.3	133	0.00
11	2-Methylnaphthalene	0.400	0.372	7.0	133	0.00
12 I	Acenaphthene-d10	0.400	0.400	0.0	132	0.00
13 S	2,4,6-Tribromophenol	0.400	0.295	26.3#	133	0.00
14 S	2-Fluorobiphenyl	0.400	0.330	17.5	99	0.00
15	Acenaphthylene	0.400	0.383	4.3	134	0.00
16	Acenaphthene	0.400	0.382	4.5	134	-0.01
17	Fluorene	0.400	0.326	18.5	129	0.00
18 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
19	4-Bromophenyl-phenylether	0.400	0.398	0.5	124	0.00
20	Hexachlorobenzene	0.400	0.401	-0.3	125	0.00
21	Pentachlorophenol	0.400	0.349	12.8	124	0.00
22	Phenanthrene	0.400	0.383	4.3	122	0.00
23	Anthracene	0.400	0.374	6.5	124	0.00
24 SURR	Fluoranthene-d10	0.400	0.352	12.0	114	0.00
25	Fluoranthene	0.400	0.359	10.3	114	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	117	-0.02
27	Pyrene	0.400	0.385	3.8	117	-0.01
28 S	Terphenyl-d14	0.400	0.384	4.0	117	-0.01
29	Benzo(a)anthracene	0.400	0.380	5.0	114	-0.01
30	Chrysene	0.400	0.397	0.8	121	-0.01
31	Indeno(1,2,3-cd)pyrene	0.400	0.403	-0.8	125	-0.02
32 I	Perylene-d12	0.400	0.400	0.0	124	-0.02
33	Benzo(b)fluoranthene	0.400	0.388	3.0	122	-0.02
34	Benzo(k)fluoranthene	0.400	0.384	4.0	119	-0.02
35 C	Benzo(a)pyrene	0.400	0.385	3.8	124	-0.02
36	Dibenzo(a,h)anthracene	0.400	0.394	1.5	127	-0.02
37	Benzo(g,h,i)perylene	0.400	0.388	3.0	122	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0