

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081019\
 Data File : BN007052.D
 Acq On : 10 Aug 2019 16:22
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Manual Integrations
 APPROVED

mohammad
 8/14/2019 2:56:02 AM

Quant Time: Aug 10 21:57:19 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sat Aug 10 21:44:50 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.43	152	6431	0.40	ng	0.00
6) Naphthalene-d8	10.19	136	23218	0.40	ng	0.00
12) Acenaphthene-d10	14.08	164	13090	0.40	ng	0.00
18) Phenanthrene-d10	16.82	188	33437	0.40	ng	0.00
26) Chrysene-d12	21.05	240	37231	0.40	ng	-0.01
32) Perylene-d12	23.16	264	35005	0.40	ng	0.00

System Monitoring Compounds

3) 2-Fluorophenol	5.06	112	5821	0.38	ng	0.00
4) Phenol-d6	6.61	99	6910	0.36	ng	0.00
7) Nitrobenzene-d5	8.56	82	6806	0.36	ng	0.00
10) 2-Methylnaphthalene-d10	11.79	152	15711	0.36	ng	0.00
13) 2,4,6-Tribromophenol	15.57	330	1958	0.28	ng	0.00
14) 2-Fluorobiphenyl	12.71	172	7981	0.42	ng	0.00
24) Fluoranthene-d10	18.87	212	44773	0.38	ng	0.00
28) Terphenyl-d14	19.49	244	38862	0.38	ng	0.00

Target Compounds

						Qvalue
2) n-Nitrosodimethylamine	3.36	42	1642m	0.388	ng	
5) bis(2-Chloroethyl)ether	6.87	93	6549	0.387	ng	99
8) Naphthalene	10.24	128	24273	0.373	ng	99
9) Hexachlorobutadiene	10.54	225	5370	0.387	ng	# 98
11) 2-Methylnaphthalene	11.87	142	16784	0.364	ng	98
15) Acenaphthylene	13.79	152	25197	0.368	ng	100
16) Acenaphthene	14.14	154	16437	0.375	ng	99
17) Fluorene	15.14	166	21636	0.330	ng	100
19) 4-Bromophenyl-phenylether	16.04	248	8123	0.393	ng	96
20) Hexachlorobenzene	16.14	284	8058	0.405	ng	100
21) Pentachlorophenol	16.48	266	2339	0.316	ng	98
22) Phenanthrene	16.87	178	39291	0.392	ng	100
23) Anthracene	16.96	178	33827	0.369	ng	100
25) Fluoranthene	18.90	202	48970	0.382	ng	100
27) Pyrene	19.26	202	49414	0.378	ng	100
29) Benzo(a)anthracene	21.03	228	51924	0.381	ng	99
30) Chrysene	21.08	228	52414	0.396	ng	98
31) Indeno(1,2,3-cd)pyrene	25.24	276	51559	0.354	ng	100
33) Benzo(b)fluoranthene	22.54	252	47639	0.397	ng	100
34) Benzo(k)fluoranthene	22.58	252	48421	0.408	ng	100
35) Benzo(a)pyrene	23.07	252	41323	0.380	ng	99
36) Dibenzo(a,h)anthracene	25.26	278	41776	0.385	ng	100
37) Benzo(g,h,i)perylene	25.87	276	43639	0.391	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081019\
Data File : BN007052.D
Acq On : 10 Aug 2019 16:22
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM

Quant Time: Aug 10 21:57:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 10 21:44:50 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.43 min Scan# 547
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion	Ratio	Lower	Upper
152	100		
150	153.5	121.7	182.5
115	57.7	46.2	69.4

Manual Integrations
APPROVED




mohammad
8/14/2019 2:56:02 AM




#2

n-Nitrosodimethylamine
Concen: 0.388 ng m
RT: 3.36 min Scan# 41
Delta R.T. 0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
42	100		
74	111.5	110.6	166.0
44	0.0	0.0	0.0





#3

2-Fluorophenol

Concen: 0.377 ng

RT: 5.06 min Scan# 252

Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N


ClientSampleId :

ICVBE081019




Tgt Ion:	112	Resp:	5821
Ion Ratio	Lower	Upper	
112	100		
64	53.2	42.6	63.8
63	28.2	22.2	33.2

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



Abundance Ion 112.00 (111.70 to 112.70): BNC
Ion 64.00 (63.70 to 64.70): BNC
Ion 63.00 (62.70 to 63.70): BNC



#4

Phenol-d6

Concen: 0.361 ng


RT: 6.61 min Scan# 445

Delta R.T. -0.00 min



Lab File: BN007052.D




Acq: 10 Aug 2019 16:22

Tgt Ion:	99	Resp:	6910
Ion Ratio	Lower	Upper	
99	100		
42	18.7	15.8	23.6
71	34.5	27.0	40.4



Abundance Ion 99.00 (98.70 to 99.70): BNC
Ion 42.00 (41.70 to 42.70): BNC
Ion 71.00 (70.70 to 71.70): BNC









#5
bis(2-Chloroethyl)ether
Concen: 0.387 ng
RT: 6.87 min Scan# 477
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion:	93	Resp:	6549
Ion	Ratio	Lower	Upper
93	100		
63	69.2	55.6	83.4
95	32.0	26.7	40.1

Instrument :
BNA_N
ClientSampleId :
ICVBE081019


Manual Integrations
APPROVED


mohammad
8/14/2019 2:56:02 AM



#6
Naphthalene-d8
Concen: 0.400 ng
RT: 10.19 min Scan# 813
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion:	136	Resp:	23218
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.4	14.0
54	7.6	6.6	9.8
68	6.8	6.0	9.0





#7

Nitrobenzene-d5

Concen: 0.364 ng

RT: 8.56 min Scan# 685

Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N





ClientSampleId :

ICVBE081019

Tgt Ion:	82	Resp:	6806
Ion Ratio	Lower	Upper	
82	100		
128	43.7	34.6	51.8
54	44.5	36.2	54.4

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#8

Naphthalene

Concen: 0.373 ng




RT: 10.24 min Scan# 817


Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Tgt Ion:	128	Resp:	24273
Ion Ratio	Lower	Upper	
128	100		
129	11.7	9.4	14.2
127	13.5	11.0	16.6





#9

Hexachlorobutadiene

Concen: 0.387 ng

RT: 10.54 min Scan# 841

Delta R.T. -0.01 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N


ClientSampleId :

ICVBE081019




Tgt Ion:	225	Resp:	5370
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.7	51.1	76.7

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E



#10

2-Methylnaphthalene-d10

Concen: 0.363 ng


RT: 11.79 min Scan# 982

Delta R.T. -0.00 min



Lab File: BN007052.D


Acq: 10 Aug 2019 16:22

Tgt Ion:	152	Resp:	15711
Ion Ratio	Lower	Upper	
152	100		
151	19.2	15.5	23.3



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E






#11
2-Methylnaphthalene
Concen: 0.364 ng
RT: 11.87 min Scan# 998
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion	Ratio	Lower	Upper
142	100		
141	87.7	69.0	103.6
115	35.2	27.1	40.7

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM






Abundance

Ion 142.00 (141.70 to 142.70): E


Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.08 min Scan# 1310
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.5	82.2	123.4
160	47.5	38.3	57.5






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.278 ng
 RT: 15.57 min Scan# 1444
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Tgt Ion	Ratio	Resp	Lower	Upper
330	100	1958		
332	0.0	0.0	0.0	0.0
141	39.8	33.6	50.4	




Manual Integrations
 APPROVED


mohammad
 8/14/2019 2:56:02 AM



#14
 2-Fluorobiphenyl
 Concen: 0.419 ng
 RT: 12.71 min Scan# 1187
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	7981		
171	33.4	26.3	39.5	
170	21.2	17.0	25.4	









#15
Acenaphthylene
Concen: 0.368 ng
RT: 13.79 min Scan# 1284
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion:	152	Resp:	25197
Ion Ratio	Lower	Upper	
152	100		
151	19.8	15.8	23.6
153	13.2	10.4	15.6




Manual Integrations
APPROVED


mohammad
8/14/2019 2:56:02 AM



#16
Acenaphthene
Concen: 0.375 ng
RT: 14.14 min Scan# 1316
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion:	154	Resp:	16437
Ion Ratio	Lower	Upper	
154	100		
153	110.6	89.5	134.3
152	55.7	44.8	67.2





#17

Fluorene

Concen: 0.330 ng

RT: 15.14 min Scan# 1405

Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N





ClientSampleId :

ICVBE081019

Tgt Ion:	166	Resp:	21636
Ion Ratio	Lower	Upper	
166	100		
165	97.4	77.9	116.9
167	13.5	10.9	16.3

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#18

Phenanthrene-d10

Concen: 0.400 ng




RT: 16.82 min Scan# 1560


Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Tgt Ion:	188	Resp:	33437
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.4	7.8	11.6









#19
4-Bromophenyl-phenylether
Concen: 0.393 ng
RT: 16.04 min Scan# 1486
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion	Ratio	Lower	Upper
248	100		
250	101.4	78.9	118.3
141	49.0	43.3	64.9




Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#20
Hexachlorobenzene
Concen: 0.405 ng
RT: 16.14 min Scan# 1496
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
284	100		
142	38.0	30.5	45.7
249	33.4	26.7	40.1









#21
 Pentachlorophenol
 Concen: 0.316 ng
 RT: 16.48 min Scan# 1528
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.1	52.3	78.5
268	63.2	51.5	77.3




Manual Integrations
 APPROVED


mohammad
 8/14/2019 2:56:02 AM



#22
 Phenanthrene
 Concen: 0.392 ng
 RT: 16.87 min Scan# 1564
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.2	22.8
179	15.4	12.3	18.5









#23
 Anthracene
 Concen: 0.369 ng
 RT: 16.96 min Scan# 1573
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Tgt Ion:	178	Resp:	33827
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.6	21.8
179	15.5	12.5	18.7




Manual Integrations
 APPROVED


mohammad
 8/14/2019 2:56:02 AM



#24
 Fluoranthene-d10
 Concen: 0.378 ng
 RT: 18.87 min Scan# 1845
 Delta R.T. -0.00 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Tgt Ion:	212	Resp:	44773
Ion Ratio	Lower	Upper	
212	100		
106	10.1	8.2	12.2
104	5.6	4.5	6.7





#25

Fluoranthene

Concen: 0.382 ng

RT: 18.90 min Scan# 1851

Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N





ClientSampleId :

ICVBE081019

Tgt Ion:	202	Resp:	48970
Ion	Ratio	Lower	Upper
202	100		
101	8.5	7.0	10.4
203	17.3	13.8	20.8

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#26

Chrysene-d12

Concen: 0.400 ng




RT: 21.05 min Scan# 2171


Delta R.T. -0.01 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Tgt Ion:	240	Resp:	37231
Ion	Ratio	Lower	Upper
240	100		
120	8.2	4.0	6.0#
236	27.4	21.3	31.9









#27
 Pyrene
 Concen: 0.378 ng
 RT: 19.26 min Scan# 1924
 Delta R.T. -0.01 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Instrument :
 BNA_N
 ClientSampleId :
 ICVBE081019

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.6	24.8
203	17.4	14.0	21.0




Manual Integrations
 APPROVED


mohammad
 8/14/2019 2:56:02 AM



#28
 Terphenyl-d14
 Concen: 0.383 ng
 RT: 19.49 min Scan# 1970
 Delta R.T. -0.01 min
 Lab File: BN007052.D
 Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.9	6.2	9.2
122	9.9	7.4	11.2





#29

Benzo(a)anthracene

Concen: 0.381 ng

RT: 21.03 min Scan# 2170

Delta R.T. -0.00 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N





ClientSampleId :

ICVBE081019

Tgt Ion:	228	Resp:	51924
Ion	Ratio	Lower	Upper
228	100		
226	26.5	21.8	32.6
229	20.0	15.6	23.4

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#30

Chrysene

Concen: 0.396 ng




RT: 21.08 min Scan# 2174


Delta R.T. -0.01 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Tgt Ion:	228	Resp:	52414
Ion	Ratio	Lower	Upper
228	100		
226	31.3	23.8	35.6
229	19.1	15.7	23.5





#31

Indeno(1,2,3-cd)pyrene

Concen: 0.354 ng

RT: 25.24 min Scan# 3128

Delta R.T. -0.01 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Instrument :

BNA_N





ClientSampleId :

ICVBE081019

Tgt Ion:	276	Resp:	51559
Ion Ratio	Lower	Upper	
276	100		
138	20.0	16.0	24.0
277	24.6	19.9	29.9

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM



#32

Perylene-d12

Concen: 0.400 ng




RT: 23.16 min Scan# 2521


Delta R.T. -0.01 min

Lab File: BN007052.D

Acq: 10 Aug 2019 16:22

Tgt Ion:	264	Resp:	35005
Ion Ratio	Lower	Upper	
264	100		
260	25.1	19.8	29.8
265	31.8	26.2	39.4









#33
Benzo(b)fluoranthene
Concen: 0.397 ng
RT: 22.54 min Scan# 2340
Delta R.T. -0.00 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.2	27.2
125	9.0	7.4	11.0




Manual Integrations
APPROVED


mohammad
8/14/2019 2:56:02 AM



#34
Benzo(k)fluoranthene
Concen: 0.408 ng
RT: 22.58 min Scan# 2351
Delta R.T. -0.01 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	18.1	27.1
125	8.9	7.4	11.0









#35
Benzo(a)pyrene
Concen: 0.380 ng
RT: 23.07 min Scan# 2494
Delta R.T. -0.01 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.5	18.5	27.7
125	10.1	8.2	12.4




Manual Integrations
APPROVED




mohammad
8/14/2019 2:56:02 AM



#36
Dibenzo(a,h)anthracene
Concen: 0.385 ng
RT: 25.26 min Scan# 3134
Delta R.T. -0.01 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion	Ratio	Lower	Upper
278	100		
139	13.9	11.3	16.9
279	24.3	19.3	28.9





#37
Benzo(g,h,i)perylene
Concen: 0.391 ng
RT: 25.87 min Scan# 3312
Delta R.T. -0.01 min
Lab File: BN007052.D
Acq: 10 Aug 2019 16:22

Tgt Ion:	276	Resp:	43639
Ion Ratio	Lower	Upper	
276	100		
277	24.5	19.8	29.8
138	17.4	14.6	22.0

Instrument :
BNA_N
ClientSampleId :
ICVBE081019

Manual Integrations
APPROVED

mohammad
8/14/2019 2:56:02 AM

