

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081123\
 Data File : BN026878.D
 Acq On : 10 Aug 2023 11:48
 Operator : MA/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Aug 11 01:50:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 01:48:37 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.109	152	6474	0.400	ng	0.00
7) Naphthalene-d8	10.938	136	18570	0.400	ng	0.00
13) Acenaphthene-d10	14.737	164	11679	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	25843	0.400	ng	0.00
29) Chrysene-d12	21.659	240	23678	0.400	ng	0.00
35) Perylene-d12	24.215	264	24877	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.618	112	80500	4.910	ng	0.00
5) Phenol-d6	7.236	99	106597	5.021	ng	0.00
8) Nitrobenzene-d5	9.294	82	77321	7.032	ng	0.00
11) 2-Methylnaphthalene-d10	12.510	152	147773	5.550	ng	0.00
14) 2,4,6-Tribromophenol	16.226	330	34998	8.802	ng	0.00
15) 2-Fluorobiphenyl	13.368	172	236854	4.967	ng	0.00
27) Fluoranthene-d10	19.504	212	365767	5.894	ng	0.00
31) Terphenyl-d14	20.090	244	267229	5.413	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.480	88	32122	4.251	ng	96
3) n-Nitrosodimethylamine	3.812	42	36449	4.321	ng	# 99
6) bis(2-Chloroethyl)ether	7.539	93	91509	4.781	ng	100
9) Naphthalene	10.992	128	261577	5.122	ng	98
10) Hexachlorobutadiene	11.226	225	46176	4.986	ng	# 100
12) 2-Methylnaphthalene	12.585	142	189025	5.424	ng	98
16) Acenaphthylene	14.459	152	312438	5.412	ng	100
17) Acenaphthene	14.801	154	187068	5.165	ng	98
18) Fluorene	15.780	166	250553	5.315	ng	99
20) 4,6-Dinitro-2-methylph...	16.934	198	2216	5.298	ng	# 1
21) 4-Bromophenyl-phenylether	16.673	248	82369	5.272	ng	96
22) Hexachlorobenzene	16.748	284	84464	4.731	ng	99
23) Atrazine	16.934	200	72187	6.997	ng	99
24) Pentachlorophenol	17.108	266	50573	11.030	ng	98
25) Phenanthrene	17.517	178	405633	5.168	ng	100
26) Anthracene	17.617	178	396696	5.657	ng	100
28) Fluoranthene	19.532	202	483928	5.537	ng	100
30) Pyrene	19.895	202	498480	4.872	ng	100
32) Benzo(a)anthracene	21.641	228	462849	5.692	ng	99
33) Chrysene	21.695	228	459111	5.108	ng	100
34) Bis(2-ethylhexyl)phtha...	21.533	149	308587	11.455	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.913	276	553778	5.455	ng	100
37) Benzo(b)fluoranthene	23.423	252	462788	4.728	ng	97
38) Benzo(k)fluoranthene	23.475	252	467781	4.651	ng	96
39) Benzo(a)pyrene	24.101	252	458318	5.135	ng	# 95
40) Dibenzo(a,h)anthracene	26.949	278	441122	5.595	ng	97
41) Benzo(g,h,i)perylene	27.750	276	469922	5.035	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081123\
Data File : BN026878.D
Acq On : 10 Aug 2023 11:48
Operator : MA/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Aug 11 01:50:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 11 01:48:37 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 8.109 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Tgt Ion:152 Resp: 6474
 Ion Ratio Lower Upper
 152 100
 150 154.5 123.8 185.6
 115 62.9 49.7 74.5



#2
 1,4-Dioxane
 Concen: 4.251 ng
 RT: 3.480 min Scan# 83
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48

Tgt Ion: 88 Resp: 32122
 Ion Ratio Lower Upper
 88 100
 43 24.1 19.2 28.8
 58 63.5 54.5 81.7








#3
n-Nitrosodimethylamine
Concen: 4.321 ng
RT: 3.812 min Scan# 131
Delta R.T. -0.014 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion: 42 Resp: 36449
Ion Ratio Lower Upper
42 100
74 149.7 119.4 179.2
44 6.5 8.7 13.1#



#4
2-Fluorophenol
Concen: 4.910 ng
RT: 5.618 min Scan# 379
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion: 112 Resp: 80500
Ion Ratio Lower Upper
112 100
64 51.2 41.5 62.3
63 27.1 21.7 32.5








#5
Phenol-d6
Concen: 5.021 ng
RT: 7.236 min Scan# 603
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion: 99 Resp: 106597
Ion Ratio Lower Upper
99 100
42 16.5 13.6 20.4
71 31.8 26.0 39.0



#6
bis(2-Chloroethyl)ether
Concen: 4.781 ng
RT: 7.539 min Scan# 645
Delta R.T. 0.007 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion: 93 Resp: 91509
Ion Ratio Lower Upper
93 100
63 68.9 55.0 82.4
95 32.6 26.2 39.2








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.938 min Scan# 998
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:	136	Resp:	18570
Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.0	13.6
54	7.2	5.5	8.3
68	6.9	5.4	8.2



#8
Nitrobenzene-d5
Concen: 7.032 ng
RT: 9.294 min Scan# 844
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:	82	Resp:	77321
Ion	Ratio	Lower	Upper
82	100		
128	40.4	34.1	51.1
54	44.9	37.0	55.4





#9

Naphthalene

Concen: 5.122 ng

RT: 10.992 min Scan# 1003

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:128 Resp: 261577

Ion Ratio Lower Upper

128 100

129 11.0 9.4 14.0

127 12.8 10.6 16.0



#10

Hexachlorobutadiene

Concen: 4.986 ng

RT: 11.226 min Scan# 1025

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:225 Resp: 46176

Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.7 50.8 76.2









#11
2-Methylnaphthalene-d10
Concen: 5.550 ng
RT: 12.510 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 147773
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5



#12
2-Methylnaphthalene
Concen: 5.424 ng
RT: 12.585 min Scan# 1337
Delta R.T. 0.004 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48


Tgt Ion:142 Resp: 189025
Ion Ratio Lower Upper
142 100
141 87.8 71.0 106.6
115 34.5 29.4 44.2







#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.737 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0





Tgt Ion:164 Resp: 11679
Ion Ratio Lower Upper
164 100
162 102.0 79.7 119.5
160 48.3 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 8.802 ng
RT: 16.226 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48



Tgt Ion:330 Resp: 34998
Ion Ratio Lower Upper
330 100
332 96.7 76.6 114.8
141 41.2 32.3 48.5








#15
2-Fluorobiphenyl
Concen: 4.967 ng
RT: 13.368 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 236854
Ion Ratio Lower Upper
172 100
171 34.2 27.4 41.0
170 23.0 18.0 27.0



#16
Acenaphthylene
Concen: 5.412 ng
RT: 14.459 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:152 Resp: 312438
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 5.165 ng
RT: 14.801 min Scan# 1616
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 187068
Ion Ratio Lower Upper
154 100
153 112.6 92.0 138.0
152 55.8 46.2 69.2



#18
Fluorene
Concen: 5.315 ng
RT: 15.780 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:166 Resp: 250553
Ion Ratio Lower Upper
166 100
165 97.7 77.4 116.2
167 13.4 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.480 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 25843
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.4 7.4 11.2



#20
4,6-Dinitro-2-methylphenol
Concen: 5.298 ng
RT: 16.934 min Scan# 1797
Delta R.T. 0.013 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:198 Resp: 2216
Ion Ratio Lower Upper
198 100
51 21.1 143.8 215.8#
105 138.5 273.5 410.3#









#21
4-Bromophenyl-phenylether
Concen: 5.272 ng
RT: 16.673 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 82369
Ion Ratio Lower Upper
248 100
250 105.7 83.0 124.4
141 31.9 29.9 44.9



#22
Hexachlorobenzene
Concen: 4.731 ng
RT: 16.748 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:284 Resp: 84464
Ion Ratio Lower Upper
284 100
142 38.4 30.0 45.0
249 30.6 24.4 36.6









#23
 Atrazine
 Concen: 6.997 ng
 RT: 16.934 min Scan# 1797
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:200 Resp: 72187
 Ion Ratio Lower Upper
 200 100
 173 20.5 17.1 25.7
 215 48.9 39.6 59.4



#24
 Pentachlorophenol
 Concen: 11.030 ng
 RT: 17.108 min Scan# 1811
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48

Tgt Ion:266 Resp: 50573
 Ion Ratio Lower Upper
 266 100
 264 62.7 49.7 74.5
 268 64.4 49.3 73.9





#25

Phenanthrene

Concen: 5.168 ng

RT: 17.517 min Scan# 1844

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:178 Resp: 405633

Ion Ratio Lower Upper

178 100

176 19.2 15.3 22.9

179 15.2 12.2 18.4



#26

Anthracene

Concen: 5.657 ng

RT: 17.617 min Scan# 1852

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:178 Resp: 396696

Ion Ratio Lower Upper

178 100

176 18.5 14.8 22.2

179 15.3 12.1 18.1





#27

Fluoranthene-d10

Concen: 5.894 ng

RT: 19.504 min Scan# 2137

Delta R.T. 0.005 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:212 Resp: 365767

Ion Ratio Lower Upper

212 100

106 13.7 10.9 16.3

104 7.5 5.9 8.9



#28

Fluoranthene

Concen: 5.537 ng

RT: 19.532 min Scan# 2144

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:202 Resp: 483928


Ion Ratio Lower Upper

202 100

101 10.9 8.9 13.3

203 17.3 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.659 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:240 Resp: 23678

Ion Ratio Lower Upper

240 100

120 13.9 11.1 16.7

236 27.9 22.2 33.2



#30

Pyrene

Concen: 4.872 ng

RT: 19.895 min Scan# 2222

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:202 Resp: 498480


Ion Ratio Lower Upper

202 100

200 20.9 16.6 25.0

203 17.9 14.4 21.6









#31
 Terphenyl-d14
 Concen: 5.413 ng
 RT: 20.090 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0


Tgt Ion:244 Resp: 267229
 Ion Ratio Lower Upper
 244 100
 212 8.3 6.9 10.3
 122 14.5 11.9 17.9



#32
 Benzo(a)anthracene
 Concen: 5.692 ng
 RT: 21.641 min Scan# 2532
 Delta R.T. 0.000 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48

Tgt Ion:228 Resp: 462849
 Ion Ratio Lower Upper
 228 100
 226 27.3 22.1 33.1
 229 19.6 15.5 23.3





#33

Chrysene

Concen: 5.108 ng

RT: 21.695 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 459111

Ion Ratio Lower Upper

228 100

226 30.6 24.6 37.0

229 19.2 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 11.455 ng

RT: 21.533 min Scan# 2520

Delta R.T. 0.000 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:149 Resp: 308587

Ion Ratio Lower Upper

149 100

167 27.7 21.4 32.0

279 3.3 3.6 5.4#





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.215 min Scan# 31

Delta R.T. 0.000 min

Lab File: BN026878.D


Acq: 10 Aug 2023 11:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0






Tgt Ion:264 Resp: 24877

Ion Ratio Lower Upper

264 100

260 25.7 20.4 30.6

265 40.3 27.0 40.6



#36

Indeno(1,2,3-cd)pyrene

Concen: 5.455 ng

RT: 26.913 min Scan# 4158

Delta R.T. 0.003 min

Lab File: BN026878.D

Acq: 10 Aug 2023 11:48




Tgt Ion:276 Resp: 553778

Ion Ratio Lower Upper

276 100

138 26.3 21.1 31.7

277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 4.728 ng
RT: 23.423 min Scan# 2964
Delta R.T. 0.003 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:252 Resp: 462788
Ion Ratio Lower Upper
252 100
253 22.1 18.4 27.6
125 10.2 10.2 15.2



#38
Benzo(k)fluoranthene
Concen: 4.651 ng
RT: 23.475 min Scan# 2982
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48

Tgt Ion:252 Resp: 467781
Ion Ratio Lower Upper
252 100
253 22.0 18.4 27.6
125 10.3 10.2 15.2









#39
Benzo(a)pyrene
Concen: 5.135 ng
RT: 24.101 min Scan# 3196
Delta R.T. 0.006 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 458318
Ion Ratio Lower Upper
252 100
253 21.5 18.6 28.0
125 11.4 11.7 17.5#



#40
Dibenzo(a,h)anthracene
Concen: 5.595 ng
RT: 26.949 min Scan# 4170
Delta R.T. 0.000 min
Lab File: BN026878.D
Acq: 10 Aug 2023 11:48


Tgt Ion:278 Resp: 441122
Ion Ratio Lower Upper
278 100
139 17.4 15.3 22.9
279 23.6 19.9 29.9





#41
 Benzo(g,h,i)perylene
 Concen: 5.035 ng
 RT: 27.750 min Scan# 44
 Delta R.T. -0.003 min
 Lab File: BN026878.D
 Acq: 10 Aug 2023 11:48

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0



Tgt Ion: 276 Resp: 469922

Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.8	29.8
138	22.9	19.0	28.4

