

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081123\
 Data File : BN026879.D
 Acq On : 10 Aug 2023 12:25
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN081123

Quant Time: Aug 11 02:06:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 01:57:19 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.430	0.436	-1.4	86	0.00
3	n-Nitrosodimethylamine	0.479	0.460	4.0	78	0.00
4 S	2-Fluorophenol	1.127	1.042	7.5	76	0.00
5 S	Phenol-d6	1.392	1.289	7.4	77	0.00
6	bis(2-Chloroethyl)ether	1.171	1.140	2.6	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
8 S	Nitrobenzene-d5	0.338	0.342	-1.2	74	0.00
9	Naphthalene	1.052	1.037	1.4	77	0.00
10	Hexachlorobutadiene	0.215	0.230	-7.0	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.636	0.679	-6.8	81	0.00
12	2-Methylnaphthalene	0.823	0.867	-5.3	80	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.213	0.181	15.0	70	0.00
15 S	2-Fluorobiphenyl	1.485	1.460	1.7	79	0.00
16	Acenaphthylene	1.909	1.843	3.5	75	0.00
17	Acenaphthene	1.179	1.167	1.0	78	0.00
18	Fluorene	1.583	1.571	0.8	76	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	76	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.007	-16.7	77	0.01
21	4-Bromophenyl-phenylether	0.228	0.219	3.9	74	0.00
22	Hexachlorobenzene	0.244	0.245	-0.4	77	0.00
23	Atrazine	0.203	0.192	5.4	72	0.00
24	Pentachlorophenol	0.123	0.103	16.3	68	0.00
25	Phenanthrene	1.150	1.123	2.3	76	0.00
26	Anthracene	1.073	0.996	7.2	73	0.00
27 SURR	Fluoranthene-d10	1.026	0.990	3.5	75	0.00
28	Fluoranthene	1.362	1.316	3.4	76	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
30	Pyrene	1.539	1.493	3.0	76	0.00
31 S	Terphenyl-d14	0.836	0.773	7.5	75	0.00
32	Benzo(a)anthracene	1.444	1.399	3.1	76	0.00
33	Chrysene	1.457	1.457	0.0	78	0.00
34	Bis(2-ethylhexyl)phthalate	1.040	0.887	14.7	65	0.00
35 I	Perylene-d12	1.000	1.000	0.0	78	0.00
36	Indeno(1,2,3-cd)pyrene	1.638	1.599	2.4	78	0.00
37	Benzo(b)fluoranthene	1.339	1.270	5.2	79	0.00
38	Benzo(k)fluoranthene	1.376	1.307	5.0	76	0.00
39 C	Benzo(a)pyrene	1.341	1.276	4.8	78	0.00
40	Dibenzo(a,h)anthracene	1.304	1.287	1.3	79	0.00
41	Benzo(g,h,i)perylene	1.403	1.386	1.2	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN081123