

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081223\
 Data File : BN026890.D
 Acq On : 11 Aug 2023 08:26
 Operator : MA/JU
 Sample : PB154594BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB154594BL

Quant Time: Aug 11 08:55:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 01:57:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.109	152	8046	0.400	ng	0.00
7) Naphthalene-d8	10.938	136	19961	0.400	ng	0.00
13) Acenaphthene-d10	14.737	164	12380	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	29605	0.400	ng	0.00
29) Chrysene-d12	21.659	240	26366	0.400	ng	0.00
35) Perylene-d12	24.206	264	29941	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.618	112	7427	0.327	ng	0.00
5) Phenol-d6	7.236	99	8740	0.312	ng	0.00
8) Nitrobenzene-d5	9.294	82	5988	0.355	ng	0.00
11) 2-Methylnaphthalene-d10	12.506	152	11462	0.361	ng	0.00
14) 2,4,6-Tribromophenol	16.226	330	1013	0.154	ng	0.00
15) 2-Fluorobiphenyl	13.368	172	17203	0.374	ng	0.00
27) Fluoranthene-d10	19.500	212	29470	0.388	ng	0.00
31) Terphenyl-d14	20.090	244	22275	0.404	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081223\
Data File : BN026890.D
Acq On : 11 Aug 2023 08:26
Operator : MA/JU
Sample : PB154594BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB154594BL

Quant Time: Aug 11 08:55:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 11 01:57:19 2023
Response via : Initial Calibration

