

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081318\
 Data File : BN002310.D
 Acq On : 13 Aug 2018 17:34
 Operator : JU/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02042

Quant Time: Aug 13 23:34:42 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 13 16:41:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.453	0.466	-2.9	96	0.00
3 S	1,4-Dioxane-d8	0.424	0.442	-4.2	94	0.00
4	Benzaldehyde	1.260	1.371	-8.8	98	0.00
5 S	Phenol-d5	1.834	1.870	-2.0	97	0.00
6	Phenol	1.868	1.930	-3.3	97	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.152	1.203	-4.4	96	0.00
8	Bis(2-Chloroethyl)ether	1.420	1.519	-7.0	98	0.00
9 S	2-Chlorophenol-d4	1.397	1.491	-6.7	98	0.00
10	2-Chlorophenol	1.406	1.485	-5.6	97	0.00
11	2-Methylphenol	1.409	1.467	-4.1	98	0.00
12	2,2'-oxybis(1-Chloropropane	2.351	2.504	-6.5	96	0.00
13 S	4-Methylphenol-d8	1.466	1.528	-4.2	97	0.00
14	Acetophenone	2.344	2.522	-7.6	97	0.00
15 P	N-Nitroso-di-n-propylamine	1.237	1.345	-8.7	98	0.00
16	4-Methylphenol	1.533	1.610	-5.0	98	0.00
17	Hexachloroethane	0.589	0.617	-4.8	97	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
19 S	Nitrobenzene-d5	0.159	0.164	-3.1	96	0.00
20	Nitrobenzene	0.381	0.401	-5.2	99	0.00
21	Isophorone	0.734	0.785	-6.9	99	0.00
22 S	2-Nitrophenol-d4	0.167	0.181	-8.4	98	0.00
23 C	2-Nitrophenol	0.176	0.190	-8.0	100	0.00
24	2,4-Dimethylphenol	0.361	0.385	-6.6	98	0.00
25	Bis(2-Chloroethoxy)methane	0.433	0.456	-5.3	98	0.00
26 S	2,4-Dichlorophenol-d3	0.308	0.326	-5.8	97	0.00
27 C	2,4-Dichlorophenol	0.298	0.316	-6.0	97	0.00
28	Naphthalene	1.029	1.073	-4.3	97	0.00
29 S	4-Chloroaniline-d4	0.333	0.353	-6.0	98	0.00
30	4-Chloroaniline	0.336	0.359	-6.8	98	0.00
31 C	Hexachlorobutadiene	0.188	0.195	-3.7	98	0.00
32	Caprolactam	0.112	0.120	-7.1	102	0.00
33 C	4-Chloro-3-methylphenol	0.343	0.366	-6.7	98	0.00
34	2-Methylnaphthalene	0.759	0.794	-4.6	98	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
36	1,2,4,5-Tetrachlorobenzene	0.606	0.635	-4.8	98	0.00
37	Hexachlorocyclopentadiene	0.368	0.372	-1.1	99	0.00
38 C	2,4,6-Trichlorophenol	0.390	0.409	-4.9	96	0.00
39	2,4,5-Trichlorophenol	0.417	0.446	-7.0	101	0.00
40	1,1'-Biphenyl	1.601	1.678	-4.8	96	0.00
41	2-Chloronaphthalene	1.248	1.313	-5.2	97	0.00
42	2-Nitroaniline	0.398	0.434	-9.0	99	0.00
43 S	Dimethylphthalate-d6	1.658	1.760	-6.2	99	0.00
44	Dimethylphthalate	1.626	1.729	-6.3	99	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081318\
 Data File : BN002310.D
 Acq On : 13 Aug 2018 17:34
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02042

Quant Time: Aug 13 23:34:42 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 13 16:41:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.329	0.358	-8.8	99	0.00
46 S	Acenaphthylene-d8	2.034	2.170	-6.7	99	0.00
47	Acenaphthylene	1.936	2.048	-5.8	97	0.00
48	3-Nitroaniline	0.313	0.339	-8.3	100	0.00
49 C	Acenaphthene	1.366	1.437	-5.2	99	0.00
50	2,4-Dinitrophenol	0.197	0.193	2.0	102	0.00
51 S	4-Nitrophenol-d4	0.313	0.324	-3.5	98	0.00
52	4-Nitrophenol	0.238	0.250	-5.0	98	0.00
53	Dibenzofuran	1.927	2.020	-4.8	98	0.00
54	2,4-Dinitrotoluene	0.495	0.541	-9.3	100	0.00
55	2,3,4,6-Tetrachlorophenol	0.364	0.391	-7.4	99	0.00
56	Diethylphthalate	1.659	1.755	-5.8	97	0.00
57 S	Fluorene-d10	1.437	1.512	-5.2	99	0.00
58	Fluorene	1.577	1.665	-5.6	98	0.00
59	4-Chlorophenyl-phenylether	0.778	0.818	-5.1	98	0.00
60	4-Nitroaniline	0.404	0.424	-5.0	96	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.127	0.130	-2.4	99	0.00
63	4,6-Dinitro-2-methylphenol	0.131	0.137	-4.6	100	0.00
64	N-Nitrosodiphenylamine	0.584	0.622	-6.5	99	0.00
65	4-Bromophenyl-phenylether	0.204	0.217	-6.4	99	0.00
66	Hexachlorobenzene	0.228	0.242	-6.1	99	0.00
67	Atrazine	0.214	0.227	-6.1	97	0.00
68 C	Pentachlorophenol	0.127	0.127	0.0	97	0.00
69	Phenanthrene	1.103	1.154	-4.6	98	0.00
70 S	Anthracene-d10	0.955	1.014	-6.2	99	0.00
71	Anthracene	1.118	1.192	-6.6	99	0.00
72	1,2,3,4-Tetrachlorobenzene	0.280	0.293	-4.6	98	0.00
73	Pentachlorobenzene	0.267	0.274	-2.6	97	0.00
74	Carbazole	0.992	1.065	-7.4	99	0.00
75	Di-n-butylphthalate	1.206	1.298	-7.6	99	0.00
76 C	Fluoranthene	1.255	1.360	-8.4	98	0.00
77 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
78 S	Pyrene-d10	0.957	1.018	-6.4	98	0.00
79	Pyrene	1.206	1.290	-7.0	98	0.00
80	Butylbenzylphthalate	0.511	0.556	-8.8	99	0.00
81	3,3'-Dichlorobenzidine	0.399	0.428	-7.3	99	0.00
82	Benzo(a)anthracene	1.225	1.302	-6.3	98	0.00
83	Bis(2-ethylhexyl)phthalate	0.766	0.828	-8.1	98	0.00
84	Chrysene	1.160	1.223	-5.4	99	0.00
85 I	Perylene-d12	1.000	1.000	0.0	95	0.00
86	Di-n-octyl phthalate	1.271	1.340	-5.4	98	0.00
87	Benzo(b)fluoranthene	1.199	1.224	-2.1	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081318\
Data File : BN002310.D
Acq On : 13 Aug 2018 17:34
Operator : JU/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTD02042

Quant Time: Aug 13 23:34:42 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Aug 13 16:41:21 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.129	1.215	-7.6	98	0.00
89 S	Benzo(a)pyrene-d12	1.046	1.095	-4.7	98	0.00
90 C	Benzo(a)pyrene	1.142	1.186	-3.9	98	0.00
91	Indeno(1,2,3-cd)pyrene	1.359	1.417	-4.3	98	0.00
92	Dibenzo(a,h)anthracene	1.136	1.186	-4.4	98	0.00
93	Benzo(g,h,i)perylene	1.134	1.175	-3.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0