

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081319\
 Data File : BN007130.D
 Acq On : 13 Aug 2019 14:07
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 14 04:48:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	134	0.00
2	n-Nitrosodimethylamine	0.400	0.332	17.0	142	0.00
3 S	2-Fluorophenol	0.400	0.422	-5.5	147	0.00
4 S	Phenol-d6	0.400	0.435	-8.7	157	0.00
5	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	141	0.00
6 I	Naphthalene-d8	0.400	0.400	0.0	143	0.00
7 S	Nitrobenzene-d5	0.400	0.410	-2.5	160	0.00
8	Naphthalene	0.400	0.375	6.3	143	0.00
9	Hexachlorobutadiene	0.400	0.381	4.8	143	-0.01
10 SURR	2-Methylnaphthalene-d10	0.400	0.369	7.8	142	0.00
11	2-Methylnaphthalene	0.400	0.372	7.0	143	0.00
12 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
13 S	2,4,6-Tribromophenol	0.400	0.335	16.3	157	0.00
14 S	2-Fluorobiphenyl	0.400	0.230	42.5#	71	0.00
15	Acenaphthylene	0.400	0.450	-12.5	163	0.00
16	Acenaphthene	0.400	0.387	3.3	141	-0.01
17	Fluorene	0.400	0.352	12.0	145	0.00
18 I	Phenanthrene-d10	0.400	0.400	0.0	122	0.00
19	4-Bromophenyl-phenylether	0.400	0.406	-1.5	126	-0.01
20	Hexachlorobenzene	0.400	0.408	-2.0	127	0.00
21	Pentachlorophenol	0.400	0.388	3.0	137	0.00
22	Phenanthrene	0.400	0.387	3.3	123	0.00
23	Anthracene	0.400	0.401	-0.3	132	-0.01
24 SURR	Fluoranthene-d10	0.400	0.358	10.5	115	0.00
25	Fluoranthene	0.400	0.368	8.0	117	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	114	-0.01
27	Pyrene	0.400	0.410	-2.5	121	0.00
28 S	Terphenyl-d14	0.400	0.398	0.5	118	-0.01
29	Benzo(a)anthracene	0.400	0.392	2.0	114	0.00
30	Chrysene	0.400	0.395	1.3	117	-0.01
31	Indeno(1,2,3-cd)pyrene	0.400	0.425	-6.2	128	-0.01
32 I	Perylene-d12	0.400	0.400	0.0	126	0.00
33	Benzo(b)fluoranthene	0.400	0.384	4.0	123	0.00
34	Benzo(k)fluoranthene	0.400	0.372	7.0	117	0.00
35 C	Benzo(a)pyrene	0.400	0.392	2.0	129	-0.01
36	Dibenzo(a,h)anthracene	0.400	0.394	1.5	129	-0.01
37	Benzo(g,h,i)perylene	0.400	0.388	3.0	124	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0