

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081319\
 Data File : BN007147.D
 Acq On : 14 Aug 2019 00:20
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 14 05:00:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	153#	0.00
2	n-Nitrosodimethylamine	0.263	0.227	13.7	169#	0.00
3 S	2-Fluorophenol	0.961	1.002	-4.3	167#	0.00
4 S	Phenol-d6	1.191	1.278	-7.3	177#	0.00
5	bis(2-Chloroethyl)ether	1.052	1.026	2.5	156#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	164#	0.00
7 S	Nitrobenzene-d5	0.323	0.321	0.6	178#	0.00
8	Naphthalene	1.122	1.048	6.6	164#	0.00
9	Hexachlorobutadiene	0.239	0.227	5.0	164#	-0.01
10 SURR	2-Methylnaphthalene-d10	0.746	0.683	8.4	162#	0.00
11	2-Methylnaphthalene	0.795	0.737	7.3	164#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	159#	0.00
13 S	2,4,6-Tribromophenol	0.215	0.172	20.0	174#	0.00
14 S	2-Fluorobiphenyl	0.582	0.308	47.1#	76	0.00
15	Acenaphthylene	2.091	2.417	-15.6	195#	0.00
16	Acenaphthene	1.339	1.251	6.6	159#	-0.01
17	Fluorene	2.004	1.717	14.3	164#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	141	0.00
19	4-Bromophenyl-phenylether	0.247	0.245	0.8	141	-0.01
20	Hexachlorobenzene	0.238	0.241	-1.3	144	0.00
21	Pentachlorophenol	0.089	0.079	11.2	145	0.00
22	Phenanthrene	1.199	1.162	3.1	141	0.00
23	Anthracene	1.095	1.073	2.0	148	-0.01
24 SURR	Fluoranthene-d10	1.419	1.259	11.3	131	0.00
25	Fluoranthene	1.533	1.395	9.0	133	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	127	-0.01
27	Pyrene	1.403	1.467	-4.6	137	0.00
28 S	Terphenyl-d14	1.090	1.110	-1.8	134	-0.01
29	Benzo(a)anthracene	1.463	1.495	-2.2	133	-0.01
30	Chrysene	1.421	1.422	-0.1	132	-0.01
31	Indeno(1,2,3-cd)pyrene	1.566	1.706	-8.9	147	-0.02
32 I	Perylene-d12	1.000	1.000	0.0	144	-0.01
33	Benzo(b)fluoranthene	1.372	1.302	5.1	139	-0.01
34	Benzo(k)fluoranthene	1.355	1.273	6.1	136	-0.01
35 C	Benzo(a)pyrene	1.244	1.210	2.7	146	-0.02
36	Dibenzo(a,h)anthracene	1.241	1.224	1.4	148	-0.02
37	Benzo(g,h,i)perylene	1.276	1.223	4.2	140	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0