

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081319\
 Data File : BN007166.D
 Acq On : 14 Aug 2019 12:58
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 14 13:37:28 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	185	0.00
2	n-Nitrosodimethylamine	0.400	0.368	8.0	217	0.00
3 S	2-Fluorophenol	0.400	0.436	-9.0	211	0.00
4 S	Phenol-d6	0.400	0.445	-11.2	222	0.00
5	bis(2-Chloroethyl)ether	0.400	0.389	2.8	188	0.00
6 I	Naphthalene-d8	0.400	0.400	0.0	195	0.00
7 S	Nitrobenzene-d5	0.400	0.423	-5.7	225	0.00
8	Naphthalene	0.400	0.380	5.0	198	0.00
9	Hexachlorobutadiene	0.400	0.394	1.5	202	-0.01
10 SURR	2-Methylnaphthalene-d10	0.400	0.372	7.0	196	0.00
11	2-Methylnaphthalene	0.400	0.386	3.5	203	0.00
12 I	Acenaphthene-d10	0.400	0.400	0.0	191	0.00
13 S	2,4,6-Tribromophenol	0.400	0.373	6.8	245	0.00
14 S	2-Fluorobiphenyl	0.400	0.222	44.5#	96	0.00
15	Acenaphthylene	0.400	0.704	-76.0#	357	0.00
16	Acenaphthene	0.400	0.412	-3.0	211	-0.01
17	Fluorene	0.400	0.387	3.3	223	0.00
18 I	Phenanthrene-d10	0.400	0.400	0.0	169	0.00
19	4-Bromophenyl-phenylether	0.400	0.401	-0.3	172	-0.01
20	Hexachlorobenzene	0.400	0.405	-1.3	173	0.00
21	Pentachlorophenol	0.400	0.434	-8.5	211	0.00
22	Phenanthrene	0.400	0.394	1.5	172	0.00
23	Anthracene	0.400	0.424	-6.0	192	0.00
24 SURR	Fluoranthene-d10	0.400	0.367	8.3	162	0.00
25	Fluoranthene	0.400	0.376	6.0	164	0.00
26 I	Chrysene-d12	0.400	0.400	0.0	153	-0.01
27	Pyrene	0.400	0.447	-11.7	176	0.00
28 S	Terphenyl-d14	0.400	0.421	-5.2	167	-0.01
29	Benzo(a)anthracene	0.400	0.422	-5.5	165	0.00
30	Chrysene	0.400	0.415	-3.7	165	0.00
31	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	173	0.00
32 I	Perylene-d12	0.400	0.400	0.0	169	0.00
33	Benzo(b)fluoranthene	0.400	0.392	2.0	168	0.00
34	Benzo(k)fluoranthene	0.400	0.382	4.5	161	0.00
35 C	Benzo(a)pyrene	0.400	0.412	-3.0	181	0.00
36	Dibenzo(a,h)anthracene	0.400	0.397	0.8	174	0.00
37	Benzo(g,h,i)perylene	0.400	0.399	0.3	170	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0