

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033368.D
 Acq On : 13 Aug 2024 18:10
 Operator : MA/JU
 Sample : SSTD02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Quant Time: Aug 14 01:43:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.569	152	227825	20.000	ng/u1	0.00
20) Naphthalene-d8	10.334	136	1075764	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.210	164	755591	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.957	188	1606824	20.000	ng/u1	0.00
79) Chrysene-d12	21.192	240	1673357	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	1984186	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	46190	8.388	ng/uL	0.00
4) Pyridine-d5	3.540	84	346276	22.338	ng/u1	0.00
7) Phenol-d5	6.746	99	421024	21.557	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.916	67	276645	24.216	ng/u1	0.00
11) 2-Chlorophenol-d4	7.105	132	329779	22.129	ng/u1	0.00
15) 4-Methylphenol-d8	8.275	113	351520	22.079	ng/u1	0.00
21) Nitrobenzene-d5	8.710	128	168230	20.453	ng/u1	0.00
24) 2-Nitrophenol-d4	9.428	143	142154	16.731	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.957	165	326830	20.604	ng/u1	0.00
31) 4-Chloroaniline-d4	10.469	131	533072	22.634	ng/u1	0.00
46) Dimethylphthalate-d6	13.634	166	1141842	21.692	ng/u1	0.00
49) Acenaphthylene-d8	13.898	160	1276969	21.098	ng/u1	0.00
54) 4-Nitrophenol-d4	14.410	143	216847	20.575	ng/u1	0.00
60) Fluorene-d10	15.210	176	982845	22.067	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.328	200	127301	15.762	ng/u1	0.00
73) Anthracene-d10	17.057	188	1547393	22.494	ng/u1	0.00
81) Pyrene-d10	19.357	212	1834160	20.483	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.833	264	2074320	21.894	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.187	88	50442	8.561	ng/uL	100
5) Pyridine	3.558	79	357743	23.080	ng/u1	100
6) Benzaldehyde	6.722	77	271234	29.608	ng/u1	100
8) Phenol	6.775	94	443180	22.296	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.011	93	362797	22.074	ng/u1	100
12) 2-Chlorophenol	7.140	128	343768	22.043	ng/u1	100
13) 2-Methylphenol	8.010	108	338285	22.035	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.093	45	602582	31.849	ng/u1	100
16) Acetophenone	8.381	105	582605	24.147	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.375	70	300899	24.066	ng/u1	100
18) 4-Methylphenol	8.334	108	382806	23.003	ng/u1	100
19) Hexachloroethane	8.640	117	142258	21.541	ng/u1	100
22) Nitrobenzene	8.752	77	421410	21.504	ng/u1	100
23) Isophorone	9.281	82	854156	21.248	ng/u1	100
25) 2-Nitrophenol	9.457	139	171862	18.514	ng/u1	100
26) 2,4-Dimethylphenol	9.528	107	408350	22.613	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.769	93	521749	20.825	ng/u1	100
29) 2,4-Dichlorophenol	9.987	162	338660	21.623	ng/u1	100
30) Naphthalene	10.387	128	1206856	21.677	ng/u1	100
32) 4-Chloroaniline	10.493	127	530100	23.183	ng/u1	100
33) Hexachlorobutadiene	10.687	225	213707	22.718	ng/u1	100
34) Caprolactam	11.246	113	82097	13.429	ng/u1	100
35) 4-Chloro-3-methylphenol	11.628	107	394511	21.550	ng/u1	100
36) 2-Methylnaphthalene	12.010	142	853890	22.438	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033368.D
 Acq On : 13 Aug 2024 18:10
 Operator : MA/JU
 Sample : SSTD02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Quant Time: Aug 14 01:43:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.228	142	863001	22.601	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.381	216	434215	21.729	ng/ul	100
40) Hexachlorocyclopentadiene	12.369	237	227504	19.754	ng/ul	100
41) 2,4,6-Trichlorophenol	12.628	196	265323	19.562	ng/ul	100
42) 2,4,5-Trichlorophenol	12.693	196	303563	20.596	ng/ul	100
43) 1,1'-Biphenyl	13.040	154	1130576	21.273	ng/ul	100
44) 2-Chloronaphthalene	13.075	162	870381	20.909	ng/ul	100
45) 2-Nitroaniline	13.275	65	258035	21.385	ng/ul	100
47) Dimethylphthalate	13.681	163	1143817	21.247	ng/ul	100
48) 2,6-Dinitrotoluene	13.787	165	210855	19.705	ng/ul	100
50) Acenaphthylene	13.928	152	1391563	20.343	ng/ul	100
51) 3-Nitroaniline	14.110	138	239254	20.912	ng/ul	100
52) Acenaphthene	14.275	153	982914	21.717	ng/ul	100
53) 2,4-Dinitrophenol	14.316	184	81672	15.118	ng/ul	100
55) 4-Nitrophenol	14.422	109	169698	21.705	ng/ul	100
56) Dibenzofuran	14.610	168	1371665	22.306	ng/ul	100
57) 2,4-Dinitrotoluene	14.575	165	318281	20.271	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.840	232	244318	19.817	ng/ul	100
59) Diethylphthalate	15.063	149	1207897	21.744	ng/ul	100
61) Fluorene	15.263	166	1148587	23.195	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.269	204	535234	22.316	ng/ul	100
63) 4-Nitroaniline	15.281	138	263119	24.158	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.339	198	149585	17.422	ng/ul	100
67) N-Nitrosodiphenylamine	15.481	169	983448	22.204	ng/ul	100
68) 4-Bromophenyl-phenylether	16.163	248	332042	21.328	ng/ul	100
69) Hexachlorobenzene	16.269	284	366374	20.763	ng/ul	100
70) Atrazine	16.445	200	371524	26.567	ng/ul	100
71) Pentachlorophenol	16.610	266	223976	19.839	ng/ul	100
72) Phenanthrene	17.004	178	1836883	22.441	ng/ul	100
74) Anthracene	17.092	178	1857756	22.399	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	12.998	216	433718	20.971	ng/ul	100
76) Pentachlorobenzene	14.534	250	456909	22.682	ng/ul	100
77) Carbazole	17.363	167	1771632	25.051	ng/ul	100
78) Di-n-butylphthalate	17.963	149	2184190	22.448	ng/ul	100
80) Fluoranthene	19.028	202	2171888	20.295	ng/ul	100
82) Pyrene	19.392	202	2309068	21.030	ng/ul	100
83) Butylbenzylphthalate	20.327	149	965880	18.625	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.110	252	785008	24.464	ng/ul	100
85) Benzo(a)anthracene	21.180	228	2340023	20.892	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.151	149	1547725	20.474	ng/ul	100
87) Chrysene	21.233	228	2249548	21.502	ng/ul	100
89) Di-n-octyl phthalate	22.239	149	2563103	19.295	ng/ul	100
90) Benzo(b)fluoranthene	23.139	252	2440470	20.577	ng/ul	100
91) Benzo(k)fluoranthene	23.198	252	2443866	20.999	ng/ul	100
93) Benzo(a)pyrene	23.898	252	2351867	21.173	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.174	276	2900681	22.662	ng/ul	100
95) Dibenzo(a,h)anthracene	27.245	278	2354788	22.533	ng/ul	100
96) Benzo(g,h,i)perylene	28.162	276	2215334	21.772	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033368.D
 Acq On : 13 Aug 2024 18:10
 Operator : MA/JU
 Sample : SSTD02076
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Quant Time: Aug 14 01:43:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.561 ng/uL
RT: 3.187 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 88 Resp: 50442
Ion Ratio Lower Upper
88 100
43 31.8 25.4 38.2
58 76.3 61.0 91.6



#5
Pyridine
Concen: 23.080 ng/uL
RT: 3.558 min Scan# 80
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion: 79 Resp: 357743
Ion Ratio Lower Upper
79 100
52 86.1 68.9 103.3
51 38.2 30.6 45.8









#6
Benzaldehyde
Concen: 29.608 ng/ul
RT: 6.722 min Scan# 618
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 77 Resp: 271234
Ion Ratio Lower Upper
77 100
105 99.2 79.4 119.0
106 94.9 75.9 113.9



#8
Phenol
Concen: 22.296 ng/ul
RT: 6.775 min Scan# 627
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion: 94 Resp: 443180
Ion Ratio Lower Upper
94 100
65 27.6 22.1 33.1
66 35.8 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 22.074 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 93 Resp: 362797
Ion Ratio Lower Upper
93 100
63 79.4 63.5 95.3
95 32.1 25.7 38.5



#12
2-Chlorophenol
Concen: 22.043 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:128 Resp: 343768
Ion Ratio Lower Upper
128 100
64 47.3 37.8 56.8
130 32.0 25.6 38.4





#13

2-Methylphenol

Concen: 22.035 ng/ul

RT: 8.010 min Scan# 81

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :





SSTD020212

Tgt Ion:108 Resp: 338285

Ion Ratio Lower Upper

108 100

107 90.7 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.849 ng/ul

RT: 8.093 min Scan# 851

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion: 45 Resp: 602582


Ion Ratio Lower Upper

45 100

77 12.3 9.8 14.8

79 8.9 7.1 10.7









#16
Acetophenone
Concen: 24.147 ng/ul
RT: 8.381 min Scan# 900
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:105 Resp: 582605
Ion Ratio Lower Upper
105 100
77 77.1 61.7 92.5
51 28.7 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 24.066 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion: 70 Resp: 300899
Ion Ratio Lower Upper
70 100
42 61.8 49.4 74.2
101 9.7 7.8 11.6
130 20.5 16.4 24.6









#18
4-Methylphenol
Concen: 23.003 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:108 Resp: 382806
Ion Ratio Lower Upper
108 100
107 111.1 88.9 133.3



#19
Hexachloroethane
Concen: 21.541 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:117 Resp: 142258
Ion Ratio Lower Upper
117 100
201 93.9 75.1 112.7
199 60.2 48.2 72.2









#22
Nitrobenzene
Concen: 21.504 ng/ul
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 77 Resp: 421410
Ion Ratio Lower Upper
77 100
123 46.0 36.8 55.2
65 13.4 10.7 16.1



#23
Isophorone
Concen: 21.248 ng/ul
RT: 9.281 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion: 82 Resp: 854156
Ion Ratio Lower Upper
82 100
95 6.6 5.3 7.9
138 17.8 14.2 21.4









#25
2-Nitrophenol
Concen: 18.514 ng/ul
RT: 9.457 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:139 Resp: 171862
Ion Ratio Lower Upper
139 100
65 53.0 42.4 63.6
109 30.0 24.0 36.0



#26
2,4-Dimethylphenol
Concen: 22.613 ng/ul
RT: 9.528 min Scan# 1095
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:107 Resp: 408350
Ion Ratio Lower Upper
107 100
121 49.9 39.9 59.9
122 84.4 67.5 101.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.825 ng/ul

RT: 9.769 min Scan# 1136

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion: 93 Resp: 521749

Ion Ratio Lower Upper

93 100

95 30.9 24.7 37.1

123 11.4 9.1 13.7



#29

2,4-Dichlorophenol

Concen: 21.623 ng/ul

RT: 9.987 min Scan# 1173

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion: 162 Resp: 338660


Ion Ratio Lower Upper

162 100

164 65.7 52.6 78.8

98 37.2 29.8 44.6





#30

Naphthalene

Concen: 21.677 ng/ul

RT: 10.387 min Scan# 1241

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:128 Resp: 1206856

Ion Ratio Lower Upper

128 100

129 11.0 8.8 13.2

127 13.0 10.4 15.6



#32

4-Chloroaniline

Concen: 23.183 ng/ul

RT: 10.493 min Scan# 1259

Delta R.T. 0.000 min

Lab File: BN033368.D




Acq: 13 Aug 2024 18:10


Tgt Ion:127 Resp: 530100

Ion Ratio Lower Upper

127 100

129 31.5 25.2 37.8





#33

Hexachlorobutadiene

Concen: 22.718 ng/ul

RT: 10.687 min Scan# 1292

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:225 Resp: 213707

Ion Ratio Lower Upper

225 100

223 62.6 50.1 75.1

227 60.1 48.1 72.1



#34

Caprolactam

Concen: 13.429 ng/ul

RT: 11.246 min Scan# 1387

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:113 Resp: 82097


Ion Ratio Lower Upper

113 100

55 182.7 146.2 219.2

56 131.4 105.1 157.7









#35
4-Chloro-3-methylphenol
Concen: 21.550 ng/ul
RT: 11.628 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:107 Resp: 394511
Ion Ratio Lower Upper
107 100
144 26.5 21.2 31.8
142 80.5 64.4 96.6



#36
2-Methylnaphthalene
Concen: 22.438 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:142 Resp: 853890
Ion Ratio Lower Upper
142 100
141 90.2 72.2 108.2









#37
 1-Methylnaphthalene
 Concen: 22.601 ng/ul
 RT: 12.228 min Scan# 1554
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212


Tgt Ion:142 Resp: 863001
 Ion Ratio Lower Upper
 142 100
 141 92.3 73.8 110.8
 116 3.7 3.0 4.4



#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 21.729 ng/ul
 RT: 12.381 min Scan# 1580
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Tgt Ion:216 Resp: 434215
 Ion Ratio Lower Upper
 216 100
 214 70.2 56.2 84.2
 179 21.4 17.1 25.7
 108 18.9 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 19.754 ng/ul
RT: 12.369 min Scan# 1578
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:237 Resp: 227504
Ion Ratio Lower Upper
237 100
235 64.0 51.2 76.8
272 10.8 8.6 13.0



#41
2,4,6-Trichlorophenol
Concen: 19.562 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:196 Resp: 265323
Ion Ratio Lower Upper
196 100
198 98.1 78.5 117.7
200 31.5 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 20.596 ng/ul

RT: 12.693 min Scan# 1633

Delta R.T. 0.000 min

Lab File: BN033368.D


Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212






Tgt Ion:196 Resp: 303563

Ion Ratio Lower Upper

196 100

198 95.5 76.4 114.6

200 30.1 24.1 36.1



#43

1,1'-Biphenyl

Concen: 21.273 ng/ul

RT: 13.040 min Scan# 1692

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:154 Resp: 1130576


Ion Ratio Lower Upper

154 100

153 40.0 32.0 48.0

76 14.3 11.4 17.2





#44

2-Chloronaphthalene

Concen: 20.909 ng/ul

RT: 13.075 min Scan# 1698

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:162 Resp: 870381

Ion Ratio Lower Upper

162 100

164 33.7 27.0 40.4

127 39.8 31.8 47.8



#45

2-Nitroaniline

Concen: 21.385 ng/ul

RT: 13.275 min Scan# 1732

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion: 65 Resp: 258035


Ion Ratio Lower Upper

65 100

92 69.5 55.6 83.4

138 104.6 83.7 125.5









#47
 Dimethylphthalate
 Concen: 21.247 ng/ul
 RT: 13.681 min Scan# 1801
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Tgt Ion:163 Resp: 1143817




Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.3	4.9
164	10.5	8.4	12.6




#48
 2,6-Dinitrotoluene
 Concen: 19.705 ng/ul
 RT: 13.787 min Scan# 1819
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Tgt Ion:165 Resp: 210855

Ion	Ratio	Lower	Upper
165	100		
89	58.5	46.8	70.2
121	22.6	18.1	27.1





#50

Acenaphthylene

Concen: 20.343 ng/ul

RT: 13.928 min Scan# 1843

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:152 Resp: 1391563

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.4

153 13.6 10.9 16.3



#51

3-Nitroaniline

Concen: 20.912 ng/ul

RT: 14.110 min Scan# 1874

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:138 Resp: 239254


Ion Ratio Lower Upper

138 100

108 13.1 10.5 15.7

92 116.8 93.4 140.2





#52

Acenaphthene

Concen: 21.717 ng/ul

RT: 14.275 min Scan# 1902

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:153 Resp: 982914

Ion Ratio Lower Upper

153 100

152 47.5 38.0 57.0

154 89.4 71.5 107.3



#53

2,4-Dinitrophenol

Concen: 15.118 ng/ul

RT: 14.316 min Scan# 1909

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:184 Resp: 81672


Ion Ratio Lower Upper

184 100

63 69.0 55.2 82.8

154 64.2 51.4 77.0









#55
4-Nitrophenol
Concen: 21.705 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:109 Resp: 169698
Ion Ratio Lower Upper
109 100
139 132.8 106.2 159.4
65 138.5 110.8 166.2



#56
Dibenzofuran
Concen: 22.306 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:168 Resp: 1371665
Ion Ratio Lower Upper
168 100
139 37.7 30.2 45.2









#57
 2,4-Dinitrotoluene
 Concen: 20.271 ng/ul
 RT: 14.575 min Scan# 1953
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Instrument :
 BNA_N
 ClientSampleId :
 SST020212

Tgt Ion:165 Resp: 318281




Ion	Ratio	Lower	Upper
165	100		
63	44.2	35.4	53.0
182	2.7	2.2	3.2




#58
 2,3,4,6-Tetrachlorophenol
 Concen: 19.817 ng/ul
 RT: 14.840 min Scan# 1998
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Tgt Ion:232 Resp: 244318

Ion	Ratio	Lower	Upper
232	100		
131	49.2	39.4	59.0
130	2.5	2.0	3.0
166	32.6	26.1	39.1





#59

Diethylphthalate

Concen: 21.744 ng/ul

RT: 15.063 min Scan# 2036

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:149 Resp: 1207897

Ion Ratio Lower Upper

149 100

177 21.3 17.0 25.6

150 12.3 9.8 14.8



#61

Fluorene

Concen: 23.195 ng/ul

RT: 15.263 min Scan# 2070

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:166 Resp: 1148587


Ion Ratio Lower Upper

166 100

165 97.2 77.8 116.6

167 13.1 10.5 15.7









#62
4-Chlorophenyl-phenylether
Concen: 22.316 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:204 Resp: 535234
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.2
141 63.0 50.4 75.6



#63
4-Nitroaniline
Concen: 24.158 ng/ul
RT: 15.281 min Scan# 2073
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:138 Resp: 263119
Ion Ratio Lower Upper
138 100
92 51.4 41.1 61.7
108 72.6 58.1 87.1









#66
4,6-Dinitro-2-methylphenol
Concen: 17.422 ng/ul
RT: 15.339 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:198 Resp: 149585
Ion Ratio Lower Upper
198 100
182 4.4 3.5 5.3
77 27.0 21.6 32.4



#67
N-Nitrosodiphenylamine
Concen: 22.204 ng/ul
RT: 15.481 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:169 Resp: 983448
Ion Ratio Lower Upper
169 100
168 68.1 54.5 81.7
167 36.5 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 21.328 ng/ul
RT: 16.163 min Scan# 2223
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:248 Resp: 332042
Ion Ratio Lower Upper
248 100
250 94.7 75.8 113.6
141 78.8 63.0 94.6



#69
Hexachlorobenzene
Concen: 20.763 ng/ul
RT: 16.269 min Scan# 2241
Delta R.T. 0.000 min
Lab File: BN033368.D
Acq: 13 Aug 2024 18:10

Tgt Ion:284 Resp: 366374
Ion Ratio Lower Upper
284 100
142 40.6 32.5 48.7
249 37.1 29.7 44.5









#70
 Atrazine
 Concen: 26.567 ng/ul
 RT: 16.445 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Tgt Ion: 200 Resp: 371524




Ion	Ratio	Lower	Upper
200	100		
173	25.8	20.6	31.0
215	46.0	36.8	55.2




#71
 Pentachlorophenol
 Concen: 19.839 ng/ul
 RT: 16.610 min Scan# 2299
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Tgt Ion: 266 Resp: 223976

Ion	Ratio	Lower	Upper
266	100		
264	60.7	48.6	72.8
268	63.1	50.5	75.7





#72

Phenanthrene

Concen: 22.441 ng/ul

RT: 17.004 min Scan# 2366

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:178 Resp: 1836883

Ion Ratio Lower Upper

178 100

179 14.6 11.7 17.5

176 18.9 15.1 22.7



#74

Anthracene

Concen: 22.399 ng/ul

RT: 17.092 min Scan# 2381

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:178 Resp: 1857756


Ion Ratio Lower Upper

178 100

179 14.7 11.8 17.6

176 17.8 14.2 21.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.971 ng/uL

RT: 12.998 min Scan# 1685

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:216 Resp: 433718

Ion Ratio Lower Upper

216 100

214 82.1 65.7 98.5

218 52.2 41.8 62.6



#76

Pentachlorobenzene

Concen: 22.682 ng/uL

RT: 14.534 min Scan# 1946

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:250 Resp: 456909


Ion Ratio Lower Upper

250 100

248 62.3 49.8 74.8

252 62.4 49.9 74.9





#77

Carbazole

Concen: 25.051 ng/ul

RT: 17.363 min Scan# 2427

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :





BNA_N

ClientSampleId :

SSTD020212

Tgt Ion:167 Resp: 1771632

Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.4	26.0
139	12.9	10.3	15.5



#78

Di-n-butylphthalate

Concen: 22.448 ng/ul

RT: 17.963 min Scan# 2529




Delta R.T. 0.000 min


Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Tgt Ion:149 Resp: 2184190

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.1	10.7
104	4.8	3.8	5.8





#80

Fluoranthene

Concen: 20.295 ng/ul

RT: 19.028 min Scan# 2710

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:202 Resp: 2171888

Ion Ratio Lower Upper

202 100

101 12.4 9.9 14.9

100 9.4 7.5 11.3



#82

Pyrene

Concen: 21.030 ng/ul

RT: 19.392 min Scan# 2772

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:202 Resp: 2309068


Ion Ratio Lower Upper

202 100

101 14.2 11.4 17.0

100 12.0 9.6 14.4





#83

Butylbenzylphthalate

Concen: 18.625 ng/ul

RT: 20.327 min Scan# 2931

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:149 Resp: 965880

Ion Ratio Lower Upper

149 100

91 69.4 55.5 83.3

206 19.7 15.8 23.6



#84

3,3'-Dichlorobenzidine

Concen: 24.464 ng/ul

RT: 21.110 min Scan# 3064

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:252 Resp: 785008


Ion Ratio Lower Upper

252 100

254 62.3 49.8 74.8

126 12.5 10.0 15.0









#85
 Benzo(a)anthracene
 Concen: 20.892 ng/ul
 RT: 21.180 min Scan# 3076
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Tgt Ion:228 Resp: 2340023




Ion	Ratio	Lower	Upper
228	100		
229	20.8	16.6	25.0
226	26.6	21.3	31.9




#86
 Bis(2-ethylhexyl)phthalate
 Concen: 20.474 ng/ul
 RT: 21.151 min Scan# 3071
 Delta R.T. 0.000 min
 Lab File: BN033368.D
 Acq: 13 Aug 2024 18:10

Tgt Ion:149 Resp: 1547725

Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.5	32.3
279	3.9	3.1	4.7





#87

Chrysene

Concen: 21.502 ng/ul

RT: 21.233 min Scan# 3085

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:228 Resp: 2249548

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.9 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 19.295 ng/ul




RT: 22.239 min Scan# 3256


Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Tgt Ion:149 Resp: 2563103





#90

Benzo(b)fluoranthene

Concen: 20.577 ng/ul

RT: 23.139 min Scan# 3409

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:252 Resp: 2440470

Ion Ratio Lower Upper

252 100

253 23.2 18.6 27.8

125 12.1 9.7 14.5



#91

Benzo(k)fluoranthene

Concen: 20.999 ng/ul

RT: 23.198 min Scan# 3419

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:252 Resp: 2443866


Ion Ratio Lower Upper

252 100

253 22.8 18.2 27.4

125 11.6 9.3 13.9





#93

Benzo(a)pyrene

Concen: 21.173 ng/ul

RT: 23.898 min Scan# 3538

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:252 Resp: 2351867

Ion Ratio Lower Upper

252 100

253 20.7 16.6 24.8

125 12.0 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.662 ng/ul

RT: 27.174 min Scan# 4095

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10




Tgt Ion:276 Resp: 2900681


Ion Ratio Lower Upper

276 100

138 22.1 17.7 26.5

277 21.7 17.4 26.0





#95

Dibenzo(a,h)anthracene

Concen: 22.533 ng/ul

RT: 27.245 min Scan# 4107

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:278 Resp: 2354788

Ion Ratio Lower Upper

278 100

139 18.3 14.6 22.0

279 23.1 18.5 27.7



#96

Benzo(g,h,i)perylene

Concen: 21.772 ng/ul

RT: 28.162 min Scan# 4263

Delta R.T. 0.000 min

Lab File: BN033368.D

Acq: 13 Aug 2024 18:10

Tgt Ion:276 Resp: 2215334

Ion Ratio Lower Upper

276 100

138 25.6 20.5 30.7

277 22.8 18.2 27.4

