

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033369.D
 Acq On : 13 Aug 2024 18:49
 Operator : MA/JU
 Sample : SST04077
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040213

Quant Time: Aug 14 01:44:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.569	152	249933	20.000	ng/ul	0.00
20) Naphthalene-d8	10.340	136	1134952	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.210	164	770901	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.957	188	1677887	20.000	ng/ul	0.00
79) Chrysene-d12	21.198	240	1636532	20.000	ng/ul	0.00
88) Perylene-d12	24.027	264	1885667	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	95626	15.829	ng/uL	0.00
4) Pyridine-d5	3.540	84	733447	43.128	ng/ul	0.00
7) Phenol-d5	6.752	99	893958	41.722	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.916	67	585136	46.688	ng/ul	0.00
11) 2-Chlorophenol-d4	7.110	132	705503	43.154	ng/ul	0.00
15) 4-Methylphenol-d8	8.275	113	752942	43.110	ng/ul	0.00
21) Nitrobenzene-d5	8.710	128	368533	42.468	ng/ul	0.00
24) 2-Nitrophenol-d4	9.428	143	344584	38.440	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.963	165	719361	42.985	ng/ul	0.00
31) 4-Chloroaniline-d4	10.475	131	1132678	45.586	ng/ul	0.00
46) Dimethylphthalate-d6	13.634	166	2372196	44.170	ng/ul	0.00
49) Acenaphthylene-d8	13.898	160	2712083	43.918	ng/ul	0.00
54) 4-Nitrophenol-d4	14.416	143	468109	43.533	ng/ul	0.00
60) Fluorene-d10	15.210	176	2008398	44.198	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.334	200	323396	38.345	ng/ul	0.00
73) Anthracene-d10	17.063	188	3156705	43.944	ng/ul	0.00
81) Pyrene-d10	19.363	212	3695641	42.200	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.845	264	4081974	45.335	ng/ul	0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.187	88	101688	15.732	ng/uL	97
5) Pyridine	3.558	79	744237	43.768	ng/ul	99
6) Benzaldehyde	6.722	77	515896	51.334	ng/ul	99
8) Phenol	6.775	94	941341	43.169	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.010	93	756208	41.940	ng/ul	100
12) 2-Chlorophenol	7.140	128	723226	42.273	ng/ul	99
13) 2-Methylphenol	8.010	108	721572	42.844	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.099	45	1231437	59.329	ng/ul	98
16) Acetophenone	8.387	105	1199923	45.333	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.381	70	622798	45.406	ng/ul	100
18) 4-Methylphenol	8.340	108	799831	43.810	ng/ul	99
19) Hexachloroethane	8.640	117	299529	41.344	ng/ul	97
22) Nitrobenzene	8.757	77	891977	43.142	ng/ul	100
23) Isophorone	9.287	82	1807939	42.628	ng/ul	100
25) 2-Nitrophenol	9.457	139	394822	40.314	ng/ul	99
26) 2,4-Dimethylphenol	9.534	107	854246	44.837	ng/ul	97
27) Bis(2-Chloroethoxy)met...	9.769	93	1092846	41.344	ng/ul	97
29) 2,4-Dichlorophenol	9.987	162	736663	44.582	ng/ul	99
30) Naphthalene	10.387	128	2518256	42.872	ng/ul	99
32) 4-Chloroaniline	10.499	127	1094356	45.365	ng/ul	98
33) Hexachlorobutadiene	10.687	225	437610	44.095	ng/ul	90
34) Caprolactam	11.263	113	159754	24.769	ng/ul	99
35) 4-Chloro-3-methylphenol	11.634	107	840652	43.525	ng/ul	98
36) 2-Methylnaphthalene	12.010	142	1792585	44.648	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033369.D
 Acq On : 13 Aug 2024 18:49
 Operator : MA/JU
 Sample : SST04077
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040213


Quant Time: Aug 14 01:44:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.234	142	1786347	44.343	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.381	216	893768	43.838	ng/ul	91
40) Hexachlorocyclopentadiene	12.375	237	514963	43.825	ng/ul	96
41) 2,4,6-Trichlorophenol	12.628	196	587068	42.424	ng/ul	96
42) 2,4,5-Trichlorophenol	12.698	196	666059	44.293	ng/ul	96
43) 1,1'-Biphenyl	13.040	154	2332425	43.015	ng/ul	99
44) 2-Chloronaphthalene	13.075	162	1811259	42.648	ng/ul	99
45) 2-Nitroaniline	13.281	65	571210	46.399	ng/ul	97
47) Dimethylphthalate	13.687	163	2359030	42.950	ng/ul	99
48) 2,6-Dinitrotoluene	13.792	165	480936	44.051	ng/ul	95
50) Acenaphthylene	13.928	152	2903353	41.601	ng/ul	99
51) 3-Nitroaniline	14.116	138	519439	44.499	ng/ul	98
52) Acenaphthene	14.275	153	2041093	44.201	ng/ul	98
53) 2,4-Dinitrophenol	14.322	184	210392	38.172	ng/ul	96
55) 4-Nitrophenol	14.434	109	359424	45.059	ng/ul	95
56) Dibenzofuran	14.616	168	2820336	44.954	ng/ul	99
57) 2,4-Dinitrotoluene	14.581	165	707593	44.170	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.839	232	540045	42.934	ng/ul	98
59) Diethylphthalate	15.069	149	2485558	43.856	ng/ul	98
61) Fluorene	15.269	166	2292434	45.375	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.275	204	1106404	45.214	ng/ul	97
63) 4-Nitroaniline	15.287	138	537331	48.355	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.345	198	367603	41.001	ng/ul	97
67) N-Nitrosodiphenylamine	15.486	169	2034658	43.993	ng/ul	97
68) 4-Bromophenyl-phenylether	16.163	248	693609	42.665	ng/ul	97
69) Hexachlorobenzene	16.269	284	794871	43.139	ng/ul	93
70) Atrazine	16.451	200	756216	51.786	ng/ul	98
71) Pentachlorophenol	16.610	266	508426	43.128	ng/ul	98
72) Phenanthrene	17.004	178	3692578	43.201	ng/ul	98
74) Anthracene	17.092	178	3773911	43.574	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.998	216	936166	43.348	ng/ul	94
76) Pentachlorobenzene	14.534	250	899771	42.774	ng/ul	91
77) Carbazole	17.363	167	3545354	48.007	ng/ul	99
78) Di-n-butylphthalate	17.963	149	4525904	44.545	ng/ul	99
80) Fluoranthene	19.027	202	4346793	41.533	ng/ul	99
82) Pyrene	19.392	202	4616026	42.987	ng/ul	99
83) Butylbenzylphthalate	20.327	149	2036810	40.160	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.116	252	1598083	50.922	ng/ul	99
85) Benzo(a)anthracene	21.180	228	4796181	43.785	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.151	149	3243818	43.876	ng/ul	100
87) Chrysene	21.239	228	4334962	42.367	ng/ul	96
89) Di-n-octyl phthalate	22.245	149	5477728	43.391	ng/ul	100
90) Benzo(b)fluoranthene	23.145	252	4820046	42.765	ng/ul	100
91) Benzo(k)fluoranthene	23.210	252	4893424	44.245	ng/ul	98
93) Benzo(a)pyrene	23.904	252	4610716	43.677	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.198	276	5820698	47.850	ng/ul	97
95) Dibenzo(a,h)anthracene	27.256	278	4755816	47.887	ng/ul	99
96) Benzo(g,h,i)perylene	28.180	276	4441325	45.929	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 14 01:44:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 14 01:18:11 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 15.732 ng/uL
 RT: 3.187 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49




Instrument :
 BNA_N
 ClientSampleId :
 SST040213


Tgt Ion: 88 Resp: 101688
 Ion Ratio Lower Upper
 88 100
 43 31.7 25.4 38.2
 58 79.6 61.0 91.6



#5
 Pyridine
 Concen: 43.768 ng/uL
 RT: 3.558 min Scan# 80
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Tgt Ion: 79 Resp: 744237
 Ion Ratio Lower Upper
 79 100
 52 85.0 68.9 103.3
 51 38.1 30.6 45.8









#6
Benzaldehyde
Concen: 51.334 ng/ul
RT: 6.722 min Scan# 618
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 77 Resp: 515896
Ion Ratio Lower Upper
77 100
105 100.8 79.4 119.0
106 94.5 75.9 113.9



#8
Phenol
Concen: 43.169 ng/ul
RT: 6.775 min Scan# 627
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion: 94 Resp: 941341
Ion Ratio Lower Upper
94 100
65 28.7 22.1 33.1
66 36.9 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 41.940 ng/ul
RT: 7.010 min Scan# 667
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 93 Resp: 756208
Ion Ratio Lower Upper
93 100
63 79.3 63.5 95.3
95 32.2 25.7 38.5



#12
2-Chlorophenol
Concen: 42.273 ng/ul
RT: 7.140 min Scan# 689
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:128 Resp: 723226
Ion Ratio Lower Upper
128 100
64 48.4 37.8 56.8
130 32.7 25.6 38.4





#13

2-Methylphenol

Concen: 42.844 ng/ul

RT: 8.010 min Scan# 81

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :





SSTD040213

Tgt Ion:108 Resp: 721572

Ion Ratio Lower Upper

108 100

107 89.4 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 59.329 ng/ul

RT: 8.099 min Scan# 852

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion: 45 Resp: 1231437


Ion Ratio Lower Upper

45 100

77 12.9 9.8 14.8

79 9.4 7.1 10.7





#16

Acetophenone

Concen: 45.333 ng/ul

RT: 8.387 min Scan# 901

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:105 Resp: 1199923

Ion Ratio Lower Upper

105 100

77 76.3 61.7 92.5

51 27.4 23.0 34.4



#17

N-Nitroso-di-n-propylamine

Concen: 45.406 ng/ul

RT: 8.381 min Scan# 900

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Tgt Ion: 70 Resp: 622798




Ion Ratio Lower Upper


70 100

42 62.0 49.4 74.2

101 9.7 7.8 11.6

130 20.4 16.4 24.6





#18

4-Methylphenol

Concen: 43.810 ng/ul

RT: 8.340 min Scan# 892

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :





SSTD040213

Tgt Ion:108 Resp: 799831

Ion Ratio Lower Upper

108 100

107 112.4 88.9 133.3



#19

Hexachloroethane

Concen: 41.344 ng/ul

RT: 8.640 min Scan# 944

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:117 Resp: 299529


Ion Ratio Lower Upper

117 100

201 98.4 75.1 112.7

199 59.6 48.2 72.2









#22
Nitrobenzene
Concen: 43.142 ng/ul
RT: 8.757 min Scan# 90
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion: 77 Resp: 891977
Ion Ratio Lower Upper
77 100
123 45.8 36.8 55.2
65 13.1 10.7 16.1



#23
Isophorone
Concen: 42.628 ng/ul
RT: 9.287 min Scan# 1054
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion: 82 Resp: 1807939
Ion Ratio Lower Upper
82 100
95 6.9 5.3 7.9
138 17.9 14.2 21.4









#25
2-Nitrophenol
Concen: 40.314 ng/ul
RT: 9.457 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:139 Resp: 394822
Ion Ratio Lower Upper
139 100
65 51.4 42.4 63.6
109 30.0 24.0 36.0



#26
2,4-Dimethylphenol
Concen: 44.837 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:107 Resp: 854246
Ion Ratio Lower Upper
107 100
121 51.8 39.9 59.9
122 86.8 67.5 101.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 41.344 ng/ul
 RT: 9.769 min Scan# 1136
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040213

Tgt Ion: 93 Resp: 1092846




Ion	Ratio	Lower	Upper
93	100		
95	33.2	24.7	37.1
123	11.3	9.1	13.7




#29
 2,4-Dichlorophenol
 Concen: 44.582 ng/ul
 RT: 9.987 min Scan# 1173
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Tgt Ion:162 Resp: 736663

Ion	Ratio	Lower	Upper
162	100		
164	65.0	52.6	78.8
98	37.8	29.8	44.6









#30
Naphthalene
Concen: 42.872 ng/ul
RT: 10.387 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:128 Resp: 2518256
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 45.365 ng/ul
RT: 10.499 min Scan# 1260
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:127 Resp: 1094356
Ion Ratio Lower Upper
127 100
129 32.7 25.2 37.8









#33
Hexachlorobutadiene
Concen: 44.095 ng/ul
RT: 10.687 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:225 Resp: 437610
Ion Ratio Lower Upper
225 100
223 67.5 50.1 75.1
227 70.9 48.1 72.1



#34
Caprolactam
Concen: 24.769 ng/ul
RT: 11.263 min Scan# 1390
Delta R.T. 0.018 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:113 Resp: 159754
Ion Ratio Lower Upper
113 100
55 184.3 146.2 219.2
56 133.5 105.1 157.7









#35
4-Chloro-3-methylphenol
Concen: 43.525 ng/ul
RT: 11.634 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:107 Resp: 840652
Ion Ratio Lower Upper
107 100
144 26.5 21.2 31.8
142 83.1 64.4 96.6



#36
2-Methylnaphthalene
Concen: 44.648 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:142 Resp: 1792585
Ion Ratio Lower Upper
142 100
141 87.2 72.2 108.2





#37

1-Methylnaphthalene

Concen: 44.343 ng/ul

RT: 12.234 min Scan# 1555

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:142 Resp: 1786347

Ion Ratio Lower Upper

142 100

141 91.7 73.8 110.8

116 3.8 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 43.838 ng/ul

RT: 12.381 min Scan# 1580

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Tgt Ion:216 Resp: 893768




Ion Ratio Lower Upper


216 100

214 81.2 56.2 84.2

179 21.8 17.1 25.7

108 19.2 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 43.825 ng/ul
RT: 12.375 min Scan# 1579
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:237 Resp: 514963
Ion Ratio Lower Upper
237 100
235 60.6 51.2 76.8
272 12.9 8.6 13.0



#41
2,4,6-Trichlorophenol
Concen: 42.424 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:196 Resp: 587068
Ion Ratio Lower Upper
196 100
198 93.5 78.5 117.7
200 29.8 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 44.293 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:196 Resp: 666059
Ion Ratio Lower Upper
196 100
198 99.5 76.4 114.6
200 31.0 24.1 36.1



#43
1,1'-Biphenyl
Concen: 43.015 ng/ul
RT: 13.040 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:154 Resp: 2332425
Ion Ratio Lower Upper
154 100
153 39.8 32.0 48.0
76 13.7 11.4 17.2









#44
2-Chloronaphthalene
Concen: 42.648 ng/ul
RT: 13.075 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:162 Resp: 1811259
Ion Ratio Lower Upper
162 100
164 32.9 27.0 40.4
127 40.4 31.8 47.8



#45
2-Nitroaniline
Concen: 46.399 ng/ul
RT: 13.281 min Scan# 1733
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion: 65 Resp: 571210
Ion Ratio Lower Upper
65 100
92 69.6 55.6 83.4
138 109.4 83.7 125.5









#47
Dimethylphthalate
Concen: 42.950 ng/ul
RT: 13.687 min Scan# 1801
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:163 Resp: 2359030
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.4 12.6



#48
2,6-Dinitrotoluene
Concen: 44.051 ng/ul
RT: 13.792 min Scan# 1820
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:165 Resp: 480936
Ion Ratio Lower Upper
165 100
89 54.3 46.8 70.2
121 20.6 18.1 27.1





#50

Acenaphthylene

Concen: 41.601 ng/ul

RT: 13.928 min Scan# 1843

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:152 Resp: 2903353

Ion Ratio Lower Upper

152 100

151 19.8 16.2 24.4

153 13.3 10.9 16.3



#51

3-Nitroaniline

Concen: 44.499 ng/ul

RT: 14.116 min Scan# 1875

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:138 Resp: 519439


Ion Ratio Lower Upper

138 100

108 11.6 10.5 15.7

92 115.0 93.4 140.2









#52
Acenaphthene
Concen: 44.201 ng/ul
RT: 14.275 min Scan# 1910
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:153 Resp: 2041093
Ion Ratio Lower Upper
153 100
152 46.8 38.0 57.0
154 87.8 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 38.172 ng/ul
RT: 14.322 min Scan# 1910
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:184 Resp: 210392
Ion Ratio Lower Upper
184 100
63 64.4 55.2 82.8
154 62.9 51.4 77.0









#55
4-Nitrophenol
Concen: 45.059 ng/ul
RT: 14.434 min Scan# 1929
Delta R.T. 0.012 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:109 Resp: 359424
Ion Ratio Lower Upper
109 100
139 144.0 106.2 159.4
65 140.4 110.8 166.2



#56
Dibenzofuran
Concen: 44.954 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:168 Resp: 2820336
Ion Ratio Lower Upper
168 100
139 37.2 30.2 45.2









#57
2,4-Dinitrotoluene
Concen: 44.170 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:165 Resp: 707593
Ion Ratio Lower Upper
165 100
63 44.7 35.4 53.0
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 42.934 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:232 Resp: 540045
Ion Ratio Lower Upper
232 100
131 47.1 39.4 59.0
130 2.5 2.0 3.0
166 31.6 26.1 39.1





#59

Diethylphthalate

Concen: 43.856 ng/ul

RT: 15.069 min Scan# 2036

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:149 Resp: 2485558

Ion Ratio Lower Upper

149 100

177 22.2 17.0 25.6

150 12.0 9.8 14.8



#61

Fluorene

Concen: 45.375 ng/ul

RT: 15.269 min Scan# 2071

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:166 Resp: 2292434


Ion Ratio Lower Upper

166 100

165 97.1 77.8 116.6

167 13.6 10.5 15.7





#62

4-Chlorophenyl-phenylether

Concen: 45.214 ng/ul

RT: 15.275 min Scan# 2072

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:204 Resp: 1106404

Ion Ratio Lower Upper

204 100

206 33.4 26.2 39.2

141 60.2 50.4 75.6



#63

4-Nitroaniline

Concen: 48.355 ng/ul

RT: 15.287 min Scan# 2074

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:138 Resp: 537331


Ion Ratio Lower Upper

138 100

92 49.2 41.1 61.7

108 75.6 58.1 87.1









#66
4,6-Dinitro-2-methylphenol
Concen: 41.001 ng/ul
RT: 15.345 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:198 Resp: 367603
Ion Ratio Lower Upper
198 100
182 3.7 3.5 5.3
77 25.1 21.6 32.4



#67
N-Nitrosodiphenylamine
Concen: 43.993 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. 0.006 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:169 Resp: 2034658
Ion Ratio Lower Upper
169 100
168 64.8 54.5 81.7
167 35.7 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 42.665 ng/ul
RT: 16.163 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:248 Resp: 693609
Ion Ratio Lower Upper
248 100
250 91.5 75.8 113.6
141 76.4 63.0 94.6



#69
Hexachlorobenzene
Concen: 43.139 ng/ul
RT: 16.269 min Scan# 2241
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:284 Resp: 794871
Ion Ratio Lower Upper
284 100
142 37.6 32.5 48.7
249 31.6 29.7 44.5









#70
 Atrazine
 Concen: 51.786 ng/ul
 RT: 16.451 min Scan# 2272
 Delta R.T. 0.006 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040213


Tgt Ion:	200	Resp:	756216
Ion Ratio	Lower	Upper	
200	100		
173	27.7	20.6	31.0
215	47.0	36.8	55.2



#71
 Pentachlorophenol
 Concen: 43.128 ng/ul
 RT: 16.610 min Scan# 2299
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Tgt Ion:	266	Resp:	508426
Ion Ratio	Lower	Upper	
266	100		
264	62.5	48.6	72.8
268	61.7	50.5	75.7





#72

Phenanthrene

Concen: 43.201 ng/ul

RT: 17.004 min Scan# 2366

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:178 Resp: 3692578

Ion Ratio Lower Upper

178 100

179 15.5 11.7 17.5

176 19.6 15.1 22.7



#74

Anthracene

Concen: 43.574 ng/ul

RT: 17.092 min Scan# 2381

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:178 Resp: 3773911


Ion Ratio Lower Upper

178 100

179 15.3 11.8 17.6

176 18.0 14.2 21.4









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 43.348 ng/uL
 RT: 12.998 min Scan# 1685
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49




Instrument :
 BNA_N
 ClientSampleId :
 SST040213


Tgt Ion:	216	Resp:	936166
Ion Ratio	Lower	Upper	
216	100		
214	78.5	65.7	98.5
218	46.0	41.8	62.6



#76
 Pentachlorobenzene
 Concen: 42.774 ng/uL
 RT: 14.534 min Scan# 1946
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Tgt Ion:	250	Resp:	899771
Ion Ratio	Lower	Upper	
250	100		
248	70.2	49.8	74.8
252	67.8	49.9	74.9









#77
 Carbazole
 Concen: 48.007 ng/ul
 RT: 17.363 min Scan# 2427
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49




Instrument :
 BNA_N
 ClientSampleId :
 SST040213


Tgt Ion:167 Resp: 3545354
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.0
 139 13.4 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 44.545 ng/ul
 RT: 17.963 min Scan# 2529
 Delta R.T. -0.000 min
 Lab File: BN033369.D
 Acq: 13 Aug 2024 18:49

Tgt Ion:149 Resp: 4525904
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.1 3.8 5.8





#80

Fluoranthene

Concen: 41.533 ng/ul

RT: 19.027 min Scan# 2710

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :





BNA_N

ClientSampleId :

SSTD040213

Tgt Ion:202 Resp: 4346793

Ion	Ratio	Lower	Upper
202	100		
101	12.8	9.9	14.9
100	9.9	7.5	11.3



#82

Pyrene

Concen: 42.987 ng/ul

RT: 19.392 min Scan# 2772




Delta R.T. -0.000 min


Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Tgt Ion:202 Resp: 4616026

Ion	Ratio	Lower	Upper
202	100		
101	14.9	11.4	17.0
100	12.3	9.6	14.4





#83

Butylbenzylphthalate

Concen: 40.160 ng/ul

RT: 20.327 min Scan# 2931

Delta R.T. -0.000 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:149 Resp: 2036810

Ion Ratio Lower Upper

149 100

91 72.7 55.5 83.3

206 20.8 15.8 23.6



#84

3,3'-Dichlorobenzidine

Concen: 50.922 ng/ul

RT: 21.116 min Scan# 3065

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:252 Resp: 1598083


Ion Ratio Lower Upper

252 100

254 63.3 49.8 74.8

126 11.8 10.0 15.0









#85
Benzo(a)anthracene
Concen: 43.785 ng/ul
RT: 21.180 min Scan# 3076
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49




Instrument :
BNA_N
ClientSampleId :
SSTD040213


Tgt Ion:228 Resp: 4796181
Ion Ratio Lower Upper
228 100
229 18.7 16.6 25.0
226 26.8 21.3 31.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.876 ng/ul
RT: 21.151 min Scan# 3071
Delta R.T. -0.000 min
Lab File: BN033369.D
Acq: 13 Aug 2024 18:49

Tgt Ion:149 Resp: 3243818
Ion Ratio Lower Upper
149 100
167 26.9 21.5 32.3
279 4.2 3.1 4.7





#87

Chrysene

Concen: 42.367 ng/ul

RT: 21.239 min Scan# 3085

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:228 Resp: 4334962

Ion Ratio Lower Upper

228 100

226 32.4 23.9 35.9

229 21.7 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 43.391 ng/ul




RT: 22.245 min Scan# 3257


Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Tgt Ion:149 Resp: 5477728





#90

Benzo(b)fluoranthene

Concen: 42.765 ng/ul

RT: 23.145 min Scan# 3409

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:252 Resp: 4820046

Ion Ratio Lower Upper

252 100

253 23.1 18.6 27.8

125 11.9 9.7 14.5



#91

Benzo(k)fluoranthene

Concen: 44.245 ng/ul

RT: 23.210 min Scan# 3421

Delta R.T. 0.012 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:252 Resp: 4893424


Ion Ratio Lower Upper

252 100

253 21.8 18.2 27.4

125 11.1 9.3 13.9





#93

Benzo(a)pyrene

Concen: 43.677 ng/ul

RT: 23.904 min Scan# 3538

Delta R.T. 0.006 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:252 Resp: 4610716

Ion Ratio Lower Upper

252 100

253 22.5 16.6 24.8

125 12.8 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 47.850 ng/ul

RT: 27.198 min Scan# 4099

Delta R.T. 0.023 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49




Tgt Ion:276 Resp: 5820698


Ion Ratio Lower Upper

276 100

138 22.7 17.7 26.5

277 23.7 17.4 26.0





#95

Dibenzo(a,h)anthracene

Concen: 47.887 ng/ul

RT: 27.256 min Scan# 4107

Delta R.T. 0.012 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Instrument :

BNA_N

ClientSampleId :

SSTD040213





Tgt Ion:278 Resp: 4755816

Ion Ratio Lower Upper

278 100

139 17.9 14.6 22.0

279 23.3 18.5 27.7



#96

Benzo(g,h,i)perylene

Concen: 45.929 ng/ul

RT: 28.180 min Scan# 4266

Delta R.T. 0.018 min

Lab File: BN033369.D

Acq: 13 Aug 2024 18:49

Tgt Ion:276 Resp: 4441325

Ion Ratio Lower Upper

276 100

138 23.8 20.5 30.7

277 22.6 18.2 27.4

