

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033371.D
 Acq On : 13 Aug 2024 20:08
 Operator : MA/JU
 Sample : SSTD16079
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Quant Time: Aug 14 01:45:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.569	152	288894	20.000	ng/u1	0.00
20) Naphthalene-d8	10.345	136	1324946	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.216	164	908184	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.969	188	1959287	20.000	ng/u1	0.01
79) Chrysene-d12	21.216	240	1861619	20.000	ng/u1	0.02
88) Perylene-d12	24.051	264	2198042	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	418127	59.877	ng/uL	0.00
4) Pyridine-d5	3.540	84	3335280	169.672	ng/u1	0.00
7) Phenol-d5	6.769	99	4307344	173.919	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.928	67	2492923	172.086	ng/u1	0.01
11) 2-Chlorophenol-d4	7.116	132	3386165	179.189	ng/u1	0.01
15) 4-Methylphenol-d8	8.293	113	3498439	173.290	ng/u1	0.02
21) Nitrobenzene-d5	8.728	128	1733135	171.080	ng/u1	0.02
24) 2-Nitrophenol-d4	9.440	143	1839648	175.795	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	9.975	165	3334595	170.686	ng/u1	0.02
31) 4-Chloroaniline-d4	10.487	131	5007891	172.646	ng/u1	0.02
46) Dimethylphthalate-d6	13.651	166	10093740	159.536	ng/u1	0.02
49) Acenaphthylene-d8	13.910	160	11315915	155.545	ng/u1	0.01
54) 4-Nitrophenol-d4	14.457	143	2198280	173.530	ng/u1	0.05
60) Fluorene-d10	15.222	176	8317457	155.369	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.363	200	1878509	190.744	ng/u1	0.04
73) Anthracene-d10	17.075	188	13116267	156.366	ng/u1	0.02
81) Pyrene-d10	19.374	212	15303075	153.615	ng/u1	0.02
92) Benzo(a)pyrene-d12	23.880	264	18432819	175.623	ng/u1	0.05
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.187	88	442052	59.165	ng/uL	96
5) Pyridine	3.558	79	3371180	171.521	ng/u1	99
6) Benzaldehyde	6.722	77	1486792	127.991	ng/u1	98
8) Phenol	6.799	94	4431964	175.837	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.022	93	3298732	158.277	ng/u1	99
12) 2-Chlorophenol	7.152	128	3403352	172.101	ng/u1	99
13) 2-Methylphenol	8.022	108	3346135	171.885	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.104	45	5087743	212.064	ng/u1	97
16) Acetophenone	8.404	105	5273076	172.348	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.422	70	2693748	169.906	ng/u1	96
18) 4-Methylphenol	8.375	108	3679397	174.357	ng/u1	98
19) Hexachloroethane	8.640	117	1341359	160.179	ng/u1	98
22) Nitrobenzene	8.769	77	3991527	165.374	ng/u1	99
23) Isophorone	9.316	82	7917441	159.910	ng/u1	98
25) 2-Nitrophenol	9.475	139	1983108	173.452	ng/u1	96
26) 2,4-Dimethylphenol	9.546	107	3850481	173.122	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.787	93	4680579	151.683	ng/u1	98
29) 2,4-Dichlorophenol	10.004	162	3349273	173.627	ng/u1	98
30) Naphthalene	10.398	128	10690957	155.910	ng/u1	100
32) 4-Chloroaniline	10.516	127	4846195	172.084	ng/u1	97
33) Hexachlorobutadiene	10.693	225	2002930	172.880	ng/u1	97
34) Caprolactam	11.345	113	702195	93.259	ng/u1	95
35) 4-Chloro-3-methylphenol	11.663	107	3849638	170.734	ng/u1	100
36) 2-Methylnaphthalene	12.016	142	7456221	159.082	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
 Data File : BN033371.D
 Acq On : 13 Aug 2024 20:08
 Operator : MA/JU
 Sample : SSTD16079
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Quant Time: Aug 14 01:45:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 14 01:18:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.240	142	7610263	161.822	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.392	216	3966534	165.145	ng/ul	92
40) Hexachlorocyclopentadiene	12.381	237	2389066	172.583	ng/ul	98
41) 2,4,6-Trichlorophenol	12.640	196	2764585	169.582	ng/ul	96
42) 2,4,5-Trichlorophenol	12.716	196	3014777	170.177	ng/ul	98
43) 1,1'-Biphenyl	13.051	154	9644163	150.973	ng/ul	98
44) 2-Chloronaphthalene	13.087	162	7830632	156.510	ng/ul	99
45) 2-Nitroaniline	13.298	65	2660236	183.424	ng/ul	94
47) Dimethylphthalate	13.704	163	10008943	154.683	ng/ul	99
48) 2,6-Dinitrotoluene	13.816	165	2289089	177.975	ng/ul	94
50) Acenaphthylene	13.945	152	11892886	144.650	ng/ul	98
51) 3-Nitroaniline	14.139	138	2266644	164.826	ng/ul#	95
52) Acenaphthene	14.286	153	8434098	155.036	ng/ul	97
53) 2,4-Dinitrophenol	14.339	184	1388317	213.811	ng/ul	94
55) 4-Nitrophenol	14.475	109	1628790	173.325	ng/ul	96
56) Dibenzofuran	14.628	168	11482109	155.351	ng/ul	96
57) 2,4-Dinitrotoluene	14.604	165	3326243	176.249	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.857	232	2543078	171.615	ng/ul	95
59) Diethylphthalate	15.081	149	10432639	156.250	ng/ul	97
61) Fluorene	15.281	166	8857534	148.820	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.281	204	4397463	152.542	ng/ul	98
63) 4-Nitroaniline	15.328	138	2066803	157.878	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.381	198	2011462	192.128	ng/ul#	91
67) N-Nitrosodiphenylamine	15.498	169	8452319	156.508	ng/ul	99
68) 4-Bromophenyl-phenylether	16.175	248	3189026	167.987	ng/ul	91
69) Hexachlorobenzene	16.280	284	3622160	168.348	ng/ul	90
70) Atrazine	16.475	200	3304422	193.787	ng/ul	96
71) Pentachlorophenol	16.622	266	2509831	182.324	ng/ul	95
72) Phenanthrene	17.016	178	14683614	147.118	ng/ul	93
74) Anthracene	17.110	178	13761064	136.068	ng/ul#	88
75) 1,2,3,4-Tetrachloroben...	13.004	216	4155627	164.785	ng/ul	92
76) Pentachlorobenzene	14.545	250	4193303	170.714	ng/ul	98
77) Carbazole	17.380	167	14210915	164.792	ng/ul	94
78) Di-n-butylphthalate	17.963	149	14410159	121.460	ng/ul#	92
80) Fluoranthene	19.033	202	15260419	128.180	ng/ul#	94
82) Pyrene	19.404	202	15018144	122.948	ng/ul#	90
83) Butylbenzylphthalate	20.333	149	8722209	151.182	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.127	252	6365726	178.316	ng/ul	97
85) Benzo(a)anthracene	21.192	228	17952579	144.074	ng/ul#	89
86) Bis(2-ethylhexyl)phtha...	21.157	149	12314793	146.430	ng/ul#	93
87) Chrysene	21.251	228	2705302	23.243	ng/ul	94
90) Benzo(b)fluoranthene	23.174	252	21417406	163.016	ng/ul	97
91) Benzo(k)fluoranthene	23.239	252	19015496	147.497	ng/ul	100
93) Benzo(a)pyrene	23.957	252	19897183	161.698	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.245	276	10230932	72.152	ng/ul	96
95) Dibenzo(a,h)anthracene	27.333	278	18582924	160.523	ng/ul	97
96) Benzo(g,h,i)perylene	28.250	276	12614908	111.916	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 14 01:45:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 14 01:18:11 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 59.165 ng/uL
 RT: 3.187 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215


Tgt Ion: 88 Resp: 442052
 Ion Ratio Lower Upper
 88 100
 43 31.0 25.4 38.2
 58 80.6 61.0 91.6



#5
 Pyridine
 Concen: 171.521 ng/uL
 RT: 3.558 min Scan# 80
 Delta R.T. -0.000 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Tgt Ion: 79 Resp: 3371180
 Ion Ratio Lower Upper
 79 100
 52 86.8 68.9 103.3
 51 37.7 30.6 45.8









#6
Benzaldehyde
Concen: 127.991 ng/ul
RT: 6.722 min Scan# 618
Delta R.T. -0.000 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 77 Resp: 1486792
Ion Ratio Lower Upper
77 100
105 97.6 79.4 119.0
106 92.3 75.9 113.9



#8
Phenol
Concen: 175.837 ng/ul
RT: 6.799 min Scan# 631
Delta R.T. 0.023 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion: 94 Resp: 4431964
Ion Ratio Lower Upper
94 100
65 28.3 22.1 33.1
66 35.6 28.6 43.0









#10
Bis(2-Chloroethyl)ether
Concen: 158.277 ng/ul
RT: 7.022 min Scan# 60
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 93 Resp: 3298732
Ion Ratio Lower Upper
93 100
63 79.9 63.5 95.3
95 32.6 25.7 38.5



#12
2-Chlorophenol
Concen: 172.101 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:128 Resp: 3403352
Ion Ratio Lower Upper
128 100
64 46.8 37.8 56.8
130 33.2 25.6 38.4





#13

2-Methylphenol

Concen: 171.885 ng/ul

RT: 8.022 min Scan# 81

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :





SSTD160215

Tgt Ion:108 Resp: 3346135

Ion Ratio Lower Upper

108 100

107 88.9 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 212.064 ng/ul

RT: 8.104 min Scan# 853

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion: 45 Resp: 5087743


Ion Ratio Lower Upper

45 100

77 13.5 9.8 14.8

79 9.7 7.1 10.7









#16
Acetophenone
Concen: 172.348 ng/ul
RT: 8.404 min Scan# 904
Delta R.T. 0.023 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:105 Resp: 5273076
Ion Ratio Lower Upper
105 100
77 76.3 61.7 92.5
51 26.8 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 169.906 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. 0.047 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion: 70 Resp: 2693748
Ion Ratio Lower Upper
70 100
42 58.2 49.4 74.2
101 10.2 7.8 11.6
130 22.6 16.4 24.6









#18
4-Methylphenol
Concen: 174.357 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.041 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:108 Resp: 3679397
Ion Ratio Lower Upper
108 100
107 109.3 88.9 133.3



#19
Hexachloroethane
Concen: 160.179 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. -0.000 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:117 Resp: 1341359
Ion Ratio Lower Upper
117 100
201 96.9 75.1 112.7
199 60.7 48.2 72.2









#22
Nitrobenzene
Concen: 165.374 ng/ul
RT: 8.769 min Scan# 90
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion: 77 Resp: 3991527
Ion Ratio Lower Upper
77 100
123 46.4 36.8 55.2
65 13.8 10.7 16.1



#23
Isophorone
Concen: 159.910 ng/ul
RT: 9.316 min Scan# 1059
Delta R.T. 0.035 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion: 82 Resp: 7917441
Ion Ratio Lower Upper
82 100
95 7.1 5.3 7.9
138 18.7 14.2 21.4









#25
2-Nitrophenol
Concen: 173.452 ng/ul
RT: 9.475 min Scan# 1086
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:139 Resp: 1983108
Ion Ratio Lower Upper
139 100
65 49.7 42.4 63.6
109 28.8 24.0 36.0



#26
2,4-Dimethylphenol
Concen: 173.122 ng/ul
RT: 9.546 min Scan# 1098
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:107 Resp: 3850481
Ion Ratio Lower Upper
107 100
121 51.0 39.9 59.9
122 86.9 67.5 101.3





#27

Bis(2-Chloroethoxy)methane

Concen: 151.683 ng/ul

RT: 9.787 min Scan# 1139

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion: 93 Resp: 4680579

Ion Ratio Lower Upper

93 100

95 32.6 24.7 37.1

123 11.5 9.1 13.7



#29

2,4-Dichlorophenol

Concen: 173.627 ng/ul

RT: 10.004 min Scan# 1176

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:162 Resp: 3349273


Ion Ratio Lower Upper

162 100

164 67.4 52.6 78.8

98 36.4 29.8 44.6





#30

Naphthalene

Concen: 155.910 ng/ul

RT: 10.398 min Scan# 1243

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:128 Resp:10690957

Ion Ratio Lower Upper

128 100

129 11.2 8.8 13.2

127 13.2 10.4 15.6



#32

4-Chloroaniline

Concen: 172.084 ng/ul

RT: 10.516 min Scan# 1263

Delta R.T. 0.023 min

Lab File: BN033371.D




Acq: 13 Aug 2024 20:08


Tgt Ion:127 Resp: 4846195

Ion Ratio Lower Upper

127 100

129 32.9 25.2 37.8









#33
Hexachlorobutadiene
Concen: 172.880 ng/ul
RT: 10.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:225 Resp: 2002930
Ion Ratio Lower Upper
225 100
223 60.8 50.1 75.1
227 63.2 48.1 72.1



#34
Caprolactam
Concen: 93.259 ng/ul
RT: 11.345 min Scan# 1404
Delta R.T. 0.100 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:113 Resp: 702195
Ion Ratio Lower Upper
113 100
55 173.1 146.2 219.2
56 127.4 105.1 157.7









#35
 4-Chloro-3-methylphenol
 Concen: 170.734 ng/ul
 RT: 11.663 min Scan# 1458
 Delta R.T. 0.035 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Tgt Ion:107 Resp: 3849638




Ion	Ratio	Lower	Upper
107	100		
144	26.3	21.2	31.8
142	80.3	64.4	96.6




#36
 2-Methylnaphthalene
 Concen: 159.082 ng/ul
 RT: 12.016 min Scan# 1518
 Delta R.T. 0.006 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Tgt Ion:142 Resp: 7456221

Ion	Ratio	Lower	Upper
142	100		
141	87.8	72.2	108.2





#37

1-Methylnaphthalene

Concen: 161.822 ng/ul

RT: 12.240 min Scan# 1554

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:142 Resp: 7610263

Ion Ratio Lower Upper

142 100

141 89.8 73.8 110.8

116 3.9 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 165.145 ng/ul

RT: 12.392 min Scan# 1582

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Tgt Ion:216 Resp: 3966534




Ion Ratio Lower Upper


216 100

214 79.7 56.2 84.2

179 22.0 17.1 25.7

108 17.8 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 172.583 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:237 Resp: 2389066
Ion Ratio Lower Upper
237 100
235 65.5 51.2 76.8
272 12.8 8.6 13.0



#41
2,4,6-Trichlorophenol
Concen: 169.582 ng/ul
RT: 12.640 min Scan# 1624
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:196 Resp: 2764585
Ion Ratio Lower Upper
196 100
198 93.3 78.5 117.7
200 30.8 25.2 37.8





#42

2,4,5-Trichlorophenol

Concen: 170.177 ng/ul

RT: 12.716 min Scan# 1

Delta R.T. 0.023 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:196 Resp: 3014777

Ion Ratio Lower Upper

196 100

198 97.4 76.4 114.6

200 31.2 24.1 36.1



#43

1,1'-Biphenyl

Concen: 150.973 ng/ul

RT: 13.051 min Scan# 1694

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:154 Resp: 9644163


Ion Ratio Lower Upper

154 100

153 41.2 32.0 48.0

76 13.5 11.4 17.2









#44
 2-Chloronaphthalene
 Concen: 156.510 ng/ul
 RT: 13.087 min Scan# 11
 Delta R.T. 0.012 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Tgt Ion:162 Resp: 7830632




Ion	Ratio	Lower	Upper
162	100		
164	33.0	27.0	40.4
127	40.4	31.8	47.8




#45
 2-Nitroaniline
 Concen: 183.424 ng/ul
 RT: 13.298 min Scan# 1736
 Delta R.T. 0.023 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Tgt Ion: 65 Resp: 2660236

Ion	Ratio	Lower	Upper
65	100		
92	71.9	55.6	83.4
138	112.0	83.7	125.5





#47

Dimethylphthalate

Concen: 154.683 ng/ul

RT: 13.704 min Scan# 1805

Delta R.T. 0.023 min

Lab File: BN033371.D


Acq: 13 Aug 2024 20:08

Instrument :

BNA_N




ClientSampleId :

SSTD160215



Tgt Ion:163 Resp:10008943

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.3	4.9
164	10.7	8.4	12.6



#48

2,6-Dinitrotoluene

Concen: 177.975 ng/ul

RT: 13.816 min Scan# 1824




Delta R.T. 0.029 min


Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Tgt Ion:165 Resp: 2289089

Ion	Ratio	Lower	Upper
165	100		
89	53.9	46.8	70.2
121	20.8	18.1	27.1





#50

Acenaphthylene

Concen: 144.650 ng/ul

RT: 13.945 min Scan# 1846

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:152 Resp:11892886

Ion Ratio Lower Upper

152 100

151 21.3 16.2 24.4

153 14.1 10.9 16.3



#51

3-Nitroaniline

Concen: 164.826 ng/ul

RT: 14.139 min Scan# 1879

Delta R.T. 0.029 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:138 Resp: 2266644


Ion Ratio Lower Upper

138 100

108 10.1 10.5 15.7#

92 111.1 93.4 140.2





#52

Acenaphthene

Concen: 155.036 ng/ul

RT: 14.286 min Scan# 1904

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:153 Resp: 8434098

Ion Ratio Lower Upper

153 100

152 48.2 38.0 57.0

154 86.3 71.5 107.3



#53

2,4-Dinitrophenol

Concen: 213.811 ng/ul

RT: 14.339 min Scan# 1913

Delta R.T. 0.023 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:184 Resp: 1388317


Ion Ratio Lower Upper

184 100

63 64.5 55.2 82.8

154 58.5 51.4 77.0









#55
4-Nitrophenol
Concen: 173.325 ng/ul
RT: 14.475 min Scan# 1927
Delta R.T. 0.053 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:109 Resp: 1628790
Ion Ratio Lower Upper
109 100
139 139.9 106.2 159.4
65 137.0 110.8 166.2



#56
Dibenzofuran
Concen: 155.351 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:168 Resp:11482109
Ion Ratio Lower Upper
168 100
139 39.9 30.2 45.2





#57

2,4-Dinitrotoluene

Concen: 176.249 ng/ul

RT: 14.604 min Scan# 1958

Delta R.T. 0.029 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:165 Resp: 3326243

Ion Ratio Lower Upper

165 100

63 43.1 35.4 53.0

182 3.0 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 171.615 ng/ul

RT: 14.857 min Scan# 2001

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Tgt Ion:232 Resp: 2543078




Ion Ratio Lower Upper


232 100

131 44.6 39.4 59.0

130 2.4 2.0 3.0

166 31.3 26.1 39.1









#59
 Diethylphthalate
 Concen: 156.250 ng/ul
 RT: 15.081 min Scan# 2039
 Delta R.T. 0.018 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Tgt Ion:149 Resp:10432639




Ion	Ratio	Lower	Upper
149	100		
177	22.5	17.0	25.6
150	13.5	9.8	14.8




#61
 Fluorene
 Concen: 148.820 ng/ul
 RT: 15.281 min Scan# 2073
 Delta R.T. 0.018 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Tgt Ion:166 Resp: 8857534

Ion	Ratio	Lower	Upper
166	100		
165	95.2	77.8	116.6
167	13.1	10.5	15.7









#62
4-Chlorophenyl-phenylether
Concen: 152.542 ng/ul
RT: 15.281 min Scan# 2073
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:204 Resp: 4397463
Ion Ratio Lower Upper
204 100
206 32.3 26.2 39.2
141 60.4 50.4 75.6



#63
4-Nitroaniline
Concen: 157.878 ng/ul
RT: 15.328 min Scan# 2081
Delta R.T. 0.047 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:138 Resp: 2066803
Ion Ratio Lower Upper
138 100
92 50.9 41.1 61.7
108 74.9 58.1 87.1









#66
4,6-Dinitro-2-methylphenol
Concen: 192.128 ng/ul
RT: 15.381 min Scan# 2083
Delta R.T. 0.041 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:198 Resp: 2011462
Ion Ratio Lower Upper
198 100
182 3.0 3.5 5.3#
77 21.8 21.6 32.4



#67
N-Nitrosodiphenylamine
Concen: 156.508 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:169 Resp: 8452319
Ion Ratio Lower Upper
169 100
168 67.6 54.5 81.7
167 37.6 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 167.987 ng/ul
RT: 16.175 min Scan# 21
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:248 Resp: 3189026
Ion Ratio Lower Upper
248 100
250 88.4 75.8 113.6
141 69.2 63.0 94.6



#69
Hexachlorobenzene
Concen: 168.348 ng/ul
RT: 16.280 min Scan# 2243
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:284 Resp: 3622160
Ion Ratio Lower Upper
284 100
142 33.8 32.5 48.7
249 32.1 29.7 44.5









#70
Atrazine
Concen: 193.787 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. 0.029 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:200 Resp: 3304422
Ion Ratio Lower Upper
200 100
173 27.0 20.6 31.0
215 48.9 36.8 55.2



#71
Pentachlorophenol
Concen: 182.324 ng/ul
RT: 16.622 min Scan# 2301
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:266 Resp: 2509831
Ion Ratio Lower Upper
266 100
264 65.4 48.6 72.8
268 66.7 50.5 75.7





#72

Phenanthrene

Concen: 147.118 ng/ul

RT: 17.016 min Scan# 21

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:178 Resp:14683614

Ion Ratio Lower Upper

178 100

179 17.5 11.7 17.5

176 21.9 15.1 22.7



#74

Anthracene

Concen: 136.068 ng/ul

RT: 17.110 min Scan# 2384

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:178 Resp:13761064


Ion Ratio Lower Upper

178 100

179 18.7 11.8 17.6#

176 23.6 14.2 21.4#









#75
1,2,3,4-Tetrachlorobenzene
Concen: 164.785 ng/uL
RT: 13.004 min Scan# 1686
Delta R.T. 0.006 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:216 Resp: 4155627
Ion Ratio Lower Upper
216 100
214 75.8 65.7 98.5
218 46.2 41.8 62.6



#76
Pentachlorobenzene
Concen: 170.714 ng/uL
RT: 14.545 min Scan# 1948
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:250 Resp: 4193303
Ion Ratio Lower Upper
250 100
248 65.8 49.8 74.8
252 62.6 49.9 74.9









#77
Carbazole
Concen: 164.792 ng/ul
RT: 17.380 min Scan# 2427
Delta R.T. 0.018 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:167 Resp:14210915
Ion Ratio Lower Upper
167 100
166 24.8 17.4 26.0
139 15.4 10.3 15.5



#78
Di-n-butylphthalate
Concen: 121.460 ng/ul
RT: 17.963 min Scan# 2529
Delta R.T. -0.000 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:149 Resp:14410159
Ion Ratio Lower Upper
149 100
150 11.9 7.1 10.7#
104 6.8 3.8 5.8#









#80
Fluoranthene
Concen: 128.180 ng/ul
RT: 19.033 min Scan# 2710
Delta R.T. 0.006 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08




Instrument :
BNA_N
ClientSampleId :
SSTD160215


Tgt Ion:202 Resp:15260419
Ion Ratio Lower Upper
202 100
101 14.8 9.9 14.9
100 11.5 7.5 11.3#



#82
Pyrene
Concen: 122.948 ng/ul
RT: 19.404 min Scan# 2774
Delta R.T. 0.012 min
Lab File: BN033371.D
Acq: 13 Aug 2024 20:08

Tgt Ion:202 Resp:15018144
Ion Ratio Lower Upper
202 100
101 18.8 11.4 17.0#
100 15.3 9.6 14.4#





#83

Butylbenzylphthalate

Concen: 151.182 ng/ul

RT: 20.333 min Scan# 2931

Delta R.T. 0.006 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:149 Resp: 8722209

Ion Ratio Lower Upper

149 100

91 72.0 55.5 83.3

206 22.8 15.8 23.6



#84

3,3'-Dichlorobenzidine

Concen: 178.316 ng/ul

RT: 21.127 min Scan# 3067

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:252 Resp: 6365726


Ion Ratio Lower Upper

252 100

254 65.2 49.8 74.8

126 12.4 10.0 15.0





#85

Benzo(a)anthracene

Concen: 144.074 ng/ul

RT: 21.192 min Scan# 3077

Delta R.T. 0.012 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:228 Resp:17952579

Ion Ratio Lower Upper

228 100

229 24.4 16.6 25.0

226 33.7 21.3 31.9#



#86

Bis(2-ethylhexyl)phthalate

Concen: 146.430 ng/ul

RT: 21.157 min Scan# 3072

Delta R.T. 0.006 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:149 Resp:12314793


Ion Ratio Lower Upper

149 100

167 30.8 21.5 32.3

279 5.9 3.1 4.7#





#87

Chrysene

Concen: 23.243 ng/ul

RT: 21.251 min Scan# 3085

Delta R.T. 0.018 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:228 Resp: 2705302

Ion Ratio Lower Upper

228 100

226 33.7 23.9 35.9

229 21.9 15.9 23.9



#90

Benzo(b)fluoranthene

Concen: 163.016 ng/ul

RT: 23.174 min Scan# 3415

Delta R.T. 0.035 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:252 Resp:21417406


Ion Ratio Lower Upper

252 100

253 21.5 18.6 27.8

125 11.3 9.7 14.5





#91

Benzo(k)fluoranthene

Concen: 147.497 ng/ul

RT: 23.239 min Scan# 3419

Delta R.T. 0.041 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:252 Resp:19015496

Ion Ratio Lower Upper

252 100

253 22.9 18.2 27.4

125 11.5 9.3 13.9



#93

Benzo(a)pyrene

Concen: 161.698 ng/ul

RT: 23.957 min Scan# 3548

Delta R.T. 0.059 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:252 Resp:19897183


Ion Ratio Lower Upper

252 100

253 22.6 16.6 24.8

125 12.3 9.6 14.4





#94

Indeno(1,2,3-cd)pyrene

Concen: 72.152 ng/ul

RT: 27.245 min Scan# 4107

Delta R.T. 0.070 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08

Instrument :

BNA_N

ClientSampleId :

SSTD160215





Tgt Ion:276 Resp:10230932

Ion Ratio Lower Upper

276 100

138 22.3 17.7 26.5

277 25.2 17.4 26.0



#95

Dibenzo(a,h)anthracene

Concen: 160.523 ng/ul

RT: 27.333 min Scan# 4122

Delta R.T. 0.088 min

Lab File: BN033371.D

Acq: 13 Aug 2024 20:08




Tgt Ion:278 Resp:18582924


Ion Ratio Lower Upper

278 100

139 17.4 14.6 22.0

279 24.6 18.5 27.7





#96
 Benzo(g,h,i)perylene
 Concen: 111.916 ng/ul
 RT: 28.250 min Scan# 41
 Delta R.T. 0.088 min
 Lab File: BN033371.D
 Acq: 13 Aug 2024 20:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTD160215

Tgt Ion:276 Resp:12614908
 Ion Ratio Lower Upper
 276 100
 138 23.9 20.5 30.7
 277 23.6 18.2 27.4

