



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081424\
Data File : BN033373.D
Acq On : 14 Aug 2024 08:21
Operator : MA/JU
Sample : DFTPP351
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP351

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081424.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Aug 14 02:12:11 2024



AutoFind: Scans 2439, 2440, 2441; Background Corrected with Scan 2433

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	245255	PASS
68	69	0.00	2	1.6	4062	PASS
69	198	0.00	100	39.1	250517	PASS
70	69	0.00	2	0.5	1209	PASS
127	198	10	80	48.6	311957	PASS
197	198	0.00	2	0.5	3099	PASS
198	198	100	100	100.0	641365	PASS
199	198	5	9	6.7	42813	PASS
275	198	10	60	26.2	168251	PASS
365	198	1	100	3.4	21708	PASS
441	198	0.01	100	10.1	64781	PASS
442	198	50	100	67.5	433088	PASS
443	442	15	24	19.7	85285	PASS

m/z	Abundance
39.00	186.0
40.00	2096.0
41.00	351.0
43.00	425.0
44.00	9575.0
44.90	337.0
47.00	183.0
49.00	456.0
50.90	249.0
54.80	283.0
57.00	221.0
70.20	159.0
72.90	447.0
75.00	249.0
77.00	309.0
83.90	426.0
85.90	265.0
95.90	599.0
118.90	197.0
133.00	401.0
146.80	187.0
177.00	171.0
190.90	367.0
192.10	176.0
192.90	333.0
207.00	4791.0
208.00	770.0
209.00	684.0
281.10	551.0
282.00	166.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
36.00	375.0
37.00	1011.0
38.00	3098.0
39.00	17904.0
40.00	2654.0
41.00	742.0
41.90	167.0
43.00	316.0
44.00	8418.0
44.90	589.0
47.00	256.0
47.90	168.0
49.00	2028.0
50.00	62104.0
51.00	238656.0
52.00	12599.0
53.00	673.0
55.00	1113.0
56.00	6959.0
57.00	16104.0
58.00	762.0
58.90	252.0
59.90	201.0
61.00	2814.0
62.00	3270.0
63.00	9201.0
64.00	1148.0
65.00	4645.0
65.90	471.0
67.00	361.0
68.00	4157.0
69.00	237056.0
70.00	1243.0
71.00	252.0
72.00	153.0
73.00	2278.0
74.00	23880.0
75.00	38560.0
76.00	12754.0
77.00	265536.0
78.00	19408.0
79.00	17400.0
80.00	13011.0
81.00	18864.0
81.90	4885.0
83.00	4733.0
84.00	763.0
85.00	2695.0
86.00	5430.0
87.00	2606.0
87.90	922.0
89.00	714.0
91.00	4194.0
92.00	5090.0
93.00	30640.0
94.00	2337.0
95.00	452.0
96.00	1641.0
97.10	678.0
98.00	21936.0
99.00	20288.0
100.00	1835.0
101.00	10736.0
102.00	532.0
102.90	3712.0
104.00	6368.0
105.00	6226.0
106.00	1762.0
107.00	77888.0
108.00	12602.0
109.00	2773.0
110.00	139584.0
111.00	23072.0
111.90	2842.0
112.90	869.0
114.80	522.0
116.00	4256.0
117.00	63656.0
118.00	5208.0
118.90	578.0
120.00	1023.0
121.10	415.0
122.00	4806.0
123.00	8429.0
124.00	3669.0
125.00	3336.0
126.00	1354.0
127.00	275776.0
128.00	20584.0
129.00	105504.0
130.00	9525.0
131.00	1938.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

132.00	910.0
133.00	945.0
134.00	3220.0
135.00	8290.0
136.00	3257.0
137.00	5410.0
137.90	670.0
138.90	606.0
140.00	1469.0
141.00	13202.0
142.00	4459.0
143.00	2552.0
144.00	876.0
145.00	703.0
145.90	2293.0
147.00	6869.0
148.00	16672.0
148.90	3327.0
150.10	896.0
151.20	1609.0
151.70	1038.0
153.00	4899.0
154.00	3457.0
155.00	7379.0
156.00	11180.0
157.10	1994.0
158.00	2219.0
159.00	1892.0
160.00	4260.0
161.00	6476.0
161.90	2130.0
163.00	741.0
164.00	855.0
165.00	4858.0
166.00	4051.0
167.00	27632.0
168.00	17536.0
169.00	2730.0
170.00	799.0
170.90	1223.0
171.90	2814.0
173.00	3573.0
174.00	5625.0
175.00	10723.0
176.00	3192.0
176.90	4406.0
178.00	1646.0
178.90	21128.0
180.00	14919.0
181.00	6953.0
181.90	1026.0
182.90	630.0
184.00	1719.0
185.00	9201.0
186.00	70928.0
187.00	20784.0
188.00	2562.0
189.00	4210.0
190.00	680.0
191.00	2731.0
192.00	6772.0
193.00	7165.0
194.00	1741.0
194.90	888.0
196.00	14958.0
197.00	2215.0
198.00	538368.0
198.90	36800.0
200.00	3162.0
201.40	1998.0
203.00	3770.0
204.00	18752.0
205.00	31920.0
206.00	138560.0
207.00	23216.0
208.00	5320.0
209.00	1889.0
210.10	2297.0
211.00	5500.0
213.00	500.0
213.80	221.0
215.00	1443.0
216.00	2811.0
217.00	35000.0
218.00	4716.0
219.00	439.0
221.00	21088.0
223.00	8273.0
224.00	73200.0
225.00	19896.0
226.00	2065.0
227.00	31432.0
228.00	4875.0
229.00	6957.0
230.00	1095.0
231.00	2669.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

232.00	605.0
233.10	509.0
233.90	1992.0
235.00	2109.0
236.00	1514.0
237.00	3043.0
238.00	381.0
239.00	946.0
239.90	1001.0
241.00	1676.0
242.00	3905.0
243.00	4029.0
244.00	57528.0
245.10	7934.0
246.00	11399.0
247.00	2055.0
248.00	516.0
249.00	2030.0
249.90	339.0
251.00	549.0
252.00	595.0
253.00	1068.0
254.10	2082.0
255.00	292672.0
256.00	41312.0
257.00	3249.0
258.00	16416.0
259.00	2842.0
259.90	404.0
260.90	534.0
263.90	382.0
264.90	6458.0
265.90	1381.0
267.10	238.0
269.80	218.0
270.90	893.0
272.00	970.0
273.00	8443.0
274.00	21800.0
275.00	128496.0
276.00	17408.0
277.00	11455.0
277.90	1960.0
278.90	430.0
280.90	621.0
281.90	480.0
283.00	1419.0
284.00	908.0
285.00	2016.0
286.00	415.0
288.90	465.0
290.00	479.0
290.80	266.0
292.00	568.0
292.90	2374.0
293.90	491.0
295.00	918.0
295.90	35296.0
297.00	5059.0
297.90	411.0
301.00	409.0
302.00	514.0
303.00	3943.0
304.00	1115.0
305.00	247.0
308.00	312.0
309.00	368.0
309.90	402.0
312.90	280.0
314.00	1570.0
315.00	3488.0
316.00	2012.0
317.00	357.0
320.10	172.0
320.90	1003.0
322.00	481.0
323.00	10107.0
324.00	2131.0
324.90	262.0
325.90	224.0
327.00	2443.0
327.90	1210.0
329.10	299.0
332.00	874.0
333.00	1191.0
334.00	7013.0
335.00	1797.0
335.90	218.0
338.90	160.0
340.00	196.0
341.00	1317.0
342.00	379.0
346.00	2380.0
346.90	427.0
351.00	211.0
352.00	3381.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

353.10	2163.0
354.00	3425.0
355.00	741.0
359.10	328.0
365.00	15445.0
366.00	2026.0
369.90	156.0
371.00	697.0
372.00	5516.0
373.00	1218.0
373.90	275.0
382.90	1545.0
383.90	560.0
389.90	732.0
391.00	610.0
391.90	410.0
400.90	324.0
402.00	2377.0
403.00	3164.0
404.00	1136.0
404.90	320.0
420.90	2337.0
422.00	2545.0
423.00	17192.0
424.00	4679.0
425.00	607.0
440.00	237.0
441.00	43080.0
442.00	264384.0
443.00	51808.0
444.00	5519.0
445.10	238.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
35.90	226.0
37.00	1275.0
38.10	3167.0
39.00	21152.0
40.00	2605.0
41.10	765.0
41.90	170.0
43.00	417.0
44.00	8690.0
45.00	769.0
46.90	222.0
48.00	155.0
49.00	2228.0
50.00	74776.0
51.00	284224.0
52.00	14804.0
52.90	744.0
55.00	1510.0
56.00	9028.0
57.00	19328.0
57.90	970.0
59.00	377.0
60.00	175.0
61.00	3705.0
62.00	4011.0
63.00	11172.0
64.00	1806.0
65.00	5860.0
65.90	460.0
67.00	427.0
68.00	4637.0
69.00	296576.0
70.00	1618.0
70.90	215.0
71.80	167.0
73.00	2598.0
74.00	29088.0
75.00	46200.0
76.00	15678.0
77.00	332352.0
78.00	23376.0
79.00	21336.0
80.00	16776.0
81.00	24008.0
82.00	5560.0
83.00	5547.0
83.90	1005.0
85.00	3420.0
86.00	6792.0
87.00	3321.0
88.00	1200.0
88.90	573.0
90.00	161.0
90.90	5033.0
92.00	6109.0
93.00	37168.0
94.00	2254.0
94.90	597.0
96.00	2065.0
97.00	811.0
98.00	30064.0
99.00	24792.0
100.00	1854.0
101.00	12799.0
102.00	707.0
103.00	4374.0
104.00	8340.0
105.00	7207.0
106.00	2986.0
107.00	103440.0
108.00	15583.0
109.00	2944.0
110.00	188480.0
111.00	27720.0
112.00	3238.0
113.00	980.0
113.90	307.0
115.00	552.0
116.00	5944.0
117.00	84624.0
118.00	7015.0
119.00	983.0
120.00	1272.0
120.90	450.0
122.00	6442.0
123.00	10157.0
124.00	4635.0
124.90	4260.0
126.10	1514.0
127.00	365632.0
128.00	27288.0
129.00	139904.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

130.00	12054.0
130.90	2318.0
132.00	1395.0
133.00	819.0
134.00	3839.0
135.00	10982.0
136.00	4757.0
137.00	6787.0
137.90	1312.0
138.90	722.0
139.90	1767.0
141.00	18512.0
142.00	6243.0
143.00	4172.0
144.00	1163.0
144.90	993.0
146.00	3077.0
147.00	9258.0
148.00	21528.0
149.00	4864.0
149.90	1217.0
151.10	2689.0
151.80	1649.0
153.00	6086.0
154.00	4415.0
155.00	10217.0
156.00	14987.0
157.00	3097.0
157.90	3373.0
159.00	2860.0
160.00	5701.0
161.00	8618.0
162.00	2163.0
163.00	760.0
164.00	827.0
164.90	7005.0
166.00	5786.0
167.00	36928.0
168.00	21760.0
169.00	3114.0
170.00	1136.0
170.90	1513.0
172.00	3645.0
173.00	4508.0
174.00	7763.0
175.00	14800.0
176.00	3761.0
177.00	6666.0
178.00	2319.0
178.90	27600.0
180.00	19448.0
181.00	8761.0
182.00	1387.0
182.90	739.0
184.00	2307.0
185.00	13273.0
186.00	99352.0
187.00	30192.0
188.00	3031.0
189.00	6425.0
189.90	1269.0
191.00	3202.0
192.00	9487.0
193.00	10067.0
194.00	2352.0
195.00	1472.0
196.00	20608.0
196.80	4183.0
198.00	751936.0
199.00	50840.0
200.00	3597.0
201.40	3311.0
203.00	5727.0
204.00	27048.0
205.00	46664.0
206.00	202880.0
207.00	32576.0
208.00	7530.0
209.00	2985.0
210.00	2758.0
211.00	7990.0
213.00	494.0
214.00	283.0
215.00	2245.0
216.00	4067.0
217.00	52552.0
217.90	6821.0
219.00	647.0
221.00	29592.0
223.00	10914.0
224.00	110872.0
225.00	28360.0
226.00	3182.0
227.00	46736.0
228.00	7152.0
229.00	9602.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

229.90	1338.0
231.00	3896.0
232.00	921.0
232.90	1051.0
234.00	2706.0
235.00	3547.0
235.90	2178.0
237.00	3996.0
238.00	589.0
239.00	1549.0
240.00	1450.0
240.90	2498.0
242.00	5719.0
243.00	6040.0
244.00	85176.0
245.00	11858.0
246.00	16752.0
247.00	3478.0
247.90	710.0
249.00	3140.0
250.00	630.0
251.00	795.0
252.00	773.0
252.90	1612.0
254.10	2896.0
255.00	426368.0
256.00	64704.0
257.00	5791.0
258.00	24128.0
259.00	3847.0
260.00	731.0
261.00	668.0
262.10	215.0
262.90	332.0
263.90	598.0
265.00	9395.0
266.00	1706.0
267.00	220.0
269.90	526.0
270.90	1475.0
272.00	1402.0
273.00	12775.0
274.00	34320.0
275.00	201792.0
276.00	26392.0
277.00	14851.0
278.00	2720.0
278.90	550.0
281.00	753.0
282.00	585.0
282.90	2050.0
284.00	1381.0
285.00	2679.0
286.00	519.0
287.90	160.0
288.90	676.0
289.90	447.0
290.90	608.0
292.00	910.0
293.00	3621.0
294.00	1118.0
295.10	1273.0
296.00	55968.0
297.00	8049.0
297.90	465.0
299.10	164.0
301.00	745.0
302.00	860.0
303.00	6086.0
304.00	1563.0
305.10	308.0
307.90	713.0
308.90	354.0
310.10	643.0
311.00	152.0
311.90	196.0
312.90	454.0
314.00	2450.0
314.90	6268.0
316.00	3533.0
316.90	847.0
320.00	259.0
321.00	1710.0
322.00	989.0
323.00	18456.0
324.00	3103.0
325.00	385.0
326.00	314.0
326.90	3135.0
328.00	1944.0
329.10	285.0
332.00	1236.0
333.00	2000.0
334.00	11844.0
335.00	3104.0
335.90	216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

338.90	372.0
339.90	225.0
341.00	2138.0
342.00	479.0
342.90	185.0
345.90	3613.0
346.80	769.0
351.00	275.0
352.00	5243.0
353.00	3975.0
354.00	5808.0
355.00	1074.0
358.90	293.0
363.80	199.0
365.00	26016.0
366.00	3905.0
367.00	293.0
370.00	512.0
371.00	1347.0
372.00	10609.0
373.00	2510.0
374.00	162.0
376.90	358.0
383.00	2658.0
383.90	668.0
384.90	311.0
390.00	1134.0
390.90	876.0
392.00	665.0
401.00	576.0
401.90	3763.0
403.00	5247.0
404.00	2044.0
405.00	491.0
410.00	208.0
420.20	164.0
420.90	3969.0
421.90	4239.0
423.00	33144.0
424.00	7728.0
424.90	798.0
441.00	69128.0
442.00	491904.0
443.10	103192.0
444.00	9296.0
445.10	701.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
35.90	293.0
37.00	1015.0
38.00	2720.0
39.00	15381.0
40.00	2588.0
41.00	630.0
43.10	355.0
44.00	8508.0
45.00	776.0
47.90	271.0
49.00	1751.0
50.00	55848.0
51.00	213632.0
52.00	10320.0
53.00	470.0
54.00	151.0
55.00	1098.0
56.00	6528.0
57.00	14781.0
57.90	617.0
58.90	328.0
59.90	207.0
61.00	3007.0
62.00	2917.0
63.00	8092.0
64.00	1209.0
65.00	4074.0
66.00	377.0
67.10	320.0
68.00	3392.0
68.90	217920.0
70.00	1242.0
70.90	168.0
73.00	2025.0
74.00	20744.0
75.00	33928.0
76.00	12004.0
77.00	249728.0
78.00	16992.0
79.00	16048.0
80.00	12780.0
81.00	18984.0
82.00	4324.0
83.00	4588.0
84.00	602.0
85.00	2633.0
86.00	5014.0
87.00	2080.0
88.00	1064.0
88.90	398.0
91.00	3865.0
92.00	4807.0
93.00	29192.0
94.00	2215.0
95.00	537.0
96.00	1553.0
97.00	476.0
98.00	21360.0
99.00	17584.0
100.00	1839.0
100.90	10217.0
102.00	667.0
103.00	3409.0
104.00	6474.0
105.00	6242.0
106.00	1707.0
107.00	78560.0
108.00	12359.0
109.00	1919.0
110.00	142464.0
111.00	21568.0
112.00	2599.0
113.00	881.0
113.90	164.0
115.00	371.0
116.00	4523.0
117.00	66536.0
118.00	4729.0
119.00	553.0
120.00	997.0
121.00	498.0
122.00	5439.0
123.00	8192.0
124.00	3707.0
125.00	3314.0
126.00	1039.0
127.00	294464.0
128.00	22792.0
129.00	114864.0
130.00	8355.0
131.00	2153.0
131.90	815.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

132.90	738.0
134.00	3089.0
135.00	8828.0
136.00	3640.0
137.00	4649.0
138.00	963.0
139.00	730.0
139.90	1773.0
141.00	14823.0
142.00	4859.0
143.00	3161.0
143.90	906.0
144.90	891.0
145.90	2706.0
147.00	7490.0
148.00	17888.0
149.00	3833.0
150.00	926.0
151.10	2015.0
151.80	1210.0
153.00	4383.0
153.90	3576.0
155.00	7807.0
156.00	12624.0
157.10	2555.0
158.00	2658.0
159.00	2327.0
160.00	4346.0
161.00	6591.0
161.90	1936.0
162.90	627.0
163.90	766.0
165.00	5820.0
166.00	4304.0
167.00	30944.0
168.00	16592.0
169.00	2552.0
169.90	898.0
170.90	1595.0
172.00	2948.0
173.00	3374.0
174.00	5882.0
175.00	12450.0
176.00	3258.0
176.90	5797.0
178.00	1991.0
179.00	21584.0
180.00	15825.0
181.00	7785.0
182.00	1259.0
182.80	634.0
184.00	1821.0
185.00	10368.0
186.00	82936.0
187.00	24616.0
188.00	2219.0
189.00	5984.0
190.00	959.0
191.00	2817.0
192.00	7321.0
193.00	8634.0
194.00	1741.0
195.00	811.0
196.00	19208.0
196.90	2898.0
198.00	633792.0
199.00	40800.0
200.00	3102.0
201.40	2819.0
203.00	4767.0
204.00	21584.0
205.00	39256.0
206.00	166976.0
207.00	26184.0
208.00	6472.0
209.00	2815.0
210.20	2420.0
211.00	6720.0
213.00	521.0
213.80	248.0
214.90	2119.0
216.00	3519.0
217.00	47352.0
218.00	5600.0
219.00	473.0
221.00	25312.0
223.00	9993.0
224.00	95488.0
225.00	24448.0
226.00	2491.0
227.00	40976.0
228.00	5080.0
229.00	9058.0
230.00	1115.0
231.00	3237.0
231.90	576.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

233.00	737.0
234.00	2635.0
234.90	2724.0
235.90	1802.0
237.00	3352.0
238.00	637.0
239.00	1567.0
239.90	1308.0
241.00	2321.0
242.00	5185.0
243.00	5605.0
244.00	71600.0
245.00	10247.0
246.00	15596.0
246.90	2909.0
248.00	693.0
249.00	3187.0
249.90	527.0
250.90	629.0
252.00	713.0
253.00	1429.0
254.00	2570.0
255.00	376256.0
256.00	56664.0
257.00	4136.0
258.00	21216.0
259.00	3618.0
259.90	741.0
260.80	679.0
263.00	350.0
263.90	552.0
265.00	8661.0
266.00	1087.0
266.80	280.0
268.00	187.0
269.90	397.0
271.00	869.0
271.90	1229.0
273.00	12373.0
274.00	29136.0
275.00	174464.0
276.00	25328.0
277.00	15525.0
278.00	2329.0
279.00	571.0
281.00	754.0
282.20	582.0
283.00	2133.0
284.00	907.0
285.00	2529.0
286.00	530.0
287.90	216.0
288.90	633.0
289.90	470.0
290.90	289.0
292.00	638.0
292.90	3242.0
293.90	949.0
295.00	1105.0
296.00	49376.0
297.00	6951.0
298.10	450.0
300.90	659.0
302.00	784.0
303.00	5759.0
304.00	1410.0
308.00	762.0
309.00	570.0
309.90	614.0
313.00	340.0
314.00	2661.0
315.00	5445.0
316.00	3296.0
317.00	572.0
320.00	221.0
321.00	1813.0
321.90	913.0
323.00	17272.0
324.00	2999.0
324.90	330.0
326.00	396.0
326.90	3414.0
328.00	1605.0
329.00	317.0
329.90	239.0
332.00	1185.0
333.00	1758.0
334.00	10296.0
335.00	2886.0
336.00	296.0
339.00	217.0
340.10	169.0
341.00	2158.0
341.90	499.0
342.90	184.0
346.00	3561.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

346.90	595.0
350.00	206.0
350.90	312.0
352.00	4924.0
353.00	3150.0
354.00	5739.0
355.00	1069.0
358.90	361.0
364.90	23664.0
366.00	3411.0
366.90	425.0
370.00	528.0
370.90	1324.0
372.00	9560.0
373.00	2284.0
374.10	316.0
377.00	199.0
383.00	2783.0
384.00	785.0
384.90	227.0
390.00	1083.0
391.00	734.0
392.10	575.0
392.90	157.0
400.80	550.0
402.00	3679.0
403.00	5664.0
404.00	2192.0
404.90	365.0
415.00	244.0
421.00	4162.0
422.00	4691.0
423.00	34168.0
424.00	7967.0
424.90	901.0
426.00	155.0
441.00	82136.0
442.00	542976.0
443.00	100856.0
444.00	9174.0
444.90	570.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351