

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081518\
 Data File : BN002341.D
 Acq On : 15 Aug 2018 18:52
 Operator : JU/SJ
 Sample : PB112091BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK91

Quant Time: Aug 16 01:51:25 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 16 01:50:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	169291	20.00	ng/ul	0.00
18) Naphthalene-d8	10.45	136	782969	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.31	164	495771	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	1162757	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	1281765	20.00	ng/ul	0.00
85) Perylene-d12	23.49	264	1381903	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.24	96	16330	4.55	ng/uL	0.00
5) Phenol-d5	6.86	99	322581	20.78	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.02	67	228057	23.39	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	261400	22.11	ng/ul	0.00
13) 4-Methylphenol-d8	8.38	113	279427	22.52	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	151947	24.38	ng/ul	0.00
22) 2-Nitrophenol-d4	9.54	143	161747	24.68	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.08	165	264598	21.97	ng/ul	0.00
29) 4-Chloroaniline-d4	10.59	131	450385	34.59	ng/ul	0.00
43) Dimethylphthalate-d6	13.72	166	1202332	29.25	ng/ul	0.00
46) Acenaphthylene-d8	14.00	160	1391542	27.60	ng/ul	0.00
51) 4-Nitrophenol-d4	14.52	143	179889	23.16	ng/ul	0.00
57) Fluorene-d10	15.30	176	992581	27.86	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.44	200	165312	22.33	ng/ul	0.00
70) Anthracene-d10	17.16	188	1617287	29.12	ng/ul	0.00
78) Pyrene-d10	19.46	212	1816404	29.62	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.35	264	1992851	27.58	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN081518\
Data File : BN002341.D
Acq On : 15 Aug 2018 18:52
Operator : JU/SJ
Sample : PB112091BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK91

Quant Time: Aug 16 01:51:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN081318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 16 01:50:44 2018
Response via : Initial Calibration

