

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081523\
 Data File : BN026894.D
 Acq On : 15 Aug 2023 10:44
 Operator : MA/JU
 Sample : PB154774BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK774

Quant Time: Aug 16 00:53:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 00:50:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	9161	0.400	ng/ul	0.00
4) Naphthalene-d8	10.938	136	30517	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.735	164	17869	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.481	188	40077	0.400	ng/ul	0.00
17) Chrysene-d12	21.659	240	29723	0.400	ng/ul	0.00
23) Perylene-d12	24.211	264	36548	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.439	96	55283	5.083	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.511	152	12685	0.282	ng/ul	0.00
18) Fluoranthene-d10	19.497	212	31597	0.337	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.725	166	4793	0.061	ng/ul#	13
20) Pyrene	19.868	202	4736	0.036	ng/ul#	1
26) Benzo(a)pyrene	24.044	252	11232	0.090	ng/ul#	65


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081523\
Data File : BN026894.D
Acq On : 15 Aug 2023 10:44
Operator : MA/JU
Sample : PB154774BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK774

Quant Time: Aug 16 00:53:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 00:50:05 2023
Response via : Initial Calibration









#12
Fluorene
Concen: 0.061 ng/ul
RT: 15.725 min Scan# 21
Delta R.T. -0.055 min
Lab File: BN026894.D
Acq: 15 Aug 2023 10:44




Instrument :
BNA_N
ClientSampleId :
SBLK774


Tgt Ion:166 Resp: 4793
Ion Ratio Lower Upper
166 100
165 6.9 79.1 118.7#
167 34.0 11.4 17.0#



#20
Pyrene
Concen: 0.036 ng/ul
RT: 19.868 min Scan# 3757
Delta R.T. -0.028 min
Lab File: BN026894.D
Acq: 15 Aug 2023 10:44


Tgt Ion:202 Resp: 4736
Ion Ratio Lower Upper
202 100
101 143.2 10.7 16.1#
100 1032.4 8.5 12.7#





#26
Benzo(a)pyrene
Concen: 0.090 ng/ul
RT: 24.044 min Scan# 51
Delta R.T. -0.055 min
Lab File: BN026894.D
Acq: 15 Aug 2023 10:44

Instrument :
BNA_N
ClientSampleId :
SBLK774



Tgt Ion:252 Resp: 11232
Ion Ratio Lower Upper
252 100
253 35.1 19.0 28.6#
125 37.0 11.4 17.0#

