

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081619\
 Data File : BN007219.D
 Acq On : 16 Aug 2019 12:12
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 16 23:03:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN081019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 12:12:02 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	-0.02
2	n-Nitrosodimethylamine	0.400	0.413	-3.2	128	-0.02
3 S	2-Fluorophenol	0.400	0.425	-6.2	108	0.00
4 S	Phenol-d6	0.400	0.420	-5.0	110	0.00
5	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	107	-0.02
6 I	Naphthalene-d8	0.400	0.400	0.0	101	-0.01
7 S	Nitrobenzene-d5	0.400	0.457	-14.2	127	-0.01
8	Naphthalene	0.400	0.391	2.3	106	-0.01
9	Hexachlorobutadiene	0.400	0.398	0.5	106	-0.03
10 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	106	-0.02
11	2-Methylnaphthalene	0.400	0.390	2.5	107	-0.02
12 I	Acenaphthene-d10	0.400	0.400	0.0	103	-0.01
13 S	2,4,6-Tribromophenol	0.400	0.387	3.3	136	0.00
14 S	2-Fluorobiphenyl	0.400	0.057	85.8#	13	0.00
15	Acenaphthylene	0.400	0.435	-8.7	118	-0.01
16	Acenaphthene	0.400	0.400	0.0	110	-0.01
17	Fluorene	0.400	0.373	6.8	115	-0.01
18 I	Phenanthrene-d10	0.400	0.400	0.0	99	-0.01
19	4-Bromophenyl-phenylether	0.400	0.417	-4.2	105	-0.01
20	Hexachlorobenzene	0.400	0.453	-13.2	114	-0.01
21	Pentachlorophenol	0.400	0.417	-4.2	119	-0.01
22	Phenanthrene	0.400	0.435	-8.7	112	-0.01
23	Anthracene	0.400	0.433	-8.2	115	-0.01
24 SURR	Fluoranthene-d10	0.400	0.412	-3.0	107	-0.01
25	Fluoranthene	0.400	0.422	-5.5	108	-0.01
26 I	Chrysene-d12	0.400	0.400	0.0	94	-0.01
27	Pvrene	0.400	0.455	-13.7	111	-0.02
28 S	Terphenyl-d14	0.400	0.456	-14.0	112	-0.02
29	Benzo(a)anthracene	0.400	0.434	-8.5	105	-0.01
30	Chrysene	0.400	0.450	-12.5	111	-0.01
31	Indeno(1,2,3-cd)pyrene	0.400	0.480	-20.0	120	-0.02
32 I	Perylene-d12	0.400	0.400	0.0	105	-0.01
33	Benzo(b)fluoranthene	0.400	0.425	-6.2	114	-0.01
34	Benzo(k)fluoranthene	0.400	0.423	-5.7	111	-0.01
35 C	Benzo(a)pyrene	0.400	0.430	-7.5	118	-0.02
36	Dibenzo(a,h)anthracene	0.400	0.443	-10.7	121	-0.03
37	Benzo(g,h,i)perylene	0.400	0.426	-6.5	114	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0