


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN081619\
 Data File : BN007218.D
 Acq On : 16 Aug 2019 10:52
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Aug 13 12:12:02 2019



AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2426

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.3	129322	PASS
68	69	0.00	2	2.0	2969	PASS
69	198	0.00	100	34.2	151039	PASS
70	69	0.00	2	0.7	1067	PASS
127	198	10	80	43.5	192170	PASS
197	198	0.00	2	0.6	2506	PASS
198	198	100	100	100.0	441557	PASS
199	198	5	9	7.3	32384	PASS
275	198	10	60	26.7	117688	PASS
365	198	1	100	3.8	16701	PASS
441	198	0.01	100	14.6	64568	PASS
442	198	50	100	94.6	417536	PASS
443	442	15	24	19.0	79328	PASS