

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/16/2023	BNA_N	<u>BN026924.D</u>
Compound Name	Response	Retention Time
DDT	2131633	20.927
DDD	21277	20.539
DDE	675	19.98
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
21952	2153585	1.02