

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081623\
 Data File : BN026925.D
 Acq On : 16 Aug 2023 10:00
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 16 15:56:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 15:56:02 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.430	0.419	2.6	86	0.00
3	n-Nitrosodimethylamine	0.479	0.461	3.8	82	0.00
4 S	2-Fluorophenol	1.127	1.060	5.9	82	0.00
5 S	Phenol-d6	1.392	1.277	8.3	80	0.00
6	bis(2-Chloroethyl)ether	1.171	1.145	2.2	82	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.338	0.278	17.8	72	0.00
9	Naphthalene	1.052	1.024	2.7	91	0.00
10	Hexachlorobutadiene	0.215	0.205	4.7	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.636	0.611	3.9	88	0.00
12	2-Methylnaphthalene	0.823	0.794	3.5	87	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
14 S	2,4,6-Tribromophenol	0.213	0.178	16.4	74	0.00
15 S	2-Fluorobiphenyl	1.485	1.473	0.8	86	0.00
16	Acenaphthylene	1.909	1.808	5.3	79	0.00
17	Acenaphthene	1.179	1.159	1.7	83	0.00
18	Fluorene	1.583	1.550	2.1	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.006	0.0	75	0.00
21	4-Bromophenyl-phenylether	0.228	0.227	0.4	79	0.00
22	Hexachlorobenzene	0.244	0.249	-2.0	81	0.00
23	Atrazine	0.203	0.184	9.4	71	0.00
24	Pentachlorophenol	0.123	0.105	14.6	71	0.00
25	Phenanthrene	1.150	1.124	2.3	78	0.00
26	Anthracene	1.073	1.007	6.2	75	0.00
27 SURR	Fluoranthene-d10	1.026	0.992	3.3	77	0.00
28	Fluoranthene	1.362	1.312	3.7	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
30	Pyrene	1.539	1.543	-0.3	78	0.00
31 S	Terphenyl-d14	0.836	0.806	3.6	78	0.00
32	Benzo(a)anthracene	1.444	1.374	4.8	74	0.00
33	Chrysene	1.457	1.423	2.3	75	0.00
34	Bis(2-ethylhexyl)phthalate	1.040	0.963	7.4	70	0.00
35 I	Perylene-d12	1.000	1.000	0.0	76	0.00
36	Indeno(1,2,3-cd)pyrene	1.638	1.571	4.1	75	0.00
37	Benzo(b)fluoranthene	1.339	1.287	3.9	78	0.00
38	Benzo(k)fluoranthene	1.376	1.288	6.4	73	0.00
39 C	Benzo(a)pyrene	1.341	1.285	4.2	77	0.00
40	Dibenzo(a,h)anthracene	1.304	1.232	5.5	73	0.00
41	Benzo(g,h,i)perylene	1.403	1.357	3.3	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4