

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015903.D
 Acq On : 17 Aug 2021 16:52
 Operator : CG/JU
 Sample : M3364-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBK84

Quant Time: Aug 17 17:22:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	249910	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	1029063	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	639593	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	1289517	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1281830	20.000	ng/u1	0.00
88) Perylene-d12	23.910	264	1265043	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	27978	4.361	ng/uL	0.00
4) Pyridine-d5	3.752	84	83205	4.640	ng/u1	0.00
7) Phenol-d5	7.058	99	139618	6.176	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	378649	27.205	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	375741	23.359	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	256608	14.595	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	234376	30.945	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	239338	33.764	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	434580	28.331	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	436385	18.689	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	1679083	35.821	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	1844175	31.416	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	50181	6.053	ng/u1	0.00
60) Fluorene-d10	15.539	176	1378744	33.720	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	237582	34.499	ng/u1	0.00
73) Anthracene-d10	17.398	188	2416056	39.974	ng/u1	0.00
81) Pyrene-d10	19.692	212	2840820	41.146	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	2747047	40.162	ng/u1	0.00
Target Compounds						
					Qvalue	
36) 2-Methylnaphthalene	12.375	142	103065	2.689	ng/u1	98
37) 1-Methylnaphthalene	12.593	142	166449	4.363	ng/u1	98
52) Acenaphthene	14.610	153	279916	6.935	ng/u1	95
61) Fluorene	15.592	166	95553	2.098	ng/u1	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
Data File : BN015903.D
Acq On : 17 Aug 2021 16:52
Operator : CG/JU
Sample : M3364-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBK84

Quant Time: Aug 17 17:22:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 17 09:43:53 2021
Response via : Initial Calibration

