

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015957.D
 Acq On : 19 Aug 2021 04:29
 Operator : CG/JU
 Sample : PB138537BL
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK537

Quant Time: Aug 19 07:04:31 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	274366	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	1150999	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	764970	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	1593200	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1501224	20.000	ng/u1	0.00
88) Perylene-d12	23.904	264	1465224	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	45179	6.414	ng/uL	0.00
4) Pyridine-d5	3.740	84	639490	32.486	ng/u1	0.00
7) Phenol-d5	7.058	99	838276	33.777	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	482344	31.566	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	634483	35.928	ng/u1	0.00
15) 4-Methylphenol-d8	8.604	113	661983	34.294	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	313947	37.059	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	335867	42.362	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	608328	35.457	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	893059	34.196	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	2007930	35.816	ng/u1	0.00
49) Acenaphthylene-d8	14.239	160	2485167	35.396	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	301847	30.441	ng/u1	0.00
60) Fluorene-d10	15.539	176	1718328	35.138	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	196897	23.141	ng/u1	0.00
73) Anthracene-d10	17.392	188	2755423	36.899	ng/u1	0.00
81) Pyrene-d10	19.686	212	3063745	37.889	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	2848897	35.961	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
Data File : BN015957.D
Acq On : 19 Aug 2021 04:29
Operator : CG/JU
Sample : PB138537BL
Misc :
ALS Vial : 63 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK537

Quant Time: Aug 19 07:04:31 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 17 09:43:53 2021
Response via : Initial Calibration

