

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015891.D
 Acq On : 17 Aug 2021 09:34
 Operator : CG/JU
 Sample : PB138361BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK361

Quant Time: Aug 17 10:11:35 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	267484	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	1118633	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	685008	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	1455844	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1386669	20.000	ng/u1	0.00
88) Perylene-d12	23.904	264	1366775	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	45732	6.660	ng/uL	0.00
4) Pyridine-d5	3.740	84	603721	31.458	ng/u1	0.00
7) Phenol-d5	7.052	99	785759	32.475	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	486605	32.664	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	612444	35.573	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	621430	33.022	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	294443	35.762	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	296251	38.446	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	559261	33.540	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	831035	32.742	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	1822083	36.295	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	2248359	35.762	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	270721	30.489	ng/u1	0.00
60) Fluorene-d10	15.540	176	1574431	35.954	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	215041	27.658	ng/u1	0.00
73) Anthracene-d10	17.398	188	2420945	35.479	ng/u1	0.00
81) Pyrene-d10	19.686	212	2733661	36.600	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.745	264	2678740	36.248	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
Data File : BN015891.D
Acq On : 17 Aug 2021 09:34
Operator : CG/JU
Sample : PB138361BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK361

Quant Time: Aug 17 10:11:35 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 17 09:43:53 2021
Response via : Initial Calibration

