

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015902.D
 Acq On : 17 Aug 2021 16:15
 Operator : CG/JU
 Sample : M3364-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBK79

Quant Time: Aug 17 16:54:09 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	265848	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	1087363	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	691075	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.292	188	1434944	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1423763	20.000	ng/u1	0.00
88) Perylene-d12	23.910	264	1405834	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	30316	4.442	ng/uL	0.00
4) Pyridine-d5	3.746	84	137708	7.220	ng/u1	0.00
7) Phenol-d5	7.052	99	125342	5.212	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	350602	23.680	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	350747	20.498	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	210758	11.268	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	219984	27.487	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	209699	27.997	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	391171	24.134	ng/u1	0.00
31) 4-Chloroaniline-d4	10.846	131	348421	14.122	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	1531733	30.243	ng/u1	0.00
49) Acenaphthylene-d8	14.240	160	1697286	26.759	ng/u1	0.00
54) 4-Nitrophenol-d4	14.728	143	39430	4.402	ng/u1	0.00
60) Fluorene-d10	15.540	176	1264006	28.611	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	196634	25.659	ng/u1	0.00
73) Anthracene-d10	17.398	188	2040855	30.344	ng/u1	0.00
81) Pyrene-d10	19.692	212	2332409	30.414	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	2280669	30.004	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
Data File : BN015902.D
Acq On : 17 Aug 2021 16:15
Operator : CG/JU
Sample : M3364-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBK79

Quant Time: Aug 17 16:54:09 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 17 09:43:53 2021
Response via : Initial Calibration

