

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015939.D
 Acq On : 18 Aug 2021 16:17
 Operator : CG/JU
 Sample : M3362-03MSD
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JMXD6MSD

Quant Time: Aug 18 17:24:27 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.899	152	182904	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	772487	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	491436	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.298	188	986394	20.000	ng/u1	0.00
79) Chrysene-d12	21.492	240	944747	20.000	ng/u1	0.00
88) Perylene-d12	23.915	264	935108	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	19483	4.149	ng/uL	0.00
4) Pyridine-d5	3.746	84	183590	13.990	ng/u1	0.00
7) Phenol-d5	7.063	99	362846	21.931	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	221456	21.740	ng/u1	0.00
11) 2-Chlorophenol-d4	7.428	132	290024	24.635	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	303695	23.600	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	140789	24.762	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	146724	27.574	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	293231	25.466	ng/u1	0.00
31) 4-Chloroaniline-d4	10.845	131	362569	20.686	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	978435	27.167	ng/u1	0.00
49) Acenaphthylene-d8	14.239	160	1132060	25.099	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	148701	23.344	ng/u1	0.00
60) Fluorene-d10	15.539	176	803414	25.573	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.657	200	125204	23.768	ng/u1	0.00
73) Anthracene-d10	17.398	188	1266461	27.393	ng/u1	0.00
81) Pyrene-d10	19.692	212	1395948	27.432	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	1247458	24.673	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.369	88	53967	11.180	ng/uL	88
5) Pyridine	3.764	79	313602	23.179	ng/u1	97
6) Benzaldehyde	7.040	77	301733	35.445	ng/u1	95
8) Phenol	7.087	94	543294	32.212	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.322	93	409181	30.967	ng/u1	96
12) 2-Chlorophenol	7.463	128	413476	34.169	ng/u1	98
13) 2-Methylphenol	8.346	108	388555	31.450	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.434	45	503543	29.841	ng/u1	98
16) Acetophenone	8.728	105	664380	31.994	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.710	70	350645	34.579	ng/u1	96
18) 4-Methylphenol	8.675	108	432368	32.699	ng/u1	94
19) Hexachloroethane	8.987	117	171435	31.198	ng/u1	89
22) Nitrobenzene	9.104	77	515070	32.038	ng/u1	99
23) Isophorone	9.634	82	912577	32.627	ng/u1	98
25) 2-Nitrophenol	9.822	139	221907	38.241	ng/u1	98
26) 2,4-Dimethylphenol	9.881	107	465923	30.061	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.122	93	551513	32.103	ng/u1	99
29) 2,4-Dichlorophenol	10.357	162	389050	34.268	ng/u1	99
30) Naphthalene	10.763	128	1292825	31.248	ng/u1	99
32) 4-Chloroaniline	10.869	127	498304	28.796	ng/u1	99
33) Hexachlorobutadiene	11.051	225	284433	32.606	ng/u1	95
34) Caprolactam	11.634	113	120951	34.003	ng/u1	96
35) 4-Chloro-3-methylphenol	11.998	107	472709	35.152	ng/u1	100
36) 2-Methylnaphthalene	12.375	142	948600	32.971	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015939.D
 Acq On : 18 Aug 2021 16:17
 Operator : CG/JU
 Sample : M3362-03MSD
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 JMXD6MSD

Quant Time: Aug 18 17:24:27 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 17 09:43:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.592	142	907103	31.675	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.745	216	524051	32.963	ng/ul	96
40) Hexachlorocyclopentadiene	12.728	237	242140	23.492	ng/ul	98
41) 2,4,6-Trichlorophenol	12.981	196	328807	37.824	ng/ul	98
42) 2,4,5-Trichlorophenol	13.051	196	350471	36.927	ng/ul	94
43) 1,1'-Biphenyl	13.381	154	1243629	32.726	ng/ul	97
44) 2-Chloronaphthalene	13.428	162	949745	32.367	ng/ul	96
45) 2-Nitroaniline	13.628	65	318969	38.363	ng/ul	98
47) Dimethylphthalate	14.004	163	1246268	34.907	ng/ul	98
48) 2,6-Dinitrotoluene	14.122	165	249626	38.608	ng/ul	90
50) Acenaphthylene	14.269	152	1434268	33.451	ng/ul	100
51) 3-Nitroaniline	14.451	138	237638	34.183	ng/ul	92
52) Acenaphthene	14.616	153	1097423	35.386	ng/ul	96
53) 2,4-Dinitrophenol	14.657	184	110518	30.763	ng/ul	91
55) 4-Nitrophenol	14.757	109	213609	34.270	ng/ul	92
56) Dibenzofuran	14.945	168	1444286	33.685	ng/ul	99
57) 2,4-Dinitrotoluene	14.910	165	364617	39.152	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.175	232	306967	39.087	ng/ul	97
59) Diethylphthalate	15.369	149	1303870	36.522	ng/ul	99
61) Fluorene	15.598	166	1231039	35.184	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.592	204	623577	34.178	ng/ul	99
63) 4-Nitroaniline	15.616	138	262267	39.373	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.669	198	180929	34.532	ng/ul#	97
67) N-Nitrosodiphenylamine	15.804	169	1014954	35.253	ng/ul	98
68) 4-Bromophenyl-phenylether	16.486	248	389053	36.755	ng/ul	96
69) Hexachlorobenzene	16.604	284	433796	35.070	ng/ul	96
70) Atrazine	16.757	200	402860	38.578	ng/ul	98
71) Pentachlorophenol	16.945	266	239377	34.496	ng/ul	96
72) Phenanthrene	17.339	178	2173615	40.535	ng/ul	99
74) Anthracene	17.433	178	1920913	35.088	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.345	216	518562	33.376	ng/ul	95
76) Pentachlorobenzene	14.869	250	510676	33.767	ng/ul	97
77) Carbazole	17.704	167	1712419	35.576	ng/ul	98
78) Di-n-butylphthalate	18.269	149	2150759	39.299	ng/ul	99
80) Fluoranthene	19.357	202	2393813	38.881	ng/ul	99
82) Pyrene	19.721	202	2376131	37.081	ng/ul	95
83) Butylbenzylphthalate	20.621	149	899381	41.737	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.410	252	628404	32.482	ng/ul	96
85) Benzo(a)anthracene	21.474	228	2233653	36.746	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.404	149	1382234	40.446	ng/ul	96
87) Chrysene	21.527	228	2173099	36.925	ng/ul	99
89) Di-n-octyl phthalate	22.339	149	2308706	39.009	ng/ul	100
90) Benzo(b)fluoranthene	23.180	252	2238248	36.407	ng/ul	100
91) Benzo(k)fluoranthene	23.221	252	2199328	35.930	ng/ul	99
93) Benzo(a)pyrene	23.810	252	1698536	30.521	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.427	276	2336814	33.865	ng/ul	98
95) Dibenzo(a,h)anthracene	26.451	278	2103384	36.923	ng/ul	97
96) Benzo(g,h,i)perylene	27.192	276	265632	4.656	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.180 ng/uL
RT: 3.369 min Scan# 48
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:	88	Resp:	53967
Ion	Ratio	Lower	Upper
88	100		
43	32.2	26.4	39.6
58	66.6	65.0	97.4



#5
Pyridine
Concen: 23.179 ng/uL
RT: 3.764 min Scan# 115
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:	79	Resp:	313602
Ion	Ratio	Lower	Upper
79	100		
52	75.9	62.9	94.3
51	38.5	32.0	48.0









#6
Benzaldehyde
Concen: 35.445 ng/ul
RT: 7.040 min Scan# 672
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion: 77 Resp: 301733
Ion Ratio Lower Upper
77 100
105 96.3 73.4 110.0
106 84.0 70.3 105.5



#8
Phenol
Concen: 32.212 ng/ul
RT: 7.087 min Scan# 680
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion: 94 Resp: 543294
Ion Ratio Lower Upper
94 100
65 31.1 27.0 40.6
66 44.4 37.8 56.8









#10
Bis(2-Chloroethyl)ether
Concen: 30.967 ng/ul
RT: 7.322 min Scan# 720
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Instrument :
BNA_N
ClientSampleId :
JMXD6MSD

Tgt Ion: 93 Resp: 409181




Ion	Ratio	Lower	Upper
93	100		
63	74.2	63.4	95.0
95	31.5	25.2	37.8




#12
2-Chlorophenol
Concen: 34.169 ng/ul
RT: 7.463 min Scan# 744
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion: 128 Resp: 413476

Ion	Ratio	Lower	Upper
128	100		
64	50.9	41.4	62.2
130	33.5	25.2	37.8









#13
2-Methylphenol
Concen: 31.450 ng/ul
RT: 8.346 min Scan# 894
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:108 Resp: 388555
Ion Ratio Lower Upper
108 100
107 98.0 74.9 112.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.841 ng/ul
RT: 8.434 min Scan# 909
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion: 45 Resp: 503543
Ion Ratio Lower Upper
45 100
77 16.4 13.2 19.8
79 12.5 8.7 13.1









#16
Acetophenone
Concen: 31.994 ng/ul
RT: 8.728 min Scan# 959
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:105 Resp: 664380
Ion Ratio Lower Upper
105 100
77 82.4 66.3 99.5
51 28.1 23.7 35.5



#17
N-Nitroso-di-n-propylamine
Concen: 34.579 ng/ul
RT: 8.710 min Scan# 956
Delta R.T. -0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion: 70 Resp: 350645
Ion Ratio Lower Upper
70 100
42 53.1 45.3 67.9
101 8.7 7.6 11.4
130 17.7 15.0 22.6









#18
4-Methylphenol
Concen: 32.699 ng/ul
RT: 8.675 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:108 Resp: 432368
Ion Ratio Lower Upper
108 100
107 112.6 85.3 127.9



#19
Hexachloroethane
Concen: 31.198 ng/ul
RT: 8.987 min Scan# 1003
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:117 Resp: 171435
Ion Ratio Lower Upper
117 100
201 104.9 74.0 111.0
199 62.8 45.0 67.4









#22
Nitrobenzene
Concen: 32.038 ng/ul
RT: 9.104 min Scan# 1023
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion: 77 Resp: 515070
Ion Ratio Lower Upper
77 100
123 40.5 32.2 48.4
65 14.1 10.9 16.3



#23
Isophorone
Concen: 32.627 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion: 82 Resp: 912577
Ion Ratio Lower Upper
82 100
95 8.1 6.1 9.1
138 17.1 13.0 19.4









#25
2-Nitrophenol
Concen: 38.241 ng/ul
RT: 9.822 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Instrument :
BNA_N
ClientSampleId :
JMXD6MSD

Tgt Ion:139 Resp: 221907




Ion	Ratio	Lower	Upper
139	100		
65	62.4	48.1	72.1
109	37.6	29.7	44.5




#26
2,4-Dimethylphenol
Concen: 30.061 ng/ul
RT: 9.881 min Scan# 1155
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:107 Resp: 465923

Ion	Ratio	Lower	Upper
107	100		
121	47.7	35.3	52.9
122	77.3	59.4	89.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.103 ng/ul
 RT: 10.122 min Scan# 1196
 Delta R.T. -0.000 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17

Instrument :
 BNA_N
 ClientSampleId :
 JMXD6MSD

Tgt Ion: 93 Resp: 551513




Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.9	40.3
123	10.6	8.3	12.5




#29
 2,4-Dichlorophenol
 Concen: 34.268 ng/ul
 RT: 10.357 min Scan# 1236
 Delta R.T. 0.006 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17

Tgt Ion: 162 Resp: 389050

Ion	Ratio	Lower	Upper
162	100		
164	65.6	51.9	77.9
98	43.2	34.2	51.2









#30
Naphthalene
Concen: 31.248 ng/ul
RT: 10.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:128 Resp: 1292825
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 14.3 11.1 16.7



#32
4-Chloroaniline
Concen: 28.796 ng/ul
RT: 10.869 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:127 Resp: 498304
Ion Ratio Lower Upper
127 100
129 32.8 25.6 38.4









#33
Hexachlorobutadiene
Concen: 32.606 ng/ul
RT: 11.051 min Scan# 1354
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:	225	Resp:	284433
Ion Ratio	Lower	Upper	
225	100		
223	59.8	50.5	75.7
227	62.5	53.9	80.9



#34
Caprolactam
Concen: 34.003 ng/ul
RT: 11.634 min Scan# 1453
Delta R.T. 0.012 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:	113	Resp:	120951
Ion Ratio	Lower	Upper	
113	100		
55	167.9	130.4	195.6
56	131.8	102.2	153.4









#35
4-Chloro-3-methylphenol
Concen: 35.152 ng/ul
RT: 11.998 min Scan# 1515
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:107 Resp: 472709
Ion Ratio Lower Upper
107 100
144 22.3 18.2 27.4
142 71.1 57.1 85.7



#36
2-Methylnaphthalene
Concen: 32.971 ng/ul
RT: 12.375 min Scan# 1579
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:142 Resp: 948600
Ion Ratio Lower Upper
142 100
141 90.0 73.4 110.0









#37
1-Methylnaphthalene
Concen: 31.675 ng/ul
RT: 12.592 min Scan# 1616
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:142 Resp: 907103
Ion Ratio Lower Upper
142 100
141 89.5 74.2 111.4
116 4.5 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.963 ng/ul
RT: 12.745 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:216 Resp: 524051
Ion Ratio Lower Upper
216 100
214 74.6 62.7 94.1
179 20.7 18.6 27.8
108 17.3 15.2 22.8









#40
Hexachlorocyclopentadiene
Concen: 23.492 ng/ul
RT: 12.728 min Scan# 1639
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:237 Resp: 242140
Ion Ratio Lower Upper
237 100
235 59.5 49.0 73.4
272 13.0 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 37.824 ng/ul
RT: 12.981 min Scan# 1682
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:196 Resp: 328807
Ion Ratio Lower Upper
196 100
198 99.1 79.9 119.9
200 32.4 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 36.927 ng/ul
RT: 13.051 min Scan# 1694
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:196 Resp: 350471
Ion Ratio Lower Upper
196 100
198 103.5 76.5 114.7
200 31.2 24.6 37.0



#43
1,1'-Biphenyl
Concen: 32.726 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:154 Resp: 1243629
Ion Ratio Lower Upper
154 100
153 38.2 32.1 48.1
76 12.9 11.0 16.6









#44
2-Chloronaphthalene
Concen: 32.367 ng/ul
RT: 13.428 min Scan# 1758
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:	162	Resp:	949745
Ion	Ratio	Lower	Upper
162	100		
164	31.6	26.6	40.0
127	42.4	36.3	54.5



#45
2-Nitroaniline
Concen: 38.363 ng/ul
RT: 13.628 min Scan# 1792
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:	65	Resp:	318969
Ion	Ratio	Lower	Upper
65	100		
92	67.7	54.5	81.7
138	96.9	75.4	113.0









#47
Dimethylphthalate
Concen: 34.907 ng/ul
RT: 14.004 min Scan# 1856
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:163 Resp: 1246268
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.0 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 38.608 ng/ul
RT: 14.122 min Scan# 1876
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17


Tgt Ion:165 Resp: 249626
Ion Ratio Lower Upper
165 100
89 65.5 59.5 89.3
121 21.2 20.0 30.0








#50
Acenaphthylene
Concen: 33.451 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17


Instrument :
BNA_N
ClientSampleId :
JMXD6MSD





Tgt Ion:152 Resp: 1434268
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.0 10.3 15.5




#51
3-Nitroaniline
Concen: 34.183 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17



Tgt Ion:138 Resp: 237638
Ion Ratio Lower Upper
138 100
108 12.0 10.3 15.5
92 120.0 103.9 155.9









#52
Acenaphthene
Concen: 35.386 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:153 Resp: 1097423
Ion Ratio Lower Upper
153 100
152 46.3 38.6 57.8
154 91.3 69.3 103.9



#53
2,4-Dinitrophenol
Concen: 30.763 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:184 Resp: 110518
Ion Ratio Lower Upper
184 100
63 68.7 64.8 97.2
154 71.7 59.7 89.5









#55
4-Nitrophenol
Concen: 34.270 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:109 Resp: 213609
Ion Ratio Lower Upper
109 100
139 102.9 84.7 127.1
65 117.4 105.0 157.6



#56
Dibenzofuran
Concen: 33.685 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:168 Resp: 1444286
Ion Ratio Lower Upper
168 100
139 38.9 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 39.152 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:165 Resp: 364617
Ion Ratio Lower Upper
165 100
63 48.3 42.8 64.2
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.087 ng/ul
RT: 15.175 min Scan# 2055
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:232 Resp: 306967
Ion Ratio Lower Upper
232 100
131 45.7 37.5 56.3
130 2.5 2.2 3.4
166 30.2 26.2 39.2









#59
Diethylphthalate
Concen: 36.522 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:149 Resp: 1303870
Ion Ratio Lower Upper
149 100
177 21.2 17.2 25.8
150 13.0 10.2 15.2



#61
Fluorene
Concen: 35.184 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:166 Resp: 1231039
Ion Ratio Lower Upper
166 100
165 99.4 78.4 117.6
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 34.178 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:204 Resp: 623577
Ion Ratio Lower Upper
204 100
206 33.6 26.1 39.1
141 65.7 53.4 80.0



#63
4-Nitroaniline
Concen: 39.373 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:138 Resp: 262267
Ion Ratio Lower Upper
138 100
92 57.7 48.6 72.8
108 100.3 80.7 121.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 34.532 ng/ul
 RT: 15.669 min Scan# 2139
 Delta R.T. -0.000 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17




Instrument :
 BNA_N
 ClientSampleId :
 JMXD6MSD


Tgt Ion:198 Resp: 180929
 Ion Ratio Lower Upper
 198 100
 182 4.6 3.0 4.4#
 77 30.8 23.4 35.0



#67
 N-Nitrosodiphenylamine
 Concen: 35.253 ng/ul
 RT: 15.804 min Scan# 2162
 Delta R.T. -0.000 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17

Tgt Ion:169 Resp: 1014954
 Ion Ratio Lower Upper
 169 100
 168 66.1 53.5 80.3
 167 36.7 27.9 41.9









#68
4-Bromophenyl-phenylether
Concen: 36.755 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:248 Resp: 389053
Ion Ratio Lower Upper
248 100
250 106.4 81.1 121.7
141 83.0 64.7 97.1



#69
Hexachlorobenzene
Concen: 35.070 ng/ul
RT: 16.604 min Scan# 2298
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:284 Resp: 433796
Ion Ratio Lower Upper
284 100
142 35.7 30.3 45.5
249 30.4 26.3 39.5









#70
Atrazine
Concen: 38.578 ng/ul
RT: 16.757 min Scan# 2324
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:200 Resp: 402860
Ion Ratio Lower Upper
200 100
173 27.0 22.2 33.2
215 45.0 37.6 56.4



#71
Pentachlorophenol
Concen: 34.496 ng/ul
RT: 16.945 min Scan# 2356
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:266 Resp: 239377
Ion Ratio Lower Upper
266 100
264 62.3 53.8 80.6
268 67.0 52.7 79.1









#72
Phenanthrene
Concen: 40.535 ng/u1
RT: 17.339 min Scan# 2423
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:178 Resp: 2173615
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 18.7 14.6 21.8



#74
Anthracene
Concen: 35.088 ng/u1
RT: 17.433 min Scan# 2439
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:178 Resp: 1920913
Ion Ratio Lower Upper
178 100
179 15.3 11.6 17.4
176 18.3 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 33.376 ng/uL
RT: 13.345 min Scan# 1744
Delta R.T. -0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:216 Resp: 518562
Ion Ratio Lower Upper
216 100
214 74.1 64.6 97.0
218 47.5 38.7 58.1



#76
Pentachlorobenzene
Concen: 33.767 ng/uL
RT: 14.869 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:250 Resp: 510676
Ion Ratio Lower Upper
250 100
248 63.2 51.9 77.9
252 65.4 54.7 82.1









#77
 Carbazole
 Concen: 35.576 ng/ul
 RT: 17.704 min Scan# 2485
 Delta R.T. 0.006 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17

Instrument :
 BNA_N
 ClientSampleId :
 JMXD6MSD

Tgt Ion:167 Resp: 1712419




Ion	Ratio	Lower	Upper
167	100		
166	20.9	17.1	25.7
139	13.1	11.6	17.4




#78
 Di-n-butylphthalate
 Concen: 39.299 ng/ul
 RT: 18.269 min Scan# 2581
 Delta R.T. -0.000 min
 Lab File: BN015939.D
 Acq: 18 Aug 2021 16:17

Tgt Ion:149 Resp: 2150759

Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.5	11.3
104	6.0	5.0	7.4









#80
Fluoranthene
Concen: 38.881 ng/ul
RT: 19.357 min Scan# 2766
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:202 Resp: 2393813
Ion Ratio Lower Upper
202 100
101 10.4 8.0 12.0
100 8.0 6.7 10.1



#82
Pyrene
Concen: 37.081 ng/ul
RT: 19.721 min Scan# 2828
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:202 Resp: 2376131
Ion Ratio Lower Upper
202 100
101 10.8 10.1 15.1
100 9.5 9.2 13.8









#83
Butylbenzylphthalate
Concen: 41.737 ng/ul
RT: 20.621 min Scan# 2981
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Instrument :
BNA_N
ClientSampleId :
JMXD6MSD

Tgt Ion:149 Resp: 899381




Ion	Ratio	Lower	Upper
149	100		
91	80.7	65.2	97.8
206	21.1	17.1	25.7




#84
3,3'-Dichlorobenzidine
Concen: 32.482 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:252 Resp: 628404

Ion	Ratio	Lower	Upper
252	100		
254	62.2	52.0	78.0
126	8.6	8.5	12.7









#85
Benzo(a)anthracene
Concen: 36.746 ng/ul
RT: 21.474 min Scan# 3126
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampled :
JMXD6MSD


Tgt Ion:228 Resp: 2233653
Ion Ratio Lower Upper
228 100
229 19.6 14.5 21.7
226 26.3 20.9 31.3



#86
Bis(2-ethylhexyl)phthalate
Concen: 40.446 ng/ul
RT: 21.404 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:149 Resp: 1382234
Ion Ratio Lower Upper
149 100
167 27.8 20.6 30.8
279 4.9 3.4 5.2









#87
Chrysene
Concen: 36.925 ng/ul
RT: 21.527 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Instrument :
BNA_N
ClientSampleId :
JMXD6MSD




Tgt Ion:228 Resp: 2173099


Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.2	36.2
229	19.4	15.9	23.9



#89
Di-n-octyl phthalate
Concen: 39.009 ng/ul
RT: 22.339 min Scan# 3273
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:149 Resp: 2308706









#90
Benzo(b)fluoranthene
Concen: 36.407 ng/ul
RT: 23.180 min Scan# 3416
Delta R.T. 0.012 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:252 Resp: 2238248
Ion Ratio Lower Upper
252 100
253 21.0 16.6 25.0
125 8.7 7.1 10.7



#91
Benzo(k)fluoranthene
Concen: 35.930 ng/ul
RT: 23.221 min Scan# 3423
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:252 Resp: 2199328
Ion Ratio Lower Upper
252 100
253 21.3 17.7 26.5
125 8.6 7.0 10.6









#93
Benzo(a)pyrene
Concen: 30.521 ng/ul
RT: 23.810 min Scan# 3523
Delta R.T. 0.012 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17




Instrument :
BNA_N
ClientSampleId :
JMXD6MSD


Tgt Ion:252 Resp: 1698536
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 8.8 7.2 10.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 33.865 ng/ul
RT: 26.427 min Scan# 3968
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:276 Resp: 2336814
Ion Ratio Lower Upper
276 100
138 18.8 15.9 23.9
277 25.3 19.5 29.3









#95
Dibenzo(a,h)anthracene
Concen: 36.923 ng/ul
RT: 26.451 min Scan# 3972
Delta R.T. 0.006 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Instrument :
BNA_N
ClientSampleId :
JMXD6MSD

Tgt Ion:278 Resp: 2103384
Ion Ratio Lower Upper
278 100
139 13.3 11.0 16.6
279 23.3 19.9 29.9



#96
Benzo(g,h,i)perylene
Concen: 4.656 ng/ul
RT: 27.192 min Scan# 4098
Delta R.T. -0.000 min
Lab File: BN015939.D
Acq: 18 Aug 2021 16:17

Tgt Ion:276 Resp: 265632
Ion Ratio Lower Upper
276 100
138 15.8 13.4 20.0
277 21.6 18.5 27.7

