

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015972.D
 Acq On : 19 Aug 2021 15:26
 Operator : CG/JU
 Sample : PB138520BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS520

Quant Time: Aug 19 16:16:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 19 15:06:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.875	152	226530	20.000	ng/u1	0.00
20) Naphthalene-d8	10.681	136	928680	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.516	164	609067	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.263	188	1293929	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	1355717	20.000	ng/u1	0.00
88) Perylene-d12	23.851	264	1360518	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.323	96	36263	6.236	ng/uL	0.00
4) Pyridine-d5	3.728	84	504228	31.023	ng/u1	0.00
7) Phenol-d5	7.040	99	674853	32.934	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.205	67	385543	30.559	ng/u1	0.00
11) 2-Chlorophenol-d4	7.405	132	513304	35.204	ng/u1	0.00
15) 4-Methylphenol-d8	8.581	113	538967	33.818	ng/u1	0.00
21) Nitrobenzene-d5	9.034	128	244567	35.780	ng/u1	0.00
24) 2-Nitrophenol-d4	9.757	143	274840	42.963	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.299	165	515405	37.232	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	642440	30.488	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	1577806	35.348	ng/u1	0.00
49) Acenaphthylene-d8	14.204	160	1931379	34.550	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	291247	36.891	ng/u1	0.00
60) Fluorene-d10	15.504	176	1340897	34.439	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.622	200	275493	39.867	ng/u1	0.00
73) Anthracene-d10	17.357	188	2172199	35.817	ng/u1	0.00
81) Pyrene-d10	19.657	212	2515091	34.442	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.698	264	2498376	33.963	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	77392	12.946	ng/uL	98
5) Pyridine	3.752	79	549629	32.800	ng/u1	98
6) Benzaldehyde	7.016	77	413499	39.220	ng/u1	99
8) Phenol	7.064	94	743232	35.580	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.299	93	561405	34.305	ng/u1	94
12) 2-Chlorophenol	7.440	128	563049	37.569	ng/u1	98
13) 2-Methylphenol	8.316	108	555682	36.316	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.405	45	687002	32.872	ng/u1	97
16) Acetophenone	8.699	105	901560	35.054	ng/u1	94
17) N-Nitroso-di-n-propyla...	8.687	70	472767	37.643	ng/u1	97
18) 4-Methylphenol	8.646	108	591159	36.098	ng/u1	87
19) Hexachloroethane	8.958	117	244246	35.888	ng/u1	95
22) Nitrobenzene	9.075	77	714841	36.985	ng/u1	98
23) Isophorone	9.605	82	1272806	37.853	ng/u1	99
25) 2-Nitrophenol	9.787	139	309842	44.415	ng/u1	97
26) 2,4-Dimethylphenol	9.852	107	693097	37.197	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.087	93	745359	36.089	ng/u1	97
29) 2,4-Dichlorophenol	10.322	162	537644	39.392	ng/u1	96
30) Naphthalene	10.728	128	1776563	35.718	ng/u1	99
32) 4-Chloroaniline	10.834	127	717407	34.485	ng/u1	98
33) Hexachlorobutadiene	11.016	225	377692	36.015	ng/u1	98
34) Caprolactam	11.599	113	185243	43.318	ng/u1	94
35) 4-Chloro-3-methylphenol	11.963	107	665788	41.183	ng/u1	99
36) 2-Methylnaphthalene	12.340	142	1285544	37.167	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015972.D
 Acq On : 19 Aug 2021 15:26
 Operator : CG/JU
 Sample : PB138520BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS520

Quant Time: Aug 19 16:16:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 19 15:06:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.557	142	1216581	35.337	ng/ul	93
39) 1,2,4,5-Tetrachloroben...	12.710	216	712492	36.160	ng/ul	96
40) Hexachlorocyclopentadiene	12.687	237	424283	33.213	ng/ul	96
41) 2,4,6-Trichlorophenol	12.946	196	452660	42.014	ng/ul	93
42) 2,4,5-Trichlorophenol	13.016	196	490558	41.704	ng/ul	99
43) 1,1'-Biphenyl	13.351	154	1632257	34.658	ng/ul	97
44) 2-Chloronaphthalene	13.393	162	1298985	35.720	ng/ul	97
45) 2-Nitroaniline	13.593	65	446562	43.336	ng/ul	97
47) Dimethylphthalate	13.969	163	1686194	38.107	ng/ul	99
48) 2,6-Dinitrotoluene	14.093	165	351268	43.836	ng/ul	89
50) Acenaphthylene	14.234	152	1961530	36.913	ng/ul	99
51) 3-Nitroaniline	14.416	138	327257	37.983	ng/ul	99
52) Acenaphthene	14.581	153	1383449	35.993	ng/ul	96
53) 2,4-Dinitrophenol	14.622	184	196482	44.128	ng/ul	91
55) 4-Nitrophenol	14.722	109	316872	41.019	ng/ul	95
56) Dibenzofuran	14.916	168	1987088	37.394	ng/ul	100
57) 2,4-Dinitrotoluene	14.875	165	519849	45.040	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.140	232	442979	45.512	ng/ul	93
59) Diethylphthalate	15.340	149	1800093	40.684	ng/ul	99
61) Fluorene	15.563	166	1644820	37.931	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.557	204	846134	37.419	ng/ul	97
63) 4-Nitroaniline	15.581	138	365739	44.303	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.640	198	289116	42.066	ng/ul	98
67) N-Nitrosodiphenylamine	15.769	169	1456920	38.577	ng/ul	99
68) 4-Bromophenyl-phenylether	16.451	248	556630	40.088	ng/ul	98
69) Hexachlorobenzene	16.569	284	633674	39.053	ng/ul	96
70) Atrazine	16.722	200	569610	41.581	ng/ul	98
71) Pentachlorophenol	16.910	266	367098	40.328	ng/ul	92
72) Phenanthrene	17.304	178	2703340	38.431	ng/ul	98
74) Anthracene	17.398	178	2745502	38.230	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.316	216	694260	34.064	ng/ul	98
76) Pentachlorobenzene	14.834	250	701439	35.357	ng/ul	99
77) Carbazole	17.663	167	2512142	39.786	ng/ul	98
78) Di-n-butylphthalate	18.234	149	3008904	41.912	ng/ul	99
80) Fluoranthene	19.322	202	3296231	37.308	ng/ul	99
82) Pyrene	19.686	202	3303357	35.924	ng/ul#	95
83) Butylbenzylphthalate	20.586	149	1359445	43.963	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.369	252	1107152	39.881	ng/ul	98
85) Benzo(a)anthracene	21.439	228	3291597	37.735	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.369	149	2024874	41.289	ng/ul	97
87) Chrysene	21.492	228	3215749	38.077	ng/ul	96
89) Di-n-octyl phthalate	22.292	149	3464090	40.229	ng/ul	100
90) Benzo(b)fluoranthene	23.121	252	3367168	37.644	ng/ul	96
91) Benzo(k)fluoranthene	23.174	252	3315911	37.233	ng/ul	98
93) Benzo(a)pyrene	23.751	252	3007124	37.139	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.351	276	3859119	38.439	ng/ul	98
95) Dibenzo(a,h)anthracene	26.368	278	3211505	38.748	ng/ul	98
96) Benzo(g,h,i)perylene	27.115	276	3131254	37.722	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.946 ng/uL
RT: 3.358 min Scan# 46
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion: 88 Resp: 77392
Ion Ratio Lower Upper
88 100
43 33.0 26.4 39.6
58 78.4 65.0 97.4



#5
Pyridine
Concen: 32.800 ng/uL
RT: 3.752 min Scan# 113
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 79 Resp: 549629
Ion Ratio Lower Upper
79 100
52 76.7 62.9 94.3
51 39.7 32.0 48.0









#6
Benzaldehyde
Concen: 39.220 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion: 77 Resp: 413499
Ion Ratio Lower Upper
77 100
105 92.2 73.4 110.0
106 88.7 70.3 105.5



#8
Phenol
Concen: 35.580 ng/ul
RT: 7.064 min Scan# 676
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 94 Resp: 743232
Ion Ratio Lower Upper
94 100
65 30.2 27.0 40.6
66 47.5 37.8 56.8









#10
Bis(2-Chloroethyl)ether
Concen: 34.305 ng/ul
RT: 7.299 min Scan# 716
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion: 93 Resp: 561405
Ion Ratio Lower Upper
93 100
63 73.4 63.4 95.0
95 33.0 25.2 37.8



#12
2-Chlorophenol
Concen: 37.569 ng/ul
RT: 7.440 min Scan# 740
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 128 Resp: 563049
Ion Ratio Lower Upper
128 100
64 52.0 41.4 62.2
130 34.6 25.2 37.8









#13
2-Methylphenol
Concen: 36.316 ng/ul
RT: 8.316 min Scan# 889
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:108 Resp: 555682
Ion Ratio Lower Upper
108 100
107 90.4 74.9 112.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.872 ng/ul
RT: 8.405 min Scan# 904
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 45 Resp: 687002
Ion Ratio Lower Upper
45 100
77 15.2 13.2 19.8
79 12.1 8.7 13.1









#16
Acetophenone
Concen: 35.054 ng/ul
RT: 8.699 min Scan# 954
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:105 Resp: 901560
Ion Ratio Lower Upper
105 100
77 77.2 66.3 99.5
51 27.8 23.7 35.5



#17
N-Nitroso-di-n-propylamine
Concen: 37.643 ng/ul
RT: 8.687 min Scan# 952
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 70 Resp: 472767
Ion Ratio Lower Upper
70 100
42 53.4 45.3 67.9
101 9.2 7.6 11.4
130 18.0 15.0 22.6









#18
4-Methylphenol
Concen: 36.098 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:108 Resp: 591159
Ion Ratio Lower Upper
108 100
107 120.2 85.3 127.9



#19
Hexachloroethane
Concen: 35.888 ng/ul
RT: 8.958 min Scan# 998
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:117 Resp: 244246
Ion Ratio Lower Upper
117 100
201 95.2 74.0 111.0
199 61.7 45.0 67.4









#22
Nitrobenzene
Concen: 36.985 ng/ul
RT: 9.075 min Scan# 1018
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion: 77 Resp: 714841
Ion Ratio Lower Upper
77 100
123 42.1 32.2 48.4
65 14.1 10.9 16.3



#23
Isophorone
Concen: 37.853 ng/ul
RT: 9.605 min Scan# 1108
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 82 Resp: 1272806
Ion Ratio Lower Upper
82 100
95 8.1 6.1 9.1
138 16.4 13.0 19.4









#25
2-Nitrophenol
Concen: 44.415 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:139 Resp: 309842
Ion Ratio Lower Upper
139 100
65 57.9 48.1 72.1
109 38.4 29.7 44.5



#26
2,4-Dimethylphenol
Concen: 37.197 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:107 Resp: 693097
Ion Ratio Lower Upper
107 100
121 45.4 35.3 52.9
122 79.3 59.4 89.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 36.089 ng/ul
 RT: 10.087 min Scan# 1190
 Delta R.T. 0.000 min
 Lab File: BN015972.D
 Acq: 19 Aug 2021 15:26

Instrument :
 BNA_N
 ClientSampleId :
 SLCS520

Tgt Ion: 93 Resp: 745359




Ion	Ratio	Lower	Upper
93	100		
95	31.4	26.9	40.3
123	10.3	8.3	12.5




#29
 2,4-Dichlorophenol
 Concen: 39.392 ng/ul
 RT: 10.322 min Scan# 1230
 Delta R.T. 0.000 min
 Lab File: BN015972.D
 Acq: 19 Aug 2021 15:26

Tgt Ion: 162 Resp: 537644

Ion	Ratio	Lower	Upper
162	100		
164	68.9	51.9	77.9
98	44.6	34.2	51.2









#30
Naphthalene
Concen: 35.718 ng/ul
RT: 10.728 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:128 Resp: 1776563
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.4 11.1 16.7



#32
4-Chloroaniline
Concen: 34.485 ng/ul
RT: 10.834 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:127 Resp: 717407
Ion Ratio Lower Upper
127 100
129 33.4 25.6 38.4









#33
Hexachlorobutadiene
Concen: 36.015 ng/ul
RT: 11.016 min Scan# 1348
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:225 Resp: 377692
Ion Ratio Lower Upper
225 100
223 64.1 50.5 75.7
227 65.9 53.9 80.9



#34
Caprolactam
Concen: 43.318 ng/ul
RT: 11.599 min Scan# 1447
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:113 Resp: 185243
Ion Ratio Lower Upper
113 100
55 148.9 130.4 195.6
56 127.3 102.2 153.4









#35
4-Chloro-3-methylphenol
Concen: 41.183 ng/ul
RT: 11.963 min Scan# 1509
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:107 Resp: 665788
Ion Ratio Lower Upper
107 100
144 24.7 18.2 27.4
142 71.6 57.1 85.7



#36
2-Methylnaphthalene
Concen: 37.167 ng/ul
RT: 12.340 min Scan# 1573
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:142 Resp: 1285544
Ion Ratio Lower Upper
142 100
141 88.5 73.4 110.0









#37
1-Methylnaphthalene
Concen: 35.337 ng/ul
RT: 12.557 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:142 Resp: 1216581
Ion Ratio Lower Upper
142 100
141 100.1 74.2 111.4
116 4.5 3.4 5.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.160 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:216 Resp: 712492
Ion Ratio Lower Upper
216 100
214 75.1 62.7 94.1
179 21.6 18.6 27.8
108 17.9 15.2 22.8









#40
Hexachlorocyclopentadiene
Concen: 33.213 ng/ul
RT: 12.687 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:237 Resp: 424283
Ion Ratio Lower Upper
237 100
235 65.0 49.0 73.4
272 12.9 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 42.014 ng/ul
RT: 12.946 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:196 Resp: 452660
Ion Ratio Lower Upper
196 100
198 91.9 79.9 119.9
200 31.5 23.6 35.4









#42
2,4,5-Trichlorophenol
Concen: 41.704 ng/ul
RT: 13.016 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:196 Resp: 490558
Ion Ratio Lower Upper
196 100
198 96.3 76.5 114.7
200 32.9 24.6 37.0



#43
1,1'-Biphenyl
Concen: 34.658 ng/ul
RT: 13.351 min Scan# 1745
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:154 Resp: 1632257
Ion Ratio Lower Upper
154 100
153 41.9 32.1 48.1
76 13.0 11.0 16.6









#44
2-Chloronaphthalene
Concen: 35.720 ng/ul
RT: 13.393 min Scan# 1752
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:162 Resp: 1298985
Ion Ratio Lower Upper
162 100
164 32.8 26.6 40.0
127 42.3 36.3 54.5



#45
2-Nitroaniline
Concen: 43.336 ng/ul
RT: 13.593 min Scan# 1786
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion: 65 Resp: 446562
Ion Ratio Lower Upper
65 100
92 69.0 54.5 81.7
138 90.3 75.4 113.0









#47
Dimethylphthalate
Concen: 38.107 ng/ul
RT: 13.969 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:163 Resp: 1686194
Ion Ratio Lower Upper
163 100
194 4.5 3.3 4.9
164 10.5 8.6 12.8



#48
2,6-Dinitrotoluene
Concen: 43.836 ng/ul
RT: 14.093 min Scan# 1871
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:165 Resp: 351268
Ion Ratio Lower Upper
165 100
89 63.7 59.5 89.3
121 21.9 20.0 30.0









#50
Acenaphthylene
Concen: 36.913 ng/ul
RT: 14.234 min Scan# 1895
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:152 Resp: 1961530
Ion Ratio Lower Upper
152 100
151 20.8 16.2 24.2
153 12.7 10.3 15.5



#51
3-Nitroaniline
Concen: 37.983 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:138 Resp: 327257
Ion Ratio Lower Upper
138 100
108 13.4 10.3 15.5
92 128.9 103.9 155.9









#52
Acenaphthene
Concen: 35.993 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:153 Resp: 1383449
Ion Ratio Lower Upper
153 100
152 47.1 38.6 57.8
154 92.1 69.3 103.9



#53
2,4-Dinitrophenol
Concen: 44.128 ng/ul
RT: 14.622 min Scan# 1961
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:184 Resp: 196482
Ion Ratio Lower Upper
184 100
63 74.7 64.8 97.2
154 65.1 59.7 89.5









#55
4-Nitrophenol
Concen: 41.019 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:109 Resp: 316872
Ion Ratio Lower Upper
109 100
139 108.0 84.7 127.1
65 123.4 105.0 157.6



#56
Dibenzofuran
Concen: 37.394 ng/ul
RT: 14.916 min Scan# 2011
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:168 Resp: 1987088
Ion Ratio Lower Upper
168 100
139 39.4 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 45.040 ng/ul
RT: 14.875 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:165 Resp: 519849
Ion Ratio Lower Upper
165 100
63 50.3 42.8 64.2
182 3.0 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 45.512 ng/ul
RT: 15.140 min Scan# 2049
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:232 Resp: 442979
Ion Ratio Lower Upper
232 100
131 41.9 37.5 56.3
130 2.3 2.2 3.4
166 28.8 26.2 39.2









#59
Diethylphthalate
Concen: 40.684 ng/ul
RT: 15.340 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:149 Resp: 1800093
Ion Ratio Lower Upper
149 100
177 21.8 17.2 25.8
150 12.5 10.2 15.2



#61
Fluorene
Concen: 37.931 ng/ul
RT: 15.563 min Scan# 2121
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:166 Resp: 1644820
Ion Ratio Lower Upper
166 100
165 95.8 78.4 117.6
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 37.419 ng/ul
RT: 15.557 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampled :
SLCS520


Tgt Ion:204 Resp: 846134
Ion Ratio Lower Upper
204 100
206 33.2 26.1 39.1
141 63.9 53.4 80.0



#63
4-Nitroaniline
Concen: 44.303 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:138 Resp: 365739
Ion Ratio Lower Upper
138 100
92 57.6 48.6 72.8
108 99.3 80.7 121.1









#66
4,6-Dinitro-2-methylphenol
Concen: 42.066 ng/ul
RT: 15.640 min Scan# 2134
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:198 Resp: 289116
Ion Ratio Lower Upper
198 100
182 3.4 3.0 4.4
77 28.1 23.4 35.0



#67
N-Nitrosodiphenylamine
Concen: 38.577 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:169 Resp: 1456920
Ion Ratio Lower Upper
169 100
168 65.4 53.5 80.3
167 34.6 27.9 41.9









#68
4-Bromophenyl-phenylether
Concen: 40.088 ng/ul
RT: 16.451 min Scan# 2272
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:248 Resp: 556630
Ion Ratio Lower Upper
248 100
250 98.8 81.1 121.7
141 79.0 64.7 97.1



#69
Hexachlorobenzene
Concen: 39.053 ng/ul
RT: 16.569 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:284 Resp: 633674
Ion Ratio Lower Upper
284 100
142 33.5 30.3 45.5
249 32.6 26.3 39.5









#70
Atrazine
Concen: 41.581 ng/ul
RT: 16.722 min Scan# 2318
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:200 Resp: 569610
Ion Ratio Lower Upper
200 100
173 27.3 22.2 33.2
215 48.4 37.6 56.4



#71
Pentachlorophenol
Concen: 40.328 ng/ul
RT: 16.910 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:266 Resp: 367098
Ion Ratio Lower Upper
266 100
264 61.6 53.8 80.6
268 58.2 52.7 79.1









#72
Phenanthrene
Concen: 38.431 ng/u1
RT: 17.304 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:178 Resp: 2703340
Ion Ratio Lower Upper
178 100
179 14.5 12.5 18.7
176 19.0 14.6 21.8



#74
Anthracene
Concen: 38.230 ng/u1
RT: 17.398 min Scan# 2433
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:178 Resp: 2745502
Ion Ratio Lower Upper
178 100
179 15.1 11.6 17.4
176 18.5 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.064 ng/uL
RT: 13.316 min Scan# 1739
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:216 Resp: 694260
Ion Ratio Lower Upper
216 100
214 79.6 64.6 97.0
218 46.7 38.7 58.1



#76
Pentachlorobenzene
Concen: 35.357 ng/uL
RT: 14.834 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:250 Resp: 701439
Ion Ratio Lower Upper
250 100
248 66.2 51.9 77.9
252 68.5 54.7 82.1









#77
 Carbazole
 Concen: 39.786 ng/ul
 RT: 17.663 min Scan# 2478
 Delta R.T. 0.000 min
 Lab File: BN015972.D
 Acq: 19 Aug 2021 15:26

Instrument :
 BNA_N
 ClientSampleId :
 SLCS520

Tgt Ion:167 Resp: 2512142




Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.1	25.7
139	13.4	11.6	17.4




#78
 Di-n-butylphthalate
 Concen: 41.912 ng/ul
 RT: 18.234 min Scan# 2575
 Delta R.T. 0.006 min
 Lab File: BN015972.D
 Acq: 19 Aug 2021 15:26

Tgt Ion:149 Resp: 3008904

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.5	11.3
104	5.5	5.0	7.4









#80
Fluoranthene
Concen: 37.308 ng/u1
RT: 19.322 min Scan# 2760
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:202 Resp: 3296231
Ion Ratio Lower Upper
202 100
101 10.0 8.0 12.0
100 8.0 6.7 10.1



#82
Pyrene
Concen: 35.924 ng/u1
RT: 19.686 min Scan# 2822
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:202 Resp: 3303357
Ion Ratio Lower Upper
202 100
101 11.4 10.1 15.1
100 8.9 9.2 13.8#









#83
Butylbenzylphthalate
Concen: 43.963 ng/ul
RT: 20.586 min Scan# 2975
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:149 Resp: 1359445
Ion Ratio Lower Upper
149 100
91 74.8 65.2 97.8
206 22.6 17.1 25.7



#84
3,3'-Dichlorobenzidine
Concen: 39.881 ng/ul
RT: 21.369 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:252 Resp: 1107152
Ion Ratio Lower Upper
252 100
254 66.2 52.0 78.0
126 10.0 8.5 12.7









#85
Benzo(a)anthracene
Concen: 37.735 ng/ul
RT: 21.439 min Scan# 3120
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:228 Resp: 3291597
Ion Ratio Lower Upper
228 100
229 19.3 14.5 21.7
226 28.1 20.9 31.3



#86
Bis(2-ethylhexyl)phthalate
Concen: 41.289 ng/ul
RT: 21.369 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:149 Resp: 2024874
Ion Ratio Lower Upper
149 100
167 27.2 20.6 30.8
279 4.6 3.4 5.2









#87
Chrysene
Concen: 38.077 ng/ul
RT: 21.492 min Scan# 3129
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:228 Resp: 3215749
Ion Ratio Lower Upper
228 100
226 27.9 24.2 36.2
229 18.4 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 40.229 ng/ul
RT: 22.292 min Scan# 3265
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:149 Resp: 3464090









#90
Benzo(b)fluoranthene
Concen: 37.644 ng/ul
RT: 23.121 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:252 Resp: 3367168
Ion Ratio Lower Upper
252 100
253 22.9 16.6 25.0
125 8.4 7.1 10.7



#91
Benzo(k)fluoranthene
Concen: 37.233 ng/ul
RT: 23.174 min Scan# 3415
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:252 Resp: 3315911
Ion Ratio Lower Upper
252 100
253 21.3 17.7 26.5
125 8.3 7.0 10.6









#93
Benzo(a)pyrene
Concen: 37.139 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. 0.000 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26




Instrument :
BNA_N
ClientSampleId :
SLCS520


Tgt Ion:252 Resp: 3007124
Ion Ratio Lower Upper
252 100
253 21.2 18.0 27.0
125 9.1 7.2 10.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 38.439 ng/ul
RT: 26.351 min Scan# 3955
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:276 Resp: 3859119
Ion Ratio Lower Upper
276 100
138 20.9 15.9 23.9
277 25.2 19.5 29.3









#95
Dibenzo(a,h)anthracene
Concen: 38.748 ng/ul
RT: 26.368 min Scan# 3958
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Instrument :
BNA_N
ClientSampleId :
SLCS520

Tgt Ion:278 Resp: 3211505
Ion Ratio Lower Upper
278 100
139 13.2 11.0 16.6
279 23.6 19.9 29.9



#96
Benzo(g,h,i)perylene
Concen: 37.722 ng/ul
RT: 27.115 min Scan# 4085
Delta R.T. 0.006 min
Lab File: BN015972.D
Acq: 19 Aug 2021 15:26

Tgt Ion:276 Resp: 3131254
Ion Ratio Lower Upper
276 100
138 17.3 13.4 20.0
277 21.2 18.5 27.7

