


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081721\
 Data File : BN015889.D
 Acq On : 17 Aug 2021 08:22
 Operator : CG/JU
 Sample : DFTPP202
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP202

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Aug 17 09:43:53 2021



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	393948	PASS
68	69	0.00	2	2.0	8987	PASS
69	198	0.00	100	44.2	444992	PASS
70	69	0.00	2	0.5	2151	PASS
127	198	10	80	47.4	477355	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1007296	PASS
199	198	5	9	6.8	68163	PASS
275	198	10	60	27.9	280875	PASS
365	198	1	100	3.8	38136	PASS
441	198	0.01	100	14.1	141941	PASS
442	198	50	100	87.9	885483	PASS
443	442	15	24	20.1	178208	PASS

m/z	Abundance
40.00	1211.0
44.00	2083.0
46.80	372.0
49.00	534.0
51.00	313.0
73.00	1524.0
75.00	337.0
84.00	923.0
85.90	778.0
95.80	598.0
133.10	349.0
133.80	395.0
147.20	362.0
191.00	708.0
191.90	351.0
193.10	439.0
207.00	4088.0
208.10	916.0
209.00	385.0
281.00	553.0
354.90	324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

m/z	Abundance
37.00	1471.0
38.00	4564.0
39.00	23952.0
40.00	2099.0
41.10	861.0
43.10	463.0
44.00	1457.0
44.90	571.0
48.00	416.0
49.10	3974.0
50.00	90688.0
51.00	326336.0
52.00	15830.0
53.00	644.0
55.00	2163.0
56.00	10717.0
57.00	21600.0
57.90	1117.0
59.00	315.0
61.10	4293.0
62.00	4799.0
63.00	15508.0
64.00	1849.0
65.00	7142.0
66.70	413.0
68.10	6910.0
69.00	350656.0
70.00	1281.0
73.00	3895.0
74.00	36576.0
75.00	57600.0
76.00	18136.0
77.00	362752.0
78.00	26376.0
79.00	30576.0
80.00	21000.0
81.00	29504.0
82.00	6801.0
83.10	5801.0
83.90	1176.0
85.00	5676.0
86.00	8861.0
87.00	3896.0
88.10	1173.0
89.10	864.0
91.00	6578.0
92.00	7860.0
93.00	48464.0
94.00	3899.0
96.00	1885.0
98.00	36440.0
99.00	27888.0
100.00	2513.0
101.00	14745.0
101.90	854.0
103.00	6642.0
104.00	10327.0
105.00	9720.0
106.10	2758.0
107.00	121864.0
108.00	19944.0
109.10	3854.0
110.00	201792.0
111.00	33312.0
112.00	3786.0
112.90	1164.0
114.70	528.0
115.00	503.0
116.00	7108.0
117.00	96232.0
118.00	7980.0
119.00	1031.0
120.10	1213.0
120.90	535.0
122.00	8328.0
123.00	11681.0
123.90	5723.0
124.90	4829.0
127.00	357504.0
128.00	28304.0
129.00	149824.0
130.00	12455.0
131.00	2313.0
132.00	1387.0
133.00	618.0
134.00	5013.0
135.00	12610.0
136.00	4644.0
137.00	6009.0
137.90	1206.0
139.10	626.0
140.00	2135.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

141.00	21064.0
142.10	5972.0
143.00	4119.0
143.90	1332.0
145.10	987.0
145.90	3497.0
147.00	10194.0
148.00	21992.0
149.00	4664.0
150.00	1317.0
151.10	2154.0
151.90	1407.0
153.00	6145.0
154.00	3837.0
155.00	11256.0
156.00	16824.0
157.10	2394.0
157.90	3839.0
158.90	3184.0
160.10	5756.0
161.00	9946.0
162.10	2136.0
162.90	1123.0
163.90	781.0
165.00	7800.0
166.20	5362.0
167.00	35960.0
168.00	21048.0
169.00	2779.0
170.00	1513.0
170.90	1726.0
172.00	3265.0
172.90	4375.0
174.00	8745.0
175.00	14009.0
175.90	3409.0
177.00	8640.0
177.90	2672.0
179.00	29360.0
180.00	21488.0
181.00	9127.0
182.00	2359.0
182.90	905.0
184.00	2312.0
185.00	14933.0
186.00	103824.0
187.00	29288.0
188.00	4057.0
188.90	7179.0
189.90	1247.0
191.00	4202.0
192.00	11009.0
193.00	11655.0
194.00	2333.0
195.10	1140.0
196.10	18840.0
198.00	721984.0
198.90	51288.0
200.00	4120.0
200.90	1062.0
201.50	2975.0
203.00	5604.0
204.00	25896.0
205.00	43920.0
206.00	189760.0
207.00	31328.0
208.00	8342.0
209.00	3368.0
210.10	3296.0
211.00	7447.0
214.20	330.0
215.00	2809.0
216.00	3892.0
217.00	50736.0
218.00	7069.0
218.90	714.0
221.10	25040.0
221.90	6784.0
223.00	11544.0
224.00	104264.0
225.10	30248.0
226.10	3178.0
227.00	43560.0
228.00	6531.0
229.00	9389.0
230.00	1347.0
231.00	4816.0
231.90	556.0
233.00	1022.0
234.00	2888.0
234.90	2692.0
236.00	2788.0
236.90	4739.0
237.90	808.0
238.90	1826.0
240.00	1670.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

241.10	2587.0
242.00	5246.0
243.10	5600.0
244.00	82616.0
245.00	11196.0
246.00	16792.0
247.00	4020.0
247.80	917.0
248.90	2974.0
250.10	844.0
250.80	991.0
252.10	557.0
252.90	1429.0
255.00	434560.0
256.00	64392.0
257.00	5577.0
258.00	24376.0
259.00	4119.0
260.00	965.0
260.80	560.0
261.10	625.0
263.90	731.0
265.00	10241.0
266.00	1597.0
269.90	724.0
270.90	1054.0
272.00	1006.0
273.00	10632.0
274.00	27488.0
275.00	185472.0
276.00	24792.0
277.00	15251.0
278.00	3121.0
278.90	768.0
281.10	936.0
281.90	490.0
283.00	1077.0
284.30	1189.0
285.00	2587.0
285.90	917.0
289.00	730.0
290.00	800.0
291.10	725.0
292.00	936.0
293.00	3622.0
294.00	1589.0
294.90	1238.0
296.00	52280.0
297.10	7660.0
298.00	771.0
300.90	487.0
302.10	901.0
303.00	4809.0
304.00	1753.0
308.00	742.0
309.00	553.0
309.90	552.0
312.90	1027.0
314.00	3521.0
314.90	5810.0
316.10	2740.0
316.90	564.0
321.00	1688.0
321.90	747.0
323.00	17048.0
323.90	3596.0
326.00	556.0
327.00	3619.0
327.90	1447.0
331.90	1254.0
332.90	1629.0
334.00	10266.0
335.00	3509.0
341.00	1814.0
341.90	667.0
345.90	2758.0
347.00	795.0
351.90	4350.0
352.90	3704.0
354.10	5311.0
355.00	1257.0
359.10	493.0
364.90	22824.0
366.00	3346.0
369.90	426.0
370.90	1088.0
372.00	8856.0
372.90	2570.0
376.90	373.0
383.00	2648.0
384.10	689.0
385.00	404.0
390.00	1356.0
390.90	329.0
391.90	462.0
401.00	513.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

401.90	3441.0
403.00	5245.0
403.90	1699.0
404.90	464.0
421.00	3554.0
422.00	4604.0
423.00	28176.0
424.10	8652.0
424.90	1391.0
441.00	75616.0
442.00	471232.0
443.00	100320.0
443.90	9618.0
444.90	820.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

m/z	Abundance
37.10	2048.0
38.10	7242.0
39.00	35024.0
40.00	2009.0
41.00	936.0
43.00	390.0
44.00	2143.0
45.00	810.0
48.10	439.0
49.10	5403.0
50.10	127616.0
51.00	458048.0
52.00	21608.0
53.00	836.0
55.00	2900.0
56.00	15430.0
57.00	31752.0
57.90	1444.0
60.10	594.0
61.00	6203.0
62.00	7093.0
63.00	18864.0
64.00	3202.0
65.00	8173.0
66.10	533.0
67.20	505.0
68.10	10796.0
69.00	524032.0
69.90	2951.0
71.00	369.0
73.10	5614.0
74.00	53416.0
75.00	82384.0
76.10	29760.0
77.00	543552.0
78.00	35936.0
79.00	41552.0
80.00	30648.0
81.00	42840.0
82.00	8802.0
83.00	9141.0
83.90	1490.0
85.00	7137.0
86.00	14154.0
87.00	5043.0
87.90	2240.0
89.00	686.0
91.00	9695.0
92.00	11687.0
93.00	69136.0
94.00	5218.0
95.00	1156.0
96.00	2677.0
97.00	1949.0
98.00	54552.0
99.00	39232.0
100.00	3089.0
101.00	20632.0
101.90	1207.0
103.00	6866.0
103.90	13673.0
105.00	12947.0
106.20	5742.0
107.00	177280.0
108.00	28576.0
109.20	5357.0
110.00	304320.0
111.00	47432.0
112.00	5466.0
113.10	2286.0
115.00	679.0
116.10	9655.0
117.00	146944.0
118.00	10429.0
119.00	1502.0
119.90	2510.0
121.20	551.0
122.00	12478.0
123.00	18024.0
124.00	7868.0
125.00	7483.0
127.00	552896.0
128.00	45400.0
129.00	226368.0
130.00	22664.0
131.00	3409.0
132.10	2157.0
132.80	1569.0
134.00	6691.0
135.00	19600.0
136.10	6718.0
137.00	10824.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

137.90	2569.0
138.90	898.0
140.00	2409.0
141.00	28632.0
142.00	8853.0
143.00	6461.0
144.00	1810.0
145.10	1677.0
146.00	4924.0
147.00	15599.0
148.00	36664.0
149.00	7001.0
150.10	2208.0
151.10	3162.0
151.90	2824.0
153.00	8328.0
153.90	8442.0
155.00	18464.0
156.00	23168.0
157.00	4486.0
157.90	6105.0
159.00	3780.0
160.00	9394.0
161.00	13271.0
161.90	3168.0
163.00	938.0
164.00	1772.0
165.00	10643.0
166.00	7961.0
167.00	55312.0
168.00	28464.0
169.00	5196.0
170.10	1532.0
170.90	2592.0
172.00	5942.0
173.00	7298.0
174.00	12300.0
175.00	23208.0
176.10	5963.0
177.00	13232.0
177.90	5100.0
179.00	46880.0
180.00	30872.0
181.00	13447.0
182.10	2200.0
183.00	998.0
184.00	3890.0
185.00	22480.0
186.00	169728.0
187.00	50752.0
188.00	5543.0
189.00	10753.0
189.90	1789.0
190.90	5804.0
192.00	14380.0
193.00	16045.0
194.00	3528.0
196.00	29496.0
198.00	1150976.0
199.00	78536.0
200.00	5457.0
201.50	5589.0
203.00	8909.0
204.00	45248.0
205.00	79560.0
206.00	323968.0
207.10	49776.0
208.00	12263.0
209.10	4533.0
210.10	4445.0
211.00	11561.0
212.00	1370.0
213.00	1012.0
214.00	431.0
215.00	4103.0
216.10	6796.0
217.00	87776.0
218.00	9171.0
219.10	932.0
221.00	41640.0
221.90	12405.0
223.00	19272.0
224.00	176448.0
225.00	48072.0
226.00	6215.0
227.00	77608.0
228.00	11884.0
229.00	15709.0
230.00	2214.0
231.10	6441.0
231.90	1000.0
233.00	1144.0
234.00	4782.0
235.00	5840.0
236.00	3173.0
237.10	7467.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

238.00	1152.0
239.00	2881.0
240.10	3443.0
241.00	4569.0
242.00	7917.0
243.10	9117.0
244.10	130608.0
245.00	17640.0
246.00	30424.0
247.00	6713.0
247.90	1018.0
248.90	5303.0
250.00	781.0
251.10	1437.0
251.90	1266.0
253.00	2867.0
255.00	702400.0
256.00	97216.0
257.00	9208.0
258.00	40944.0
259.00	7008.0
260.00	1364.0
261.00	1466.0
262.00	447.0
262.90	455.0
264.10	1253.0
265.00	16608.0
266.00	3424.0
268.10	382.0
270.00	825.0
270.90	1340.0
272.00	2143.0
273.00	19840.0
274.00	54232.0
275.00	327616.0
276.00	45768.0
277.00	30256.0
278.00	4795.0
279.00	1222.0
281.00	1046.0
282.10	612.0
283.10	2566.0
284.00	3133.0
285.00	5186.0
286.00	1026.0
287.80	465.0
288.90	882.0
290.00	987.0
290.90	484.0
292.00	1433.0
293.00	6038.0
294.00	1358.0
296.00	89120.0
297.00	12232.0
298.00	924.0
299.00	450.0
301.00	851.0
302.00	1677.0
303.00	9796.0
304.00	3938.0
305.20	630.0
307.80	1104.0
309.00	830.0
310.00	1327.0
310.80	319.0
312.10	640.0
312.90	1078.0
314.00	4066.0
315.00	9569.0
316.00	5878.0
317.00	1062.0
320.00	462.0
320.90	2623.0
322.10	2054.0
323.00	27568.0
324.00	4698.0
325.10	416.0
325.60	544.0
327.00	5564.0
328.00	2678.0
329.10	494.0
330.90	392.0
332.00	1890.0
333.20	2581.0
334.00	20096.0
335.00	5678.0
335.90	633.0
338.80	447.0
340.10	453.0
341.00	3569.0
341.90	965.0
344.80	319.0
346.00	6530.0
346.70	998.0
347.10	1124.0
349.90	666.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

351.00	765.0
352.00	9682.0
353.00	5592.0
354.00	9855.0
355.00	2052.0
359.00	629.0
365.00	42112.0
366.10	5463.0
366.90	428.0
369.80	761.0
370.90	2607.0
372.00	14841.0
373.00	4786.0
373.90	474.0
377.00	546.0
382.90	4137.0
384.00	1655.0
385.10	372.0
390.00	2131.0
391.00	1157.0
391.90	938.0
401.20	859.0
402.00	6717.0
403.00	8886.0
404.00	4101.0
405.00	769.0
414.90	504.0
421.00	8215.0
422.00	8099.0
423.00	57344.0
424.00	14841.0
425.00	1811.0
441.00	161344.0
442.00	1008128.0
443.00	187584.0
444.00	18024.0
444.90	1081.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

m/z	Abundance
37.00	2201.0
38.10	5044.0
39.00	28560.0
40.00	2644.0
41.00	631.0
44.00	1878.0
44.90	719.0
49.10	3485.0
50.00	110840.0
51.00	398400.0
52.00	19384.0
53.00	982.0
55.00	2751.0
56.00	12949.0
57.00	28008.0
58.00	1080.0
58.90	355.0
61.00	5900.0
62.00	6349.0
63.00	16172.0
64.00	2031.0
65.00	7923.0
66.00	542.0
66.80	389.0
67.00	388.0
68.10	9254.0
69.00	460288.0
70.00	2221.0
73.00	5138.0
74.00	46576.0
75.00	70304.0
76.10	24488.0
77.10	489024.0
78.00	31720.0
79.00	38624.0
80.00	28936.0
81.00	40496.0
82.00	8525.0
83.00	7693.0
84.00	1638.0
85.00	6799.0
86.00	11995.0
86.90	4578.0
88.00	1629.0
89.00	702.0
91.00	9095.0
92.00	9873.0
93.00	65920.0
93.90	3930.0
95.00	842.0
95.90	2902.0
97.10	1139.0
98.00	50072.0
99.00	34808.0
99.90	3813.0
101.00	20408.0
102.00	1256.0
103.00	7993.0
104.00	14504.0
105.00	11310.0
106.10	3432.0
107.00	161024.0
108.00	25520.0
109.10	6344.0
110.00	285056.0
111.00	40824.0
112.00	5632.0
112.90	1606.0
114.20	577.0
115.00	836.0
116.00	8979.0
117.00	132992.0
118.00	8649.0
118.90	790.0
120.00	1233.0
121.10	957.0
122.00	11259.0
123.00	15859.0
124.00	7358.0
125.00	6374.0
127.00	521664.0
128.00	40464.0
129.00	215744.0
130.00	16984.0
131.00	3449.0
132.00	1983.0
132.90	952.0
134.00	6519.0
135.00	17720.0
135.90	7521.0
137.00	9473.0
137.80	1610.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

139.10	1311.0
139.90	2536.0
141.00	31960.0
142.00	9597.0
143.00	6574.0
144.00	1991.0
145.10	1335.0
145.90	6157.0
147.00	14450.0
148.00	32432.0
149.00	7015.0
150.00	1742.0
151.20	4294.0
151.90	2894.0
153.00	10180.0
154.00	6320.0
155.00	17576.0
156.00	22864.0
157.00	4444.0
158.00	6014.0
159.00	3953.0
160.00	9587.0
161.00	13172.0
162.00	3983.0
163.00	1214.0
164.10	1247.0
165.00	11215.0
165.90	8456.0
167.00	56328.0
168.00	23800.0
169.00	4488.0
170.00	2018.0
171.00	2988.0
171.90	6770.0
173.00	6543.0
174.00	12899.0
175.00	21680.0
176.00	6377.0
177.00	9124.0
178.00	3611.0
179.00	42400.0
180.00	29872.0
181.00	14173.0
182.00	2691.0
182.80	1506.0
184.00	3519.0
185.00	22840.0
186.00	163392.0
187.00	51224.0
188.10	4582.0
189.00	9906.0
190.00	1539.0
191.00	5944.0
192.00	14340.0
193.00	14019.0
194.00	3453.0
194.90	1866.0
196.00	27568.0
198.00	1148928.0
199.00	74664.0
200.10	6488.0
201.40	4285.0
203.00	8637.0
204.00	42608.0
205.00	70560.0
206.00	332224.0
207.10	44552.0
208.00	12265.0
209.00	4301.0
210.20	5444.0
211.10	12794.0
213.00	1049.0
214.00	453.0
214.90	3781.0
216.10	7764.0
217.00	85416.0
218.00	10349.0
219.00	862.0
221.00	44496.0
221.90	13017.0
223.10	17944.0
224.00	171392.0
225.00	47760.0
226.10	5735.0
227.00	84784.0
228.00	9495.0
229.00	17184.0
230.00	2081.0
231.00	6761.0
232.00	1264.0
233.10	1765.0
234.00	4619.0
235.00	5464.0
235.90	3476.0
236.90	5699.0
238.10	909.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

238.90	3239.0
240.00	3191.0
240.90	4677.0
242.00	9810.0
243.10	10566.0
244.00	121344.0
245.00	17944.0
246.00	28992.0
247.00	6416.0
248.00	1563.0
248.90	5165.0
250.10	477.0
250.90	1256.0
251.90	1428.0
253.10	3053.0
255.00	771328.0
256.00	106864.0
257.00	10927.0
258.00	39304.0
259.00	7457.0
260.00	1133.0
261.00	1124.0
262.10	309.0
262.90	598.0
264.10	1169.0
265.00	19664.0
266.00	3084.0
266.90	666.0
270.10	797.0
271.00	2209.0
272.20	2083.0
273.00	23400.0
274.00	57240.0
275.00	329536.0
276.00	43136.0
277.00	29264.0
278.00	5646.0
279.00	890.0
281.00	743.0
282.00	776.0
283.10	3540.0
284.00	1953.0
285.10	3908.0
286.00	906.0
287.90	628.0
288.90	1279.0
290.10	1115.0
291.00	575.0
292.00	1385.0
293.10	5673.0
294.10	1940.0
295.10	3218.0
296.00	90568.0
297.00	12588.0
298.10	1030.0
301.00	1317.0
302.00	1902.0
303.00	12593.0
304.00	3549.0
305.00	647.0
308.00	1544.0
309.00	908.0
310.10	1272.0
311.00	507.0
312.10	501.0
312.90	1000.0
314.00	5719.0
315.00	10615.0
316.00	7220.0
317.00	885.0
320.20	533.0
321.00	3127.0
322.00	1806.0
323.00	33624.0
324.00	5337.0
325.10	653.0
326.10	1068.0
327.00	6242.0
328.00	1993.0
328.90	621.0
332.00	2561.0
333.10	3865.0
334.00	21624.0
335.00	6027.0
335.90	486.0
336.10	496.0
338.90	567.0
339.80	455.0
341.00	4432.0
341.90	2007.0
345.90	6519.0
347.00	1134.0
348.10	315.0
350.80	747.0
352.00	10703.0
353.00	6967.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202

354.00	10034.0
355.00	2396.0
358.80	1056.0
364.90	49472.0
366.00	6901.0
366.80	757.0
369.80	498.0
370.30	709.0
371.00	2483.0
372.00	18552.0
373.00	4462.0
374.00	386.0
377.00	403.0
383.00	5632.0
384.00	911.0
384.80	355.0
390.00	2521.0
390.90	1880.0
392.00	1942.0
396.90	325.0
401.10	955.0
402.00	7900.0
403.00	11128.0
404.00	4981.0
404.90	697.0
414.80	537.0
416.00	492.0
421.00	10686.0
422.00	8845.0
423.00	70944.0
424.00	17240.0
425.00	1951.0
439.40	351.0
441.00	188864.0
442.00	1177088.0
443.00	246720.0
444.00	23880.0
445.00	1472.0

Instrument :
BNA_N
ClientSampleId :
DFTPP202