

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081723\
 Data File : BN026942.D
 Acq On : 17 Aug 2023 14:59
 Operator : MA/JU
 Sample : 04024-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP05-20230815

Quant Time: Aug 18 03:12:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 15:56:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.109	152	7448	0.400	ng	0.00
7) Naphthalene-d8	10.938	136	18397	0.400	ng	0.00
13) Acenaphthene-d10	14.726	164	12911	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	29465	0.400	ng	0.00
29) Chrysene-d12	21.659	240	19814	0.400	ng	0.00
35) Perylene-d12	24.203	264	18725	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.618	112	2972	0.142	ng	0.00
5) Phenol-d6	7.235	99	2143	0.083	ng	0.00
8) Nitrobenzene-d5	9.294	82	4527	0.291	ng	0.00
11) 2-Methylnaphthalene-d10	12.506	152	10215	0.349	ng	0.00
14) 2,4,6-Tribromophenol	16.214	330	1988	0.289	ng	0.00
15) 2-Fluorobiphenyl	13.357	172	16132	0.337	ng	0.00
27) Fluoranthene-d10	19.495	212	30335	0.401	ng	0.00
31) Terphenyl-d14	20.080	244	25425	0.614	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.473	88	35393	4.419	ng	Qvalue 95
6) bis(2-Chloroethyl)ether	7.531	93	1233	0.057	ng	97
34) Bis(2-ethylhexyl)phtha...	21.524	149	98087	1.905	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081723\
Data File : BN026942.D
Acq On : 17 Aug 2023 14:59
Operator : MA/JU
Sample : 04024-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP05-20230815

Quant Time: Aug 18 03:12:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 15:56:02 2023
Response via : Initial Calibration

