

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081823\
 Data File : BN026952.D
 Acq On : 18 Aug 2023 10:01
 Operator : MA/JU
 Sample : PB154872BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK872

Quant Time: Aug 18 13:12:49 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 18 13:11:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.101	152	10230	0.400	ng/ul	0.00
4) Naphthalene-d8	10.927	136	31934	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.731	164	19378	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.477	188	43376	0.400	ng/ul	0.00
17) Chrysene-d12	21.650	240	32288	0.400	ng/ul	0.00
23) Perylene-d12	24.196	264	37803	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.443	96	74947	6.171	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.500	152	11535	0.245	ng/ul	0.00
18) Fluoranthene-d10	19.492	212	31705	0.311	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081823\
Data File : BN026952.D
Acq On : 18 Aug 2023 10:01
Operator : MA/JU
Sample : PB154872BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK872

Quant Time: Aug 18 13:12:49 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN080323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 18 13:11:26 2023
Response via : Initial Calibration

